



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105588

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/09/2019	11:20
2	K5258-03		1	9	100.00	>212.0	>212.0	10/09/2019	11:50
3	K5258-03DUP		1	9	100.00	>212.0	>212.0	10/09/2019	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB 105588

WorkList Name : fp10919 WorkList ID : 131862 Department : Wet-Chemistry Date : 10-09-2019 09:00:27

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2019	Solid	K5258-03	B Flash Point	Cool 4 deg C	PSEG01	N31	PL-OIL-DRUM	10/08/2019	1010A

Date/Time 10/09/19 11:20
 Received by: KS
 Relinquished by: JP

Date/Time 10/09/19 12:20
 Received by: JP
 Relinquished by: KS