

7470A


LB105593

INSTRUMENT ID: CV1

Linear ▼

 μ Abs.:

42892



A= 0.0000e+000

B= 2.3367e-004

C= -3.3709e-002

Rho= 0.9999902

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.017	0.017	217	0.000	217				
0.2	0.200	0.184	-0.016	930	0.0 %	930				
2.5	2.500	2.481	-0.019	10764	0.0 %	10764				
5.0	5.000	5.023	0.023	21642	0.0 %	21642				
7.5	7.500	7.506	0.006	32267	0.0 %	32267				
10.0	10.000	9.989	-0.011	42892	0.0 %	42892				

LB105593 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 09 Oct 2019 13:49:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	217	-	0.0000	7470A	PPB	09 Oct 2019 13:53:55	S	Std
0.2 - 1	50.2	930	-	0.2000	7470A	PPB	09 Oct 2019 13:56:04	S	Std
2.5 - 1	52.5	10764	-	2.5000	7470A	PPB	09 Oct 2019 13:58:13	S	Std
5.0 - 1	55.0	21642	-	5.0000	7470A	PPB	09 Oct 2019 14:00:22	S	Std
7.5 - 1	57.5	32267	-	7.5000	7470A	PPB	09 Oct 2019 14:02:31	S	Std
10.0 - 1	510.0	42892	-	10.0000	7470A	PPB	09 Oct 2019 14:04:39	S	Std
ICV46 - 1	ICV46	16274	3.7690	-	7470A	PPB	09 Oct 2019 14:11:11	U	SMPL
ICB46 - 1	ICB46	199	0.0128	-	7470A	PPB	09 Oct 2019 14:13:18	U	SMPL
CCV85 - 1	CCV85	22175	5.1478	-	7470A	PPB	09 Oct 2019 14:15:26	U	SMPL
CCB85 - 1	CCB85	191	0.0109	-	7470A	PPB	09 Oct 2019 14:17:33	U	SMPL
CRA - 1	CRA	940	0.1859	-	7470A	PPB	09 Oct 2019 14:19:40	U	SMPL
HighStd - 1	HighStd	42899	9.9903	-	7470A	PPB	09 Oct 2019 14:21:47	U	SMPL
ChkStd - 1	ChkStd	29742	6.9160	-	7470A	PPB	09 Oct 2019 14:23:53	U	SMPL
PB123803BL - 1	PBW	65	-0.0185	-	7470A	PPB	09 Oct 2019 14:26:01	U	SMPL
PB123803BS - 1	LCSW	17795	4.1244	-	7470A	PPB	09 Oct 2019 14:28:08	U	SMPL
K5147-04 - 1	OK-03-100719	114	-0.0071	-	7470A	PPB	09 Oct 2019 14:30:17	U	SMPL
K5147-04DUP - 1	OK-03-100719DUP	126	-0.0043	-	7470A	PPB	09 Oct 2019 14:32:25	U	SMPL
K5147-04MS - 1	OK-03-100719MS	17273	4.0024	-	7470A	PPB	09 Oct 2019 14:34:34	U	SMPL
K5147-04MSD - 1	OK-03-100719MSD	17663	4.0935	-	7470A	PPB	09 Oct 2019 14:36:43	U	SMPL
K5152-05 - 1	SU-02-100719	89	-0.0129	-	7470A	PPB	09 Oct 2019 14:38:50	U	SMPL
CCV86 - 1	CCV86	22042	5.1167	-	7470A	PPB	09 Oct 2019 14:40:58	U	SMPL
CCB86 - 1	CCB86	140	-0.0010	-	7470A	PPB	09 Oct 2019 14:43:05	U	SMPL
K5238-06 - 1	TP-COMP	44	-0.0234	-	7470A	PPB	09 Oct 2019 14:45:12	U	SMPL
PB123804BL - 1	PBW	105	-0.0092	-	7470A	PPB	09 Oct 2019 14:47:20	U	SMPL
PB123804BS - 1	LCSW	17472	4.0489	-	7470A	PPB	09 Oct 2019 14:49:27	U	SMPL
K5285-04 - 1	TP-1-S	189	0.0105	-	7470A	PPB	09 Oct 2019 14:51:34	U	SMPL
K5285-08 - 1	TP-1-D	266	0.0284	-	7470A	PPB	09 Oct 2019 14:53:42	U	SMPL
K5285-12 - 1	TP-2-S	273	0.0301	-	7470A	PPB	09 Oct 2019 14:55:50	U	SMPL
K5285-16 - 1	TP-2-D	272	0.0298	-	7470A	PPB	09 Oct 2019 14:57:58	U	SMPL
K5285-16DUP - 1	TP-2-DDUP	271	0.0296	-	7470A	PPB	09 Oct 2019 15:00:07	U	SMPL
K5285-16MS - 1	TP-2-DMS	17710	4.1045	-	7470A	PPB	09 Oct 2019 15:02:16	U	SMPL
K5285-16MSD - 1	TP-2-DMSD	17525	4.0613	-	7470A	PPB	09 Oct 2019 15:04:23	U	SMPL
CCV87 - 1	CCV87	22120	5.1350	-	7470A	PPB	09 Oct 2019 15:06:30	U	SMPL
CCB87 - 1	CCB87	76	-0.0160	-	7470A	PPB	09 Oct 2019 15:08:36	U	SMPL
K5286-02 - 1	E. OUTFALL-CREEK-SOIL-SP	197	0.0123	-	7470A	PPB	09 Oct 2019 15:10:44	U	SMPL
PB123803TB - 1	PB123720TB	127	-0.0040	-	7470A	PPB	09 Oct 2019 15:12:52	U	SMPL
PB123804TB - 1	PB123775TB	160	0.0037	-	7470A	PPB	09 Oct 2019 15:14:59	U	SMPL
K5147-04LX5 - 1		13	-0.0307	-	7470A	PPB	09 Oct 2019 15:17:06	U	SMPL
K5147-04A - 1		17490	4.0531	-	7470A	PPB	09 Oct 2019 15:19:14	U	SMPL
K5285-16LX5 - 1		56	-0.0206	-	7470A	PPB	09 Oct 2019 15:21:21	U	SMPL
K5285-16A - 1		17629	4.0856	-	7470A	PPB	09 Oct 2019 15:23:30	U	SMPL
CCV88 - 1	CCV88	22154	5.1429	-	7470A	PPB	09 Oct 2019 15:25:39	U	SMPL
CCB88 - 1	CCB88	53	-0.0213	-	7470A	PPB	09 Oct 2019 15:27:48	U	SMPL