



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105601

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/09/2019	16:20
2	CAL2	1	Water	NA	NA	20.1	7.01	10/09/2019	16:21
3	CAL3	1	Water	NA	NA	20.1	10.04	10/09/2019	16:22
4	ICV	1	Water	NA	NA	20.1	6.99	10/09/2019	16:24
5	CCV1	1	Water	NA	NA	20.1	2.02	10/09/2019	16:25
6	K5297-03	1	Water	NA	NA	13.5	6.75	10/09/2019	16:26
7	K5297-03DUP	1	Water	NA	NA	13.6	6.76	10/09/2019	16:27
8	CCV2	1	Water	NA	NA	20.1	12.02	10/09/2019	16:30

28 105601

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K4297 WorkList ID : 131892 Department : Wet-Chemistry Date : 10-09-2019 16:10:05

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/09/2019	Water	K5297-03	-√- pH	Cool 4 deg C	PSEG01	N31	HH-WATER	10/09/2019	9040C

Date/Time 10.09.19 16:13
 Received by: for
 Relinquished by: for

Date/Time 10.09.19 17:10
 Received by: for
 Relinquished by: for