


LB105617

ISM02.4

INSTRUMENT ID : CV1



Linear

A= 0.0000e+000

B= 2.2769e-004

C= 1.2629e-003

Rho= 0.9998294

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.032	0.032	136	0.000	136				
0.05	0.050									
0.20	0.200	0.209	0.009	914	0.0 %	914				
2.50	2.500	2.529	0.029	11102	0.0 %	11102				
5.00	5.000	4.851	-0.149	21299	0.0 %	21299				
7.50	7.500	7.556	0.056	33181	0.0 %	33181				
10.0	10.000	10.022	0.022	44010	0.0 %	44010				

10
1
5
3
0

rs 10.10.19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	136	-		0 ISM02.4	PPB	10/10/2019 12:57	Std
	0.2 S0.2	914	-		0.2 ISM02.4	PPB	10/10/2019 13:00	Std
	2.5 S2.5	11102	-		2.5 ISM02.4	PPB	10/10/2019 13:06	Std
	5 S5.0	21299	-		5 ISM02.4	PPB	10/10/2019 13:08	Std
	7.5 S7.5	33181	-		7.5 ISM02.4	PPB	10/10/2019 13:10	Std
	10 S10.0	44010	-		10 ISM02.4	PPB	10/10/2019 13:13	Std
ICV	ICV	17773	4.0481	-	ISM02.4	PPB	10/10/2019 13:18	SMPL
ICB	ICB	56	0.014	-	ISM02.4	PPB	10/10/2019 13:20	SMPL
CCV001	CCV001	21575	4.9137	-	ISM02.4	PPB	10/10/2019 13:23	SMPL
CCB001	CCB001	58	0.0145	-	ISM02.4	PPB	10/10/2019 13:25	SMPL
PB123833BL	PBW833	92	0.0222	-	ISM02.4	PPB	10/10/2019 13:27	SMPL
K5065-01	MDBAD0	18576	4.2309	-	ISM02.4	PPB	10/10/2019 13:29	SMPL
K5065-04	MDBAL3	138	0.0327	-	ISM02.4	PPB	10/10/2019 13:32	SMPL
K5065-05	MDBAL5	149	0.0352	-	ISM02.4	PPB	10/10/2019 13:34	SMPL
K5065-06	MDBAL5D	136	0.0322	-	ISM02.4	PPB	10/10/2019 13:36	SMPL
K5065-07	MDBAL5S	5134	1.1702	-	ISM02.4	PPB	10/10/2019 13:39	SMPL
K5065-08	MDBAL6	140	0.0331	-	ISM02.4	PPB	10/10/2019 13:41	SMPL
K5065-09	MDBAL7	173	0.0407	-	ISM02.4	PPB	10/10/2019 13:43	SMPL
K5065-10	MDBAL8	202	0.0473	-	ISM02.4	PPB	10/10/2019 13:45	SMPL
K5065-11	MDBAL9	194	0.0454	-	ISM02.4	PPB	10/10/2019 13:48	SMPL
K5065-12	MDBAM0	559	0.1285	-	ISM02.4	PPB	10/10/2019 13:50	SMPL
K5065-13	MDBAM1	335	0.0775	-	ISM02.4	PPB	10/10/2019 13:52	SMPL
K5065-14	MDBAM2	196	0.0459	-	ISM02.4	PPB	10/10/2019 13:55	SMPL
K5065-15	MDBAL4	150	0.0354	-	ISM02.4	PPB	10/10/2019 13:57	SMPL
K5065-19	MDBAD1	22416	5.1052	-	ISM02.4	PPB	10/10/2019 13:59	SMPL
PB123834BL	PBW834	67	0.0165	-	ISM02.4	PPB	10/10/2019 14:01	SMPL
K5103-01	MJLMG9	160	0.0377	-	ISM02.4	PPB	10/10/2019 14:04	SMPL
K5103-02	MJLMK0	99	0.0238	-	ISM02.4	PPB	10/10/2019 14:06	SMPL
K5103-03	MJLMH0	215	0.0502	-	ISM02.4	PPB	10/10/2019 14:08	SMPL
K5103-04	MJLMJ3	222	0.0518	-	ISM02.4	PPB	10/10/2019 14:11	SMPL
CCV002	CCV002	21419	4.8782	-	ISM02.4	PPB	10/10/2019 14:13	SMPL
CCB002	CCB002	61	0.0152	-	ISM02.4	PPB	10/10/2019 14:15	SMPL
K5103-05	MJLMH9	126	0.03	-	ISM02.4	PPB	10/10/2019 14:17	SMPL
K5103-06	MJLMJ1	130	0.0309	-	ISM02.4	PPB	10/10/2019 14:20	SMPL
K5103-07	MJLMJ5	166	0.0391	-	ISM02.4	PPB	10/10/2019 14:22	SMPL
K5103-08	MJLMH1	162	0.0381	-	ISM02.4	PPB	10/10/2019 14:24	SMPL
K5103-09	MJLMH1D	136	0.0322	-	ISM02.4	PPB	10/10/2019 14:26	SMPL
K5103-10	MJLMH1S	5225	1.191	-	ISM02.4	PPB	10/10/2019 14:29	SMPL
K5103-11	MJLMK7	301	0.0698	-	ISM02.4	PPB	10/10/2019 14:31	SMPL
K5103-12	MJLMK8	77	0.0188	-	ISM02.4	PPB	10/10/2019 14:33	SMPL
K5103-13	MJLMK9	105	0.0252	-	ISM02.4	PPB	10/10/2019 14:36	SMPL
PB123835BL	PBW835	93	0.0224	-	ISM02.4	PPB	10/10/2019 14:38	SMPL
K5104-01	MJLMJ0	89	0.0215	-	ISM02.4	PPB	10/10/2019 14:40	SMPL
K5104-02	MJLMJ4	144	0.0341	-	ISM02.4	PPB	10/10/2019 14:42	SMPL
K5104-03	MJLMJ2	103	0.0247	-	ISM02.4	PPB	10/10/2019 14:45	SMPL
K5104-04	MJLMJ6	125	0.0297	-	ISM02.4	PPB	10/10/2019 14:47	SMPL

K5104-05	MJLMJ7	150	0.0354 -	ISM02.4	PPB	10/10/2019 14:49	SMPL
K5104-06	MJLMJ7D	136	0.0322 -	ISM02.4	PPB	10/10/2019 14:51	SMPL
K5104-07	MJLMJ7S	4932	1.1242 -	ISM02.4	PPB	10/10/2019 14:54	SMPL
PB123836BL	PBW836	69	0.017 -	ISM02.4	PPB	10/10/2019 14:56	SMPL
K5164-01	MH0FW7	172	0.0404 -	ISM02.4	PPB	10/10/2019 14:58	SMPL
K5164-02	MH0FW9	168	0.0395 -	ISM02.4	PPB	10/10/2019 15:01	SMPL
CCV003	CCV003	21605	4.9206 -	ISM02.4	PPB	10/10/2019 15:03	SMPL
CCB003	CCB003	66	0.0163 -	ISM02.4	PPB	10/10/2019 15:05	SMPL
K5164-03	MH0FX1	184	0.0432 -	ISM02.4	PPB	10/10/2019 15:07	SMPL
K5164-04	MH0FX3	217	0.0507 -	ISM02.4	PPB	10/10/2019 15:10	SMPL
K5164-05	MH0FX5	336	0.0778 -	ISM02.4	PPB	10/10/2019 15:12	SMPL
K5164-06	MH0FX7	187	0.0438 -	ISM02.4	PPB	10/10/2019 15:14	SMPL
K5164-07	MH0FX9	182	0.0427 -	ISM02.4	PPB	10/10/2019 15:17	SMPL
K5164-08	MH0FY1	170	0.04 -	ISM02.4	PPB	10/10/2019 15:19	SMPL
K5164-09	MH0FY3	182	0.0427 -	ISM02.4	PPB	10/10/2019 15:21	SMPL
K5164-10	MH0FY5	153	0.0361 -	ISM02.4	PPB	10/10/2019 15:23	SMPL
K5164-11	MH0FY7	153	0.0361 -	ISM02.4	PPB	10/10/2019 15:26	SMPL
K5164-12	MH0FY9	166	0.0391 -	ISM02.4	PPB	10/10/2019 15:28	SMPL
K5164-13	MH0FZ1	187	0.0438 -	ISM02.4	PPB	10/10/2019 15:30	SMPL
K5164-14	MH0FZ3	284	0.0659 -	ISM02.4	PPB	10/10/2019 15:33	SMPL
K5164-15	MH0FZ3D	285	0.0662 -	ISM02.4	PPB	10/10/2019 15:35	SMPL
K5164-16	MH0FZ3S	5150	1.1739 -	ISM02.4	PPB	10/10/2019 15:37	SMPL
K5164-17	MH0FZ5	288	0.0668 -	ISM02.4	PPB	10/10/2019 15:40	SMPL
K5164-18	MH0FZ7	356	0.0823 -	ISM02.4	PPB	10/10/2019 15:42	SMPL
K5164-19	MH0FZ9	164	0.0386 -	ISM02.4	PPB	10/10/2019 15:44	SMPL
K5164-20	MH0GA1	356	0.0823 -	ISM02.4	PPB	10/10/2019 15:46	SMPL
K5164-21	MH0GA3	262	0.0609 -	ISM02.4	PPB	10/10/2019 15:49	SMPL
K5164-22	MH0GA5	249	0.058 -	ISM02.4	PPB	10/10/2019 15:51	SMPL
CCV004	CCV004	21825	4.9707 -	ISM02.4	PPB	10/10/2019 15:53	SMPL
CCB004	CCB004	84	0.0204 -	ISM02.4	PPB	10/10/2019 15:56	SMPL
PB123837BL	PBW837	88	0.0213 -	ISM02.4	PPB	10/10/2019 15:58	SMPL
K5085-07	MCOB15	58	0.0145 -	ISM02.4	PPB	10/10/2019 16:00	SMPL
PB123838BL	PBW838	56	0.014 -	ISM02.4	PPB	10/10/2019 16:12	SMPL
K5086-02	MCOB00	68	0.0167 -	ISM02.4	PPB	10/10/2019 16:14	SMPL
K5086-22	MCOB51	31	0.0083 -	ISM02.4	PPB	10/10/2019 16:16	SMPL
CCV005	CCV005	21755	4.9547 -	ISM02.4	PPB	10/10/2019 16:18	SMPL
CCB005	CCB005	54	0.0136 -	ISM02.4	PPB	10/10/2019 16:21	SMPL