


LB105634

245.1

INSTRUMENT ID : CU1

 μ Abs.:

38457



Linear

A= 0.0000e+000

B= 2.5923e-004

C= -2.3477e-002

Rho= 0.9999355

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.007	-0.007	64	0.000	64				
0.05	0.050	0.038	-0.012	236	0.0 %	236				
0.2	0.200	0.188	-0.012	816	0.0 %	816				
2.5	2.500	2.538	0.038	9882	0.0 %	9882				
5.0	5.000	4.966	-0.034	19249	0.0 %	19249				
7.5	7.500	7.581	0.081	29335	0.0 %	29335				
10.0	10.000	9.946	-0.054	38457	0.0 %	38457				

LB105634 INSTRUMENT ID : CV1

Method: 245.1

Operator: Admin

Date of Analysis: 11 Oct 2019 08:26:59

Sample ID	Extended ID	μ Abs.	Conc.	Std. Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	64	-	0.0000	245.1	PPB	11 Oct 2019 08:33:13	S	Std
0.05 - 1	50.05	236	-	0.0500	245.1	PPB	11 Oct 2019 08:35:32	S	Std
0.2 - 1	50.2	816	-	0.2000	245.1	PPB	11 Oct 2019 08:37:50	S	Std
2.5 - 1	52.5	9882	-	2.5000	245.1	PPB	11 Oct 2019 08:40:08	S	Std
5.0 - 1	55.0	19249	-	5.0000	245.1	PPB	11 Oct 2019 08:42:26	S	Std
7.5 - 1	57.5	29335	-	7.5000	245.1	PPB	11 Oct 2019 08:44:43	S	Std
10.0 - 1	510.0	38457	-	10.0000	245.1	PPB	11 Oct 2019 08:46:59	S	Std
ICV50 - 1	ICV50	14884	3.8349	-245.1		PPB	11 Oct 2019 08:55:04	U	SMPL
ICB50 - 1	ICB50	103	0.0032	-245.1		PPB	11 Oct 2019 08:57:19	U	SMPL
CCV06 - 1	CCV06	19574	5.0506	-245.1		PPB	11 Oct 2019 08:59:35	U	SMPL
CCB06 - 1	CCB06	164	0.0190	-245.1		PPB	11 Oct 2019 09:01:50	U	SMPL
CRA - 1	CRA	240	0.0387	-245.1		PPB	11 Oct 2019 09:04:05	U	SMPL
HighStd - 1	HighStd	39747	10.2800	-245.1		PPB	11 Oct 2019 09:06:21	U	SMPL
ChkStd - 1	ChkStd	27268	7.0451	-245.1		PPB	11 Oct 2019 09:08:37	U	SMPL
PB123868BL - 1	PBW	-50	-0.0364	-245.1		PPB	11 Oct 2019 09:10:54	U	SMPL
PB123868BS - 1	LCSW	17475	4.5065	-245.1		PPB	11 Oct 2019 09:13:11	U	SMPL
K5174-01 - 1	COMP	537	0.1157	-245.1		PPB	11 Oct 2019 09:15:27	U	SMPL
K5301-02 - 1	EFF-WASTE-WATER	225	0.0348	-245.1		PPB	11 Oct 2019 09:17:44	U	SMPL
K5301-02DUP - 1	EFF-WASTE-WATERDUP	244	0.0398	-245.1		PPB	11 Oct 2019 09:20:01	U	SMPL
K5301-02MS - 1	EFF-WASTE-WATERMS	13976	3.5995	-245.1		PPB	11 Oct 2019 09:22:19	U	SMPL
K5301-02MSD - 1	EFF-WASTE-WATERMSD	13562	3.4922	-245.1		PPB	11 Oct 2019 09:24:35	U	SMPL
CCV07 - 1	CCV07	19170	4.9459	-245.1		PPB	11 Oct 2019 09:26:53	U	SMPL
CCB07 - 1	CCB07	25	-0.0170	-245.1		PPB	11 Oct 2019 09:29:11	U	SMPL
PB123869BL - 1	PBW	53	-0.0097	-245.1		PPB	11 Oct 2019 09:31:27	U	SMPL
PB123869BS - 1	LCSW	16347	4.2141	-245.1		PPB	11 Oct 2019 09:33:43	U	SMPL
K5111-12 - 1	MDL-MDL-WATER-06-QT4-2019	239	0.0385	-245.1		PPB	11 Oct 2019 09:38:56	U	SMPL
K5301-02LX5 - 1		58	-0.0084	-245.1		PPB	11 Oct 2019 09:41:12	U	SMPL
K5301-02A - 1		14227	3.6645	-245.1		PPB	11 Oct 2019 09:43:28	U	SMPL
CCV08 - 1	CCV08	19301	4.9799	-245.1		PPB	11 Oct 2019 09:45:46	U	SMPL
CCB08 - 1	CCB08	-25	-0.0300	-245.1		PPB	11 Oct 2019 09:48:05	U	SMPL