


LB105636

ISM02.4

INSTRUMENT ID: CVI

Linear



A= 0.0000e+000

B= 1.9979e-004

C= -7.5989e-002

Rho= 0.9998437

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.037	-0.037	196	0.000	196					—
0.05	0.050										
0.20	0.200	0.168	-0.032	1219	0.0 %	1219					-16
2.50	2.500	2.537	0.037	13077	0.0 %	13077					—
5.00	5.000	5.038	0.038	25599	0.0 %	25599					—
7.50	7.500	7.598	0.098	38409	0.0 %	38409					—
10.0	10.000	9.896	-0.104	49914	0.0 %	49914					-1

RS 10.11.19

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	196	-		0 ISM02.4	PPB	10/11/2019 10:09	Std
	0.2 S0.2	1219	-		0.2 ISM02.4	PPB	10/11/2019 10:11	Std
	2.5 S2.5	13077	-		2.5 ISM02.4	PPB	10/11/2019 10:13	Std
	5 S5.0	25599	-		5 ISM02.4	PPB	10/11/2019 10:16	Std
	7.5 S7.5	38409	-		7.5 ISM02.4	PPB	10/11/2019 10:18	Std
	10 S10.0	49914	-		10 ISM02.4	PPB	10/11/2019 10:20	Std
ICV	ICV	19289	3.7778 -		ISM02.4	PPB	10/11/2019 10:24	SMPL
ICB	ICB	135	-0.049 -		ISM02.4	PPB	10/11/2019 10:26	SMPL
CCV009	CCV009	25550	5.0287 -		ISM02.4	PPB	10/11/2019 10:28	SMPL
CCB009	CCB009	137	-0.0486 -		ISM02.4	PPB	10/11/2019 10:31	SMPL
PB123876BL	PBS876	166	-0.0428 -		ISM02.4	PPB	10/11/2019 10:33	SMPL
K5098-01	MJLMA2	613	0.0465 -		ISM02.4	PPB	10/11/2019 10:35	SMPL
K5098-02	MJLMA3	745	0.0729 -		ISM02.4	PPB	10/11/2019 10:37	SMPL
K5098-03	MJLMA4	1911	0.3058 -		ISM02.4	PPB	10/11/2019 10:40	SMPL
K5098-04	MJLMA5	1400	0.2037 -		ISM02.4	PPB	10/11/2019 10:42	SMPL
K5098-05	MJLMD9	943	0.1124 -		ISM02.4	PPB	10/11/2019 10:44	SMPL
K5098-06	MJLME1	1274	0.1785 -		ISM02.4	PPB	10/11/2019 10:46	SMPL
K5098-07	MJLME2	2259	0.3753 -		ISM02.4	PPB	10/11/2019 10:49	SMPL
K5098-08	MJLME4	1086	0.141 -		ISM02.4	PPB	10/11/2019 10:51	SMPL
K5098-09	MJLME5	2088	0.3412 -		ISM02.4	PPB	10/11/2019 10:53	SMPL
K5098-10	MJLME6	1654	0.2545 -		ISM02.4	PPB	10/11/2019 10:56	SMPL
K5098-11	MJLME7	2572	0.4379 -		ISM02.4	PPB	10/11/2019 10:58	SMPL
K5098-12	MJLME9	1653	0.2543 -		ISM02.4	PPB	10/11/2019 11:00	SMPL
K5098-13	MJLMF0	1900	0.3036 -		ISM02.4	PPB	10/11/2019 11:02	SMPL
K5098-14	MJLMF1	1598	0.2433 -		ISM02.4	PPB	10/11/2019 11:05	SMPL
K5098-15	MJLMF2	2269	0.3773 -		ISM02.4	PPB	10/11/2019 11:07	SMPL
K5098-16	MJLMF4	917	0.1072 -		ISM02.4	PPB	10/11/2019 11:09	SMPL
K5098-17	MJLMF5	2351	0.3937 -		ISM02.4	PPB	10/11/2019 11:11	SMPL
K5098-18	MJLMF6	1430	0.2097 -		ISM02.4	PPB	10/11/2019 11:14	SMPL
K5098-19	MJLMF7	2438	0.4111 -		ISM02.4	PPB	10/11/2019 11:16	SMPL
CCV010	CCV010	25482	5.0151 -		ISM02.4	PPB	10/11/2019 11:18	SMPL
CCB010	CCB010	129	-0.0502 -		ISM02.4	PPB	10/11/2019 11:21	SMPL
K5098-20	MJLM82	1495	0.2227 -		ISM02.4	PPB	10/11/2019 11:23	SMPL
K5098-21	MJLM82D	1507	0.2251 -		ISM02.4	PPB	10/11/2019 11:25	SMPL
K5098-22	MJLM82S	12912	2.5037 -		ISM02.4	PPB	10/11/2019 11:27	SMPL
PB123877BL	PBS877	146	-0.0468 -		ISM02.4	PPB	10/11/2019 11:30	SMPL
K5101-01	MCOB28	1355	0.1947 -		ISM02.4	PPB	10/11/2019 11:32	SMPL
K5101-02	MCOB29	1060	0.1358 -		ISM02.4	PPB	10/11/2019 11:34	SMPL
K5101-03	MCOB30	2235	0.3705 -		ISM02.4	PPB	10/11/2019 11:37	SMPL
K5101-04	MCOB31	1796	0.2828 -		ISM02.4	PPB	10/11/2019 11:39	SMPL
K5101-05	MCOB32	1670	0.2577 -		ISM02.4	PPB	10/11/2019 11:41	SMPL
K5101-06	MCOB33	2043	0.3322 -		ISM02.4	PPB	10/11/2019 11:44	SMPL
K5101-07	MCOB34	1791	0.2818 -		ISM02.4	PPB	10/11/2019 11:46	SMPL
K5101-08	MCOB35	699	0.0637 -		ISM02.4	PPB	10/11/2019 11:48	SMPL
K5101-09	MCOB35D	700	0.0639 -		ISM02.4	PPB	10/11/2019 11:50	SMPL
K5101-10	MCOB35S	12183	2.3581 -		ISM02.4	PPB	10/11/2019 11:58	SMPL

LB105636
INSTRUMENT ID : CV1

K5101-11	MC0B36	520	0.0279 -	ISM02.4	PPB	10/11/2019 12:00	SMPL
K5101-12	MC0B37	3287	0.5807 -	ISM02.4	PPB	10/11/2019 12:02	SMPL
K5101-13	MC0B38	1964	0.3164 -	ISM02.4	PPB	10/11/2019 12:05	SMPL
K5101-14	MC0B39	1706	0.2649 -	ISM02.4	PPB	10/11/2019 12:07	SMPL
K5101-15	MC0B40	1652	0.2541 -	ISM02.4	PPB	10/11/2019 12:09	SMPL
K5101-16	MC0B41	2082	0.34 -	ISM02.4	PPB	10/11/2019 12:11	SMPL
CCV011	CCV011	25559	5.0305 -	ISM02.4	PPB	10/11/2019 12:14	SMPL
CCB011	CCB011	128	-0.0504 -	ISM02.4	PPB	10/11/2019 12:16	SMPL
K5101-17	MC0B42	1006	0.125 -	ISM02.4	PPB	10/11/2019 12:18	SMPL
K5101-18	MC0B43	1048	0.1334 -	ISM02.4	PPB	10/11/2019 12:21	SMPL
K5101-19	MC0B44	1488	0.2213 -	ISM02.4	PPB	10/11/2019 12:23	SMPL
K5101-20	MC0B45	1114	0.1466 -	ISM02.4	PPB	10/11/2019 12:25	SMPL
K5101-21	MC0B46	1020	0.1278 -	ISM02.4	PPB	10/11/2019 12:27	SMPL
K5101-22	MC0B47	1380	0.1997 -	ISM02.4	PPB	10/11/2019 12:30	SMPL
PB123878BL	PBS878	107	-0.0546 -	ISM02.4	PPB	10/11/2019 12:32	SMPL
K5102-01	MBEPA9	60279	11.9672 -	ISM02.4	PPB	10/11/2019 12:34	SMPL
K5102-02	MBEPB0	9664	1.8548 -	ISM02.4	PPB	10/11/2019 12:36	SMPL
K5102-03	MBEPB1	8492	1.6206 -	ISM02.4	PPB	10/11/2019 12:39	SMPL
K5102-04	MBEPB2	9867	1.8953 -	ISM02.4	PPB	10/11/2019 12:41	SMPL
K5102-05	MBEPB3	14208	2.7626 -	ISM02.4	PPB	10/11/2019 12:43	SMPL
K5102-06	MBEPC4	24475	4.8139 -	ISM02.4	PPB	10/11/2019 12:46	SMPL
K5102-07	MBEPC5	41805	8.2763 -	ISM02.4	PPB	10/11/2019 12:48	SMPL
K5102-08	MBEPK6	1900	0.3036 -	ISM02.4	PPB	10/11/2019 12:50	SMPL
K5102-09	MBEPK7	1776	0.2788 -	ISM02.4	PPB	10/11/2019 12:52	SMPL
K5102-10	MBEPK8	4158	0.7547 -	ISM02.4	PPB	10/11/2019 12:55	SMPL
K5102-11	MBEPK9	7358	1.3941 -	ISM02.4	PPB	10/11/2019 12:57	SMPL
K5102-12	MBEPL0	988	0.1214 -	ISM02.4	PPB	10/11/2019 12:59	SMPL
K5102-13	MBEPL1	1140	0.1518 -	ISM02.4	PPB	10/11/2019 13:02	SMPL
CCV012	CCV012	25539	5.0265 -	ISM02.4	PPB	10/11/2019 13:04	SMPL
CCB012	CCB012	130	-0.05 -	ISM02.4	PPB	10/11/2019 13:06	SMPL
K5102-14	MBEPB4	13760	2.6731 -	ISM02.4	PPB	10/11/2019 13:08	SMPL
K5102-15	MBEPB5	10654	2.0526 -	ISM02.4	PPB	10/11/2019 13:11	SMPL
K5102-16	MBEPB6	18115	3.5432 -	ISM02.4	PPB	10/11/2019 13:13	SMPL
K5102-17	MBEPB7	16971	3.3147 -	ISM02.4	PPB	10/11/2019 13:15	SMPL
K5102-18	MBEPB8	24072	4.7334 -	ISM02.4	PPB	10/11/2019 13:17	SMPL
K5102-19	MBEPB9	19321	3.7842 -	ISM02.4	PPB	10/11/2019 13:20	SMPL
K5102-20	MBEPB9D	19330	3.786 -	ISM02.4	PPB	10/11/2019 13:22	SMPL
K5102-21	MBEPB9S	30426	6.0028 -	ISM02.4	PPB	10/11/2019 13:24	SMPL
K5102-22	MBEPC1	22335	4.3863 -	ISM02.4	PPB	10/11/2019 13:27	SMPL
PB123879BL	PBS879	85	-0.059 -	ISM02.4	PPB	10/11/2019 13:29	SMPL
K5126-01	MJLM50	2045	0.3326 -	ISM02.4	PPB	10/11/2019 13:31	SMPL
K5126-02	MJLM51	2088	0.3412 -	ISM02.4	PPB	10/11/2019 13:33	SMPL
K5126-03	MJLM54	46602	9.2347 -	ISM02.4	PPB	10/11/2019 13:36	SMPL
K5126-04	MJLM66	11453	2.2122 -	ISM02.4	PPB	10/11/2019 13:38	SMPL
K5126-05	MJLM67	1457	0.2151 -	ISM02.4	PPB	10/11/2019 13:40	SMPL
K5126-06	MJLM74	1573	0.2383 -	ISM02.4	PPB	10/11/2019 13:43	SMPL
K5126-07	MJLM75	7903	1.503 -	ISM02.4	PPB	10/11/2019 13:45	SMPL

INSTRUMENT ID : CV1

K5126-08	MJLM78	5569	1.0366 -	ISM02.4 PPB	10/11/2019 13:47 SMPL
K5126-09	MJLM79	34276	6.772 -	ISM02.4 PPB	10/11/2019 13:49 SMPL
K5126-10	MJLM80	36727	7.2617 -	ISM02.4 PPB	10/11/2019 13:52 SMPL
CCV013	CCV013	25715	5.0616 -	ISM02.4 PPB	10/11/2019 13:54 SMPL
CCB013	CCB013	8	-0.0744 -	ISM02.4 PPB	10/11/2019 13:56 SMPL
K5126-11	MJLM83	4399	0.8029 -	ISM02.4 PPB	10/11/2019 13:59 SMPL
K5126-12	MJLM83D	4413	0.8057 -	ISM02.4 PPB	10/11/2019 14:01 SMPL
K5126-13	MJLM83S	17278	3.376 -	ISM02.4 PPB	10/11/2019 14:03 SMPL
K5126-14	MJLM90	1305	0.1847 -	ISM02.4 PPB	10/11/2019 14:05 SMPL
K5126-15	MJLM91	157	-0.0446 -	ISM02.4 PPB	10/11/2019 14:15 SMPL
K5126-16	MJLM92	134	-0.0492 -	ISM02.4 PPB	10/11/2019 14:17 SMPL
K5126-17	MJLMA6	1683	0.2603 -	ISM02.4 PPB	10/11/2019 14:19 SMPL
K5126-18	MJLMA7	1796	0.2828 -	ISM02.4 PPB	10/11/2019 14:21 SMPL
K5126-19	MJLMA8	1850	0.2936 -	ISM02.4 PPB	10/11/2019 14:24 SMPL
K5102-01DLX2		30361	5.9899 -	ISM02.4 PPB	10/11/2019 14:26 SMPL
CCV014	CCV014	25840	5.0866 -	ISM02.4 PPB	10/11/2019 14:28 SMPL
CCB014	CCB014	70	-0.062 -	ISM02.4 PPB	10/11/2019 14:31 SMPL