



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105655

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/11/2019	15:25
2	CAL2	1	Water	NA	NA	20.1	7.01	10/11/2019	15:26
3	CAL3	1	Water	NA	NA	20.1	10.04	10/11/2019	15:27
4	ICV	1	Water	NA	NA	20.1	6.99	10/11/2019	15:30
5	CCV1	1	Water	NA	NA	20.1	2.01	10/11/2019	15:32
6	K5337-01	1	Water	NA	NA	22.0	7.04	10/11/2019	15:35
7	K5337-01DUP	1	Water	NA	NA	22.1	7.05	10/11/2019	15:36
8	CCV2	1	Water	NA	NA	20.1	12.02	10/11/2019	15:39

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5337

Department : Wet-Chemistry

WorkList ID : 132017

Date : 10-11-2019 15:18:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2019	Water	K5337-01-17	pH	Cool 4 deg C	PSEG01	N33	HHT-WATER	10/11/2019	9040C

Date/Time 10-11-19 15:20
 Received by: J2
 Relinquished by: J2

Date/Time 10-11-19 16:30
 Received by: J2
 Relinquished by: J2