



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105674

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/14/2019	11:40
2	K5337-02		1	8	100.00	>212.0	>212.0	10/14/2019	12:10
3	K5337-02DUP		1	8	100.00	>212.0	>212.0	10/14/2019	12:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

LB 105674

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp101419 WorkList ID : 132059 Department : Wet-Chemistry Date : 10-14-2019 08:44:26

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/17/2019	Water	K5337-02 C	Flash Point	Cool 4 deg C	PSEG01	N33	HHT-WATER	10/11/2019	1010A

Date/Time 10/14/19 11:15
Received by: KS
Relinquished by: JP

Date/Time 10/14/19 12:50
Received by: JP
Relinquished by: KS