



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105677

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/14/2019	13:20
2	CAL2	1	Water	NA	NA	20.1	7.00	10/14/2019	13:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/14/2019	13:22
4	ICV	1	Water	NA	NA	20.1	7.01	10/14/2019	13:25
5	CCV1	1	Water	NA	NA	20.1	2.02	10/14/2019	13:26
6	K5326-02	1	Water	NA	NA	17.5	7.45	10/14/2019	13:27
7	K5326-02DUP	1	Water	NA	NA	17.6	7.46	10/14/2019	13:28
8	CCV2	1	Water	NA	NA	20.1	12.02	10/14/2019	13:30

WORKLIST(Hardcopy Internal Chain)

20105677

WorkList Name : CORROSIONITY K5326 V_h WorkList ID : 132094 Department : Wet-Chemistry Date : 10-14-2019 13:00:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2019	Water	K5326-02-10	Corrosivity	Cool 4 deg C	HALE01	L31	IDW-AQ-101019	10/10/2019	9040C

Date/Time 10.14.19 13:10
 Received by: JA
 Relinquished by: CA

Date/Time 10.14.19 14:00
 Received by: CA
 Relinquished by: JA