

7470A


LB105681

INSTRUMENT ID : CV1

Linear

 μ Abs.:

42266



A= 0.0000e+000

B= 2.3546e-004

C= -7.4003e-004

Rho= 0.9995082

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.013	0.013	58	0.000	58				
0.2	0.200	0.206	0.006	880	0.0 %	880				
2.5	2.500	2.568	0.068	10908	0.0 %	10908				
5.0	5.000	4.790	-0.210	20348	0.0 %	20348				
7.5	7.500	7.671	0.171	32582	0.0 %	32582				
10.0	10.000	9.951	-0.049	42266	0.0 %	42266				

LB105681 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 14 Oct 2019 14:06:28				
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type	Type
0.0 - 1	50-0	58	-	0.00007470A	PPB	14 Oct 2019 14:30:23	S	Std
0.2 - 1	50-2	880	-	0.20007470A	PPB	14 Oct 2019 14:32:32	S	Std
2.5 - 1	52-5	10908	-	2.50007470A	PPB	14 Oct 2019 14:34:41	S	Std
5.0 - 1	55-0	20348	-	5.00007470A	PPB	14 Oct 2019 14:36:51	S	Std
7.5 - 1	57-5	32582	-	7.50007470A	PPB	14 Oct 2019 14:39:00	S	Std
10.0 - 1	510-0	42266	-	10.00007470A	PPB	14 Oct 2019 14:41:08	S	Std
ICV54 - 1	ICV54	16902	3.9791	-7470A	PPB	14 Oct 2019 14:55:53	U	SMPL
ICB54 - 1	ICB54	41	0.0089	-7470A	PPB	14 Oct 2019 14:58:00	U	SMPL
CCV29 - 1	CCV29	20032	4.7161	-7470A	PPB	14 Oct 2019 15:00:07	U	SMPL
CCB29 - 1	CCB29	27	0.0056	-7470A	PPB	14 Oct 2019 15:02:14	U	SMPL
CRA - 1	CRA	881	0.2067	-7470A	PPB	14 Oct 2019 15:04:21	U	SMPL
HighStd - 1	HighStd	42196	9.9349	-7470A	PPB	14 Oct 2019 15:06:30	U	SMPL
ChkStd - 1	ChkStd	29640	6.9784	-7470A	PPB	14 Oct 2019 15:08:45	U	SMPL
PB123960BL - 1	PBW	-11	-0.0033	-7470A	PPB	14 Oct 2019 15:10:53	U	SMPL
PB123960BS - 1	LCSW	17853	4.2030	-7470A	PPB	14 Oct 2019 15:13:01	U	SMPL
K5266-13 - 1	AMENDED-COMPOSITE-A	791	0.1855	-7470A	PPB	14 Oct 2019 15:15:08	U	SMPL
K5266-13DUP - 1	AMENDED-COMPOSITE-ADUP	828	0.1942	-7470A	PPB	14 Oct 2019 15:17:16	U	SMPL
K5266-13MS - 1	AMENDED-COMPOSITE-AMS	19094	4.4952	-7470A	PPB	14 Oct 2019 15:19:25	U	SMPL
K5266-13MSD - 1	AMENDED-COMPOSITE-AMSD	18675	4.3965	-7470A	PPB	14 Oct 2019 15:21:34	U	SMPL
K5266-14 - 1	AMENDED-COMPOSITE-C	483	0.1130	-7470A	PPB	14 Oct 2019 15:23:41	U	SMPL
CCV30 - 1	CCV30	20124	4.7377	-7470A	PPB	14 Oct 2019 15:25:48	U	SMPL
CCB30 - 1	CCB30	20	0.0040	-7470A	PPB	14 Oct 2019 15:27:55	U	SMPL
K5292-06 - 1	20190855-AMENDED-COMP-B	1892	0.4448	-7470A	PPB	14 Oct 2019 15:30:02	U	SMPL
K5292-07 - 1	20180657-AMENDED-COMP-D	1046	0.2456	-7470A	PPB	14 Oct 2019 15:32:09	U	SMPL
K5292-08 - 1	20190858-AMENDED-COMP-E	1578	0.3708	-7470A	PPB	14 Oct 2019 15:34:16	U	SMPL
K5293-02 - 1	KEARNY-TOPSOIL	127	0.0292	-7470A	PPB	14 Oct 2019 15:36:23	U	SMPL
PB123960TB - 1	PB123826TB	34	0.0073	-7470A	PPB	14 Oct 2019 15:38:31	U	SMPL
K5266-13LX5 - 1		151	0.0348	-7470A	PPB	14 Oct 2019 15:40:38	U	SMPL
K5266-13A - 1		19056	4.4862	-7470A	PPB	14 Oct 2019 15:42:47	U	SMPL
CCV31 - 1	CCV31	20101	4.7323	-7470A	PPB	14 Oct 2019 15:44:57	U	SMPL
CCB31 - 1	CCB31	91	0.0207	-7470A	PPB	14 Oct 2019 15:47:06	U	SMPL