



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105682

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP78922
Residual chlorine LOD, 0.05PPM	WP78920
Residual chlorine ICV-LCS, 0.4PPM	WP78919
Chlorine Calibration std, 0.1ppm	WP78914
Chlorine Calibration std, 0.2ppm	WP78915
Chlorine Calibration std, 0.8ppm	WP78916
Chlorine Calibration std, 0.0ppm	WP78913
Chlorine Calibration std, 1.6ppm	WP78917
Residual Chlorine LOQ 0.1ppm	WP78921

Intercept: -0.0006

Slope: 1.0076

Regression: 0.999992

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	10/14/2019	13:28
2	CAL2	0.1	1	0.000	0.100	0.100	10/14/2019	13:29
3	CAL3	0.2	1	0.000	0.200	0.200	10/14/2019	13:29
4	CAL4	0.4	1	0.000	0.400	0.400	10/14/2019	13:30
5	CAL5	0.8	1	0.000	0.810	0.810	10/14/2019	13:30
6	CAL6	1.6	1	0.000	1.610	1.610	10/14/2019	13:32



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105682

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3880	10/14/2019	13:32
2	ICB				1	0.0000	0.0000	0.0000	0.0010	10/14/2019	13:33
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4080	10/14/2019	13:33
4	CCB1				1	0.0000	0.0000	0.0000	0.0010	10/14/2019	13:34
5	LB105682BL	50	50		1	0.0000	0.0000	0.0000	0.0010	10/14/2019	13:34
6	LB105682BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3980	10/14/2019	13:35
7	K5111-07	50	50		1	0.0000	0.0510	0.0510	0.0510	10/14/2019	13:35
8	K5111-08	50	50		1	0.0000	0.1200	0.1200	0.1200	10/14/2019	13:36
9	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4080	10/14/2019	13:36
10	CCB2				1	0.0000	0.0000	0.0000	0.0010	10/14/2019	13:37