



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105683

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/14/2019	15:00
2	CAL2	1	Water	NA	NA	20.1	7.00	10/14/2019	15:01
3	CAL3	1	Water	NA	NA	20.1	10.02	10/14/2019	15:02
4	ICV	1	Water	NA	NA	20.1	6.99	10/14/2019	15:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/14/2019	15:06
6	K5326-01	1	Solid	20.02	20	21.2	9.13	10/14/2019	15:08
7	K5355-07	1	Solid	20.03	20	21.6	10.19	10/14/2019	15:10
8	K5355-08	1	Solid	20.04	20	21.5	10.19	10/14/2019	15:11
9	K5355-09	1	Solid	20.03	20	21.3	10.26	10/14/2019	15:15
10	K5355-09DUP	1	Solid	20.04	20	21.4	10.27	10/14/2019	15:16
11	CCV2	1	Water	NA	NA	20.1	12.02	10/14/2019	15:25

✓ 105683

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5356 WorkList ID : 132092 Department : Wet-Chemistry Date : 10-14-2019 12:59:42

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/10/2019	Solid	K5326-01	✓ Corrosivity	Cool 4 deg C	HALE01	L31	IDW-SO-101019	10/10/2019	9045D
10/11/2019	Solid	K5355-07	✓ Corrosivity	Cool 4 deg C	TECT01		COMP-31	10/11/2019	9045D
10/11/2019	Solid	K5355-08	✓ Corrosivity	Cool 4 deg C	TECT01		COMP-32	10/11/2019	9045D
10/11/2019	Solid	K5355-09	✓ Corrosivity	Cool 4 deg C	TECT01		COMP-33	10/11/2019	9045D

Date/Time 10.14.19 13:00
 Received by: Jo
 Relinquished by: Jo

Date/Time 10.14.19 17:00
 Received by: CP
 Relinquished by: Jo