


LB10569p

ISM02.4

INSTRUMENT ID : CV1



A= 0.0000e+000

B= 2.1857e-004

C= -1.8688e-002

Rho= 0.9991247

Accept=Accepted

slope
y intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.004	-0.004	69	0.000	69				
0.05	0.050									
0.20	0.200	0.204	0.004	1017	0.0 %	1017				
2.50	2.500	2.573	0.073	11859	0.0 %	11859				
5.00	5.000	5.079	0.079	23323	0.0 %	23323				
7.50	7.500	7.179	-0.321	32931	0.0 %	32931				
10.0	10.000	10.168	0.168	46607	0.0 %	46607				

2.18

2
3
2
-4
2

ns 10.15.19

LB105696
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	69	-	0	ISM02.4	PPB	10/15/2019 9:26	Std
0.2 S0.2	1017	-	0.2	ISM02.4	PPB	10/15/2019 9:29	Std
2.5 S2.5	11859	-	2.5	ISM02.4	PPB	10/15/2019 9:31	Std
5 S5.0	23323	-	5	ISM02.4	PPB	10/15/2019 9:33	Std
7.5 S7.5	32931	-	7.5	ISM02.4	PPB	10/15/2019 9:36	Std
10 S10.0	46607	-	10	ISM02.4	PPB	10/15/2019 9:38	Std
ICV	ICV	17414	3.7876	-	ISM02.4	PPB	10/15/2019 9:43 SMPL
ICB	ICB	34	-0.0113	-	ISM02.4	PPB	10/15/2019 9:45 SMPL
CCV032	CCV032	22725	4.9484	-	ISM02.4	PPB	10/15/2019 9:47 SMPL
CCB032	CCB032	32	-0.0117	-	ISM02.4	PPB	10/15/2019 9:50 SMPL
PB123975E	PBS975	72	-0.003	-	ISM02.4	PPB	10/15/2019 9:52 SMPL
K5161-01	MH0GF5	2737	0.5795	-	ISM02.4	PPB	10/15/2019 9:54 SMPL
K5161-02	MH0GF6	1144	0.2314	-	ISM02.4	PPB	10/15/2019 9:56 SMPL
K5161-03	MH0GF7	1024	0.2051	-	ISM02.4	PPB	10/15/2019 9:59 SMPL
K5161-04	MH0GF8	467	0.0834	-	ISM02.4	PPB	10/15/2019 10:01 SMPL
K5161-05	MH0GF9	551	0.1017	-	ISM02.4	PPB	10/15/2019 10:03 SMPL
K5161-06	MH0GG0	418	0.0727	-	ISM02.4	PPB	10/15/2019 10:05 SMPL
K5161-07	MH0GG1	1209	0.2456	-	ISM02.4	PPB	10/15/2019 10:08 SMPL
K5161-08	MH0GG2	2356	0.4963	-	ISM02.4	PPB	10/15/2019 10:10 SMPL
K5161-09	MH0GG3	3290	0.7004	-	ISM02.4	PPB	10/15/2019 10:12 SMPL
K5161-10	MH0GG4	1918	0.4005	-	ISM02.4	PPB	10/15/2019 10:15 SMPL
K5161-11	MH0GG5	2573	0.5437	-	ISM02.4	PPB	10/15/2019 10:17 SMPL
K5161-12	MH0GG6	7101	1.5334	-	ISM02.4	PPB	10/15/2019 10:19 SMPL
K5161-13	MH0GG7	9780	2.119	-	ISM02.4	PPB	10/15/2019 10:21 SMPL
K5161-14	MH0GG8	6497	1.4014	-	ISM02.4	PPB	10/15/2019 10:24 SMPL
K5161-15	MH0GG8D	6539	1.4106	-	ISM02.4	PPB	10/15/2019 10:26 SMPL
K5161-16	MH0GG8S	21294	4.6356	-	ISM02.4	PPB	10/15/2019 10:28 SMPL
K5161-17	MH0GG9	3866	0.8263	-	ISM02.4	PPB	10/15/2019 10:30 SMPL
K5161-18	MH0GH0	9352	2.0254	-	ISM02.4	PPB	10/15/2019 10:33 SMPL
K5161-19	MH0GH1	13304	2.8892	-	ISM02.4	PPB	10/15/2019 10:35 SMPL
CCV033	CCV033	23266	5.0666	-	ISM02.4	PPB	10/15/2019 10:37 SMPL
CCB033	CCB033	17	-0.015	-	ISM02.4	PPB	10/15/2019 10:40 SMPL
K5161-20	MH0GH2	4433	0.9502	-	ISM02.4	PPB	10/15/2019 10:42 SMPL
K5161-21	MH0GH3	6619	1.4281	-	ISM02.4	PPB	10/15/2019 10:44 SMPL
K5161-22	MH0GH4	1489	0.3068	-	ISM02.4	PPB	10/15/2019 10:46 SMPL
PB123976E	PBS976	25	-0.0132	-	ISM02.4	PPB	10/15/2019 10:49 SMPL
K5162-02	MH0GH6	1742	0.3621	-	ISM02.4	PPB	10/15/2019 10:53 SMPL
K5162-03	MH0GH7	55386	12.0872	-	ISM02.4	PPB	10/15/2019 10:55 SMPL
K5162-04	MH0GH8	3867	0.8265	-	ISM02.4	PPB	10/15/2019 10:58 SMPL
K5162-05	MH0GH9	7115	1.5365	-	ISM02.4	PPB	10/15/2019 11:00 SMPL
K5162-06	MH0GJ0	1490	0.307	-	ISM02.4	PPB	10/15/2019 11:02 SMPL
K5162-07	MH0GJ1	1217	0.2473	-	ISM02.4	PPB	10/15/2019 11:05 SMPL
K5162-08	MH0GJ2	4062	0.8692	-	ISM02.4	PPB	10/15/2019 11:07 SMPL
K5162-09	MH0GJ3	204	0.0259	-	ISM02.4	PPB	10/15/2019 11:09 SMPL
K5162-10	MH0GJ4	2640	0.5583	-	ISM02.4	PPB	10/15/2019 11:11 SMPL
K5162-11	MH0GJ5	1680	0.3485	-	ISM02.4	PPB	10/15/2019 11:14 SMPL

INSTRUMENT ID : CV1

K5162-12	MH0GJ6	3467	0.7391 -	ISM02.4	PPB	10/15/2019 11:16 SMPL
K5162-13	MH0GJ7	1487	0.3063 -	ISM02.4	PPB	10/15/2019 11:18 SMPL
K5162-14	MH0GJ8	1075	0.2163 -	ISM02.4	PPB	10/15/2019 11:20 SMPL
K5162-15	MH0GJ9	1456	0.2996 -	ISM02.4	PPB	10/15/2019 11:23 SMPL
K5162-16	MH0GK0	294	0.0456 -	ISM02.4	PPB	10/15/2019 11:25 SMPL
CCV034	CCV034	22851	4.9759 -	ISM02.4	PPB	10/15/2019 11:27 SMPL
CCB034	CCB034	49	-0.008 -	ISM02.4	PPB	10/15/2019 11:30 SMPL
K5162-17	MH0GK0D	365	0.0611 -	ISM02.4	PPB	10/15/2019 11:32 SMPL
K5162-18	MH0GK0S	14564	3.1646 -	ISM02.4	PPB	10/15/2019 11:34 SMPL
K5162-19	MH0GK1	129	0.0095 -	ISM02.4	PPB	10/15/2019 11:36 SMPL
K5162-20	MH0GK2	118	0.0071 -	ISM02.4	PPB	10/15/2019 11:39 SMPL
K5162-21	MH0GK3	399	0.0685 -	ISM02.4	PPB	10/15/2019 11:41 SMPL
K5162-22	MH0GK4	6023	1.2978 -	ISM02.4	PPB	10/15/2019 11:43 SMPL
PB123977E	PBS977	34	-0.0113 -	ISM02.4	PPB	10/15/2019 11:46 SMPL
K5179-01	MBEP83	22671	4.9366 -	ISM02.4	PPB	10/15/2019 11:48 SMPL
K5179-02	MBEP84	22041	4.7989 -	ISM02.4	PPB	10/15/2019 11:50 SMPL
K5179-03	MBEP84D	22154	4.8236 -	ISM02.4	PPB	10/15/2019 11:52 SMPL
K5179-04	MBEP84S	35150	7.6642 -	ISM02.4	PPB	10/15/2019 11:55 SMPL
K5179-05	MBEP85	1480	0.3048 -	ISM02.4	PPB	10/15/2019 11:57 SMPL
K5179-06	MBEP86	1754	0.3647 -	ISM02.4	PPB	10/15/2019 11:59 SMPL
K5179-07	MBEP87	1807	0.3763 -	ISM02.4	PPB	10/15/2019 12:02 SMPL
K5179-08	MBEP88	1473	0.3033 -	ISM02.4	PPB	10/15/2019 12:04 SMPL
K5179-09	MBEP89	11510	2.4971 -	ISM02.4	PPB	10/15/2019 12:06 SMPL
K5179-10	MBEPF7	34345	7.4882 -	ISM02.4	PPB	10/15/2019 12:08 SMPL
K5179-11	MBEPF8	14398	3.1283 -	ISM02.4	PPB	10/15/2019 12:11 SMPL
K5179-12	MBEPF9	13679	2.9712 -	ISM02.4	PPB	10/15/2019 12:13 SMPL
K5179-13	MBEPG0	8504	1.8401 -	ISM02.4	PPB	10/15/2019 12:15 SMPL
CCV035	CCV035	24123	5.254 -	ISM02.4	PPB	10/15/2019 12:18 SMPL
CCB035	CCB035	37	-0.0106 -	ISM02.4	PPB	10/15/2019 12:20 SMPL
K5179-14	MBEPG1	9971	2.1607 -	ISM02.4	PPB	10/15/2019 12:22 SMPL
K5179-15	MBEPG2	20585	4.4806 -	ISM02.4	PPB	10/15/2019 12:26 SMPL
K5179-16	MBEPG3	31773	6.926 -	ISM02.4	PPB	10/15/2019 12:32 SMPL
K5179-17	MBEPG4	10067	2.1817 -	ISM02.4	PPB	10/15/2019 12:34 SMPL
K5179-18	MBEP90	15214	3.3067 -	ISM02.4	PPB	10/15/2019 12:36 SMPL
K5179-19	MBEP91	7378	1.5939 -	ISM02.4	PPB	10/15/2019 12:39 SMPL
K5179-20	MBEP92	8761	1.8962 -	ISM02.4	PPB	10/15/2019 12:41 SMPL
K5179-21	MBEPG5	18659	4.0597 -	ISM02.4	PPB	10/15/2019 12:43 SMPL
K5179-22	MBEPG6	1642	0.3402 -	ISM02.4	PPB	10/15/2019 12:45 SMPL
K5162-01	MH0GH5	3315	0.7059 -	ISM02.4	PPB	10/15/2019 12:48 SMPL
K5162-03DLX2		26579	5.7908 -	ISM02.4	PPB	10/15/2019 12:50 SMPL
CCV036	CCV036	24052	5.2384 -	ISM02.4	PPB	10/15/2019 12:52 SMPL
CCB036	CCB036	20	-0.0143 -	ISM02.4	PPB	10/15/2019 12:55 SMPL