



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105700

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/15/2019	10:58
2	K5326-02		1	9	100.00	>212.0	>212.0	10/15/2019	11:25
3	K5326-02DUP		1	9	100.00	>212.0	>212.0	10/15/2019	11:50

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB105700

WorkList Name : fp101519 WorkList ID : 132135 Department : Wet-Chemistry Date : 10-15-2019 09:02:47

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2019	Water	K5326-02	Flash Point	Cool 4 deg C	HALE01	L31	IDW-AQ-101019	10/10/2019	1010A

Date/Time 10/15/19 10:30
 Received by: KS
 Relinquished by: JP

Date/Time 10/15/19 11:55
 Received by: JP
 Relinquished by: KS