



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105702

Reagent/Standard	Lot/Log #
Residual chlorine std, and CCV , 0.4PPM	WP78980
Residual chlorine LOD, 0.05PPM	WP78982
Residual chlorine ICV-LCS, 0.4PPM	WP78981
Chlorine Calibration std, 0.1ppm	WP78976
Chlorine Calibration std, 0.2ppm	WP78977
Chlorine Calibration std, 0.8ppm	WP78978
Chlorine Calibration std, 0.0ppm	WP78975
Chlorine Calibration std, 1.6ppm	WP78979
Total Chlorine Powder Pillows	W2249

Intercept: 0.0014

Slope: 1.0069

Regression: 0.999983

Seq	Lab ID	True Val (mg/l)	Dilution	Initial Reading	Final Reading	Difference	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	10/15/2019	12:18
2	CAL2	0.1	1	0.000	0.100	0.100	10/15/2019	12:18
3	CAL3	0.2	1	0.000	0.200	0.200	10/15/2019	12:19
4	CAL4	0.4	1	0.000	0.410	0.410	10/15/2019	12:20
5	CAL5	0.8	1	0.000	0.810	0.810	10/15/2019	12:20
6	CAL6	1.6	1	0.000	1.610	1.610	10/15/2019	12:21



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: AMANDEEP

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: apatel

Run Number: LB105702

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.3900	0.3900	0.3840	10/15/2019	12:21
2	ICB				1	0.0000	0.0000	0.0000	-0.0040	10/15/2019	12:21
3	CCV1			0.4	1	0.0000	0.4000	0.4000	0.3940	10/15/2019	12:22
4	CCB1				1	0.0000	0.0000	0.0000	-0.0040	10/15/2019	12:22
5	LB105702BL	50	50		1	0.0000	0.0000	0.0000	-0.0040	10/15/2019	12:23
6	LB105702BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.3940	10/15/2019	12:23
7	K5111-09	50	50		1	0.0000	0.0500	0.0500	0.0460	10/15/2019	12:24
8	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4040	10/15/2019	12:24
9	CCB2				1	0.0000	0.0000	0.0000	-0.0040	10/15/2019	12:25