



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105703

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/15/2019	11:20
2	CAL2	1	Water	NA	NA	20.1	6.99	10/15/2019	11:21
3	CAL3	1	Water	NA	NA	20.1	10.03	10/15/2019	11:22
4	ICV	1	Water	NA	NA	20.1	7.01	10/15/2019	11:24
5	CCV1	1	Water	NA	NA	20.1	2.02	10/15/2019	11:25
6	K5372-01	1	Water	NA	NA	19.3	5.01	10/15/2019	11:26
7	K5372-01DUP	1	Water	NA	NA	19.4	5.02	10/15/2019	11:27
8	K5372-02	1	Water	NA	NA	19.4	4.56	10/15/2019	11:30
9	CCV2	1	Water	NA	NA	20.1	12.03	10/15/2019	12:32

LB 105703

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5372 WorkList ID : 132142 Department : Wet-Chemistry Date : 10-15-2019 11:12:38

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2019	Water	K5372-01	pH	Cool 4 deg C	STER16	P21	MAIN-CELL	10/11/2019	SM 4500-H B
10/11/2019	Water	K5372-02	pH	Cool 4 deg C	STER16	P21	EX-CELL	10/11/2019	SM 4500-H B

Date/Time 10-15-19 11:15
Received by: JD
Relinquished by: JD

Date/Time 10-15-19 12:40
Received by: JD
Relinquished by: JD