

LOD/LOQ for all analytes

IC-2 300.0

10-15-19 Ak

Clear table


ident	concentra tion F-	concentrati on CL-	concentrati on NO2	concentrati on BR-	concentrati on NO3	concentrati on HPO4	concentrati on SO4	file name	date time	Initial wt/ Final Vol	Analyst
STD1	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-10-01_	10/1/19 11:47		CT/AP
STD2	0.2370	0.3710	0.3770	1.2480	0.3160	0.3260	1.9060	2019-10-01_	10/1/19 12:18		CT/AP
STD3	0.4310	0.6440	0.6470	2.1690	0.5450	1.0010	3.2530	2019-10-01_	10/1/19 12:49		CT/AP
STD4	0.8150	1.2060	1.2010	4.0210	1.0100	2.1860	6.0250	2019-10-01_	10/1/19 13:20		CT/AP
STD5	1.9150	2.8590	2.8480	9.4780	2.3580	5.0660	14.1850	2019-10-01_	10/1/19 13:51		CT/AP
STD6	3.9010	5.8910	5.9030	19.6450	4.9000	9.8870	29.4500	2019-10-01_	10/1/19 14:22		CT/AP
STD7	5.1010	7.6290	7.6240	25.4390	6.3710	12.5350	38.1810	2019-10-01_	10/1/19 14:53		CT/AP
ICV	2.0230	2.9740	2.9490	9.8290	2.4750	5.3960	15.0740	2019-10-01_	10/1/19 15:27		CT/AP
ICB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-10-01_	10/1/19 16:02		CT/AP
CCV	2.0280	2.9690	2.9750	9.8980	2.4680	5.3030	14.7830	2019-10-01_	10/1/19 16:33		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-10-01_	10/1/19 17:04		CT/AP
LB105705BLW	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-10-01_	10/1/19 17:35		CT/AP
LB105705BSW	1.9950	2.8670	2.8670	9.5340	2.3820	5.2300	14.2610	2019-10-01_	10/1/19 18:06		CT/AP
K5111-07	0.1550	0.2520	0.2530	0.8050	0.2150	0.1310	1.3810	2019-10-01_	10/1/19 19:08		CT/AP
K5111-07RE	0.2330	0.3810	0.3810	1.2310	0.3240	0.4020	1.9500	2019-10-01_	10/1/19 19:39		CT/AP
K5111-08	0.2380	0.3820	0.3820	1.2360	0.3230	0.4180	1.9450	2019-10-01_	10/1/19 20:09		CT/AP
K5111-08RE	0.4320	0.6580	0.6530	2.1610	0.5540	1.0660	3.3030	2019-10-01_	10/1/19 20:40		CT/AP
CCV	2.0620	2.9750	2.9820	9.9080	2.4730	5.3630	14.7360	2019-10-01_	10/1/19 21:12		CT/AP
CCB	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	0.0000	2019-10-01_	10/1/19 21:43		CT/AP

CALIBRATION OF COMPONENT F-

IC-2

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.0168603 \cdot A + 1.59581$
 RSD: 4.181 %
 Correlation coefficient: 0.999261

10-




K3 = 0 K2 = 0 K1 = 0.0168603 K0 = 1.59581
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1		0	0	0		0
2	5.217	186.5	0.2		20	
3	11.15	416.3	0.4		20	
4	25.52	872.7	0.8		20	
5	62.75	2177	2		20	
6	130.1	4532	4		20	
7	180	5956	5		20	

000

CALIBRATION OF COMPONENT CL-

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.0243407 \cdot A + 2.12645$
 RSD: 3.796 %
 Correlation coefficient: 0.999391




K3 = 0 K2 = 0 K1 = 0.0243407 K0 = 2.12645
 Base: Area
 Ref.channel: chl
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used File
1		0	0	0		0
2	13.31	217.4	0.3		20	
3	27.32	442	0.6		20	
4	55.74	903.8	1.2		20	
5	140	2261	3		20	
6	299.3	4753	6		20	
7	392.2	6181	7.5		20	

CALIBRATION OF COMPONENT NO2

Method: AnionsIC2-100119.mtw
Equation: $Q = 0.0105965 \cdot A + 2.67926$
RSD: 3.819 %
Correlation coefficient: 0.999383




K3 = 0 K2 = 0 K1 = 0.0105965 K0 = 2.67926
Base: Area
Ref.channel: chl
ISTD:
Formula: Linear
Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	22.83	458.8	0.3	20			
3	48.33	967.9	0.6	20			
4	99.37	2014	1.2	20			
5	249.6	5123	3	20			
6	514.2	1.089e+04	6	20			
7	650.9	1.414e+04	7.5	20			

5?&

CALIBRATION OF COMPONENT BR-

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.0542276 \cdot A + 7.84878$
 RSD: 3.997 %
 Correlation coefficient: 0.999325




K3 = 0 K2 = 0 K1 = 0.0542276 K0 = 7.84878
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0		0		0
2	14.04		315.7		1		20
3	29.21		655.4		2		20
4	60.29		1338		4		20
5	153.6		3351		10		20
6	334.7		7101		20		20
7	439.4		9238		25		20

CALIBRATION OF COMPONENT NO3

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.00958397 \cdot A + 2.05722$
 RSD: 4.381 %
 Correlation coefficient: 0.999189




K3 = 0 K2 = 0 K1 = 0.00958397 K0 = 2.05722
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1		0	0		0		0
2	16.28		445.2	0.25			20
3	34.07		922.7	0.5			20
4	70.26		1893	1			20
5	177.5		4706	2.5			20
6	387.2	1.001e+04		5			20
7	511.7	1.308e+04		6.25			20

CALIBRATION OF COMPONENT HPO4

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.0208939 \cdot A - 4.09382$
 RSD: 2.789 %
 Correlation coefficient: 0.999671




K3 = 0 K2 = 0 K1 = 0.0208939 K0 = -4.09382
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	10.86	508.3	0.5	20			
3	25	1154	1	20			
4	51.74	2288	2	20			
5	119.7	5045	5	20			
6	238.5	9660	10	20			
7	306.5	1.219e+04	12.5	20			

CALIBRATION OF COMPONENT SO4

Method: AnionsIC2-100119.mtw
 Equation: $Q = 0.0322739 \cdot A + 9.52592$
 RSD: 4.157 %
 Correlation coefficient: 0.999269



K3 = 0 K2 = 0 K1 = 0.0322739 K0 = 9.52592
 Base: Area
 Ref.channel: ch1
 ISTD:
 Formula: Linear
 Weight: 1

Level	Height	Area	Conc.	Vol/Dil	Retention	Used	File
1	0	0	0	0	0		
2	22.09	886	1.5	20			
3	43.59	1721	3	20			
4	87.87	3439	6	20			
5	220	8495	15	20			
6	469	1.795e+04	30	20			
7	610.5	2.337e+04	37.5	20			

Report date: 10/7/2019 3:43:47 PM
Printed by: wet


Ident: STD1
Analysis from: 10/1/2019 11:47:17 AM
File: _2019-10-01_

Last save: 10/1/2019 3:23:

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22632

Last save: 10/1/20

SAMPLE:
: CT/AP
Vial number: 2
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/7/2019 3:43:57 PM
Printed by: wet


Ident: STD2
Analysis from: 10/1/2019 12:18:16 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23:

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22633

Last save: 10/1/20

SAMPLE:
: CT/AP
Vial number: 3
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.62	0.468	5.22	4.99	186.541
2	8.12	0.255	13.31	12.72	217.377
3	9.62	0.315	22.82	21.81	458.778
4	11.97	0.348	14.04	13.42	315.664
5	13.59	0.424	16.28	15.56	445.202
6	21.12	0.561	10.85	10.37	508.257
7	24.60	0.507	22.09	21.11	885.981
7	28.00	0.411	104.62	99.99	3017.800

Report date: 10/7/2019 3:44:07 PM
Printed by: wet


Ident: STD3
Analysis from: 10/1/2019 12:49:14 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23:

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22634

Last save: 10/1/20

SAMPLE:
: CT/AP
Vial number: 4
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.63	0.479	11.14	5.10	416.290
2	8.12	0.251	27.32	12.50	441.974
3	9.61	0.312	48.33	22.10	967.889
4	11.96	0.340	29.21	13.36	655.386
5	13.57	0.413	34.07	15.58	922.664
6	21.11	0.549	24.99	11.43	1153.705
7	24.59	0.498	43.59	19.93	1720.614
7	28.00	0.406	218.65	100.00	6278.521

This report has been created by IC Net
METROHM LTD

Report date: 10/7/2019 3:44:18 PM
Printed by: wet


Ident: STD4
Analysis from: 10/1/2019 1:20:14 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22635

Last save: 10/1/2019 3:23

SAMPLE:
: CT/AP
Vial number: 5
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.61	0.437	25.52	5.66	872.665
2	8.12	0.249	55.73	12.36	903.761
3	9.61	0.315	99.37	22.04	2014.426
4	11.94	0.332	60.29	13.37	1338.145
5	13.54	0.405	70.26	15.59	1892.987
6	21.10	0.525	51.74	11.48	2288.201
7	24.59	0.494	87.87	19.49	3438.579
7	28.00	0.394	450.78	100.00	12748.764

Report date: 10/7/2019 3:44:27 PM
Printed by: wet


Ident: STD5
Analysis from: 10/1/2019 1:51:13 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22636

Last save: 10/1/2019

SAMPLE:
: CT/AP
Vial number: 6
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.63	0.459	62.75	5.59	2176.991
2	8.12	0.244	140.02	12.47	2261.454
3	9.60	0.317	249.62	22.22	5122.872
4	11.94	0.319	153.55	13.67	3350.828
5	13.52	0.395	177.46	15.80	4706.320
6	21.09	0.503	119.66	10.65	5045.007
7	24.58	0.488	220.04	19.59	8495.077
7	28.00	0.389	1123.11	100.00	31158.548

This report has been created by IC Net
METROHM LTD

Report date: 10/7/2019 3:44:36 PM
Printed by: wet


Ident: STD6
Analysis from: 10/1/2019 2:22:11 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23:

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22637

Last save: 10/1/20

SAMPLE:
: CT/AP
Vial number: 7
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.65	0.469	130.08	5.48	4532.376
2	8.11	0.238	299.25	12.61	4753.405
3	9.58	0.329	514.18	21.67	10887.895
4	11.93	0.304	334.64	14.10	7100.605
5	13.49	0.380	387.13	16.32	10010.635
6	21.04	0.490	238.54	10.05	9659.749
7	24.54	0.488	468.99	19.76	17954.862
7	28.00	0.385	2372.81	100.00	64899.526

This report has been created by IC Net
METROHM LTD

Report date: 10/7/2019 3:44:44 PM
Printed by: wet


Ident: STD7
Analysis from: 10/1/2019 2:53:10 PM
File: _2019-10-01_

Last save: 10/1/2019 3:23:

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22638

Last save: 10/1/20

SAMPLE:
: CT/AP
Vial number: 8
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.63	0.442	180.02	5.82	5956.337
2	8.11	0.234	392.21	12.69	6180.922
3	9.57	0.338	650.91	21.06	14137.076
4	11.93	0.298	439.42	14.21	9237.529
5	13.48	0.372	511.70	16.55	13079.948
6	21.02	0.485	306.50	9.91	12194.476
7	24.52	0.489	610.46	19.75	23365.712
7	28.00	0.380	3091.23	100.00	84152.000

This report has been created by IC Net
METROHM LTD

Report date: 10/7/2019 3:44:56 PM
Printed by: wet


Ident: ICV
Analysis from: 10/1/2019 3:27:08 PM
File: _2019-10-01_

Last save: 10/1/2019 3:55

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22639

Last save: 10/1/2019

SAMPLE:
: CT/AP
Vial number: 9
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec
1	5.63	0.471	65.11	5.54	2304.605
2	8.11	0.240	145.50	12.37	2355.990
3	9.59	0.317	255.18	21.70	5312.747
4	11.94	0.309	160.27	13.63	3480.214
5	13.51	0.383	187.36	15.93	4951.250
6	21.07	0.499	127.32	10.83	5361.438
7	24.56	0.484	235.08	19.99	9046.134
7	28.00	0.386	1175.81	100.00	32812.379

Report date: 10/7/2019 3:45:06 PM
Printed by: wet


Ident: ICB
Analysis from: 10/1/2019 4:02:11 PM
File: _2019-10-01_

Last save: 10/1/2019 4:30

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22640

Last save: 10/1/2019 4:30

SAMPLE:
: CT/AP
Vial number: 10
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

300.0

Report date: 10/15/2019 2:10:19 PM
Printed by: wet


Ident: CCV
Analysis from: 10/1/2019 4:33:09 PM
File: _2019-10-01_

Last save: 10/1/2019 5:01:08 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22641

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.61	0.455	66.84	5.57	2311.531	7.09
2	8.10	0.226	150.17	12.51	2352.272	7.21
3	9.57	0.308	261.13	21.74	5362.013	16.44
4	11.92	0.290	167.06	13.91	3505.984	10.75
5	13.49	0.356	194.55	16.20	4936.613	15.14
6	21.04	0.490	127.52	10.62	5271.831	16.17
7	24.54	0.477	233.53	19.45	8865.900	27.19
7	28.00	0.372	1200.80	100.00	32606.144	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 2:10:30 PM
Printed by: wet


Ident: CCB
Analysis from: 10/1/2019 5:04:06 PM
File: _2019-10-01_

Last save: 10/1/2019 5:32:07 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22642

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 2:10:39 PM
Printed by: wet


Ident: LB105705BLW
Analysis from: 10/1/2019 5:35:05 PM
File: _2019-10-01_

Last save: 10/15/2019 2:01:06 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22643

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 13
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 2:10:49 PM
Printed by: wet


Ident: LB105705BSW
Analysis from: 10/1/2019 6:06:04 PM
File: _2019-10-01_

Last save: 10/15/2019 2:01:06 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22644

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 14
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.61	0.463	64.61	5.58	2271.971	7.20
2	8.10	0.226	144.14	12.46	2268.741	7.19
3	9.58	0.307	250.09	21.61	5158.952	16.34
4	11.93	0.290	160.14	13.84	3371.531	10.68
5	13.51	0.355	186.63	16.13	4755.357	15.06
6	21.04	0.489	125.57	10.85	5202.548	16.48
7	24.53	0.475	225.82	19.52	8542.430	27.06
7	28.00	0.372	1157.00	100.00	31571.530	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:34:54 PM
Printed by: wet


Ident: K5111-07 0.25
Analysis from: 10/1/2019 7:08:03 PM
File: _2019-10-01_

Last save: 10/1/2019 7:36:03 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22646

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 15
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.62	0.538	2.27	3.81	88.803	5.22
2	8.10	0.219	7.81	13.10	119.574	7.03
3	9.60	0.282	11.56	19.37	224.859	13.21
4	11.95	0.290	7.36	12.33	152.156	8.94
5	13.57	0.353	9.10	15.24	234.138	13.76
6	21.08	0.546	6.81	11.41	321.255	18.88
7	24.55	0.477	14.75	24.72	560.915	32.96
7	28.00	0.387	59.66	99.99	1701.701	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:35:05 PM
Printed by: wet


Ident: K5111-07RE 0.5
Analysis from: 10/1/2019 7:39:01 PM
File: _2019-10-01_

Last save: 10/1/2019 8:07:01 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22647

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 16
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.60	0.496	4.84	4.31	181.226	5.78
2	8.10	0.225	14.31	12.75	225.503	7.19
3	9.60	0.288	23.65	21.06	466.043	14.85
4	11.95	0.295	14.74	13.12	309.308	9.86
5	13.57	0.355	17.94	15.97	461.829	14.72
6	21.07	0.524	12.98	11.56	580.951	18.51
7	24.55	0.479	23.80	21.20	912.939	29.09
7	28.00	0.380	112.27	99.99	3137.801	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:35:15 PM
Printed by: wet


Ident: K5111-08 0.5
Analysis from: 10/1/2019 8:09:59 PM
File: _2019-10-01_

Last save: 10/1/2019 8:37:59 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22648

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 17
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.62	0.510	4.89	4.36	187.080	5.92
2	8.11	0.226	14.21	12.68	226.298	7.16
3	9.60	0.287	23.70	21.16	467.635	14.81
4	11.96	0.296	14.67	13.09	310.946	9.84
5	13.58	0.356	17.76	15.86	459.741	14.56
6	21.08	0.529	13.07	11.66	596.363	18.88
7	24.56	0.480	23.72	21.18	910.391	28.82
7	28.00	0.383	112.01	99.99	3158.455	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:35:27 PM
Printed by: wet


Ident: K5111-08RE 1.0
Analysis from: 10/1/2019 8:40:59 PM
File: _2019-10-01_

Last save: 10/1/2019 9:08:59 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22649

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 18
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.62	0.520	10.53	4.62	418.024	6.52
2	8.10	0.226	28.37	12.44	453.077	7.07
3	9.59	0.289	49.28	21.60	979.683	15.28
4	11.95	0.295	30.31	13.29	652.320	10.17
5	13.55	0.353	36.46	15.98	942.023	14.69
6	21.06	0.522	27.05	11.86	1215.985	18.96
7	24.54	0.476	46.09	20.21	1751.663	27.32
7	28.00	0.383	228.09	99.99	6412.776	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:35:38 PM
Printed by: wet


Ident: CCV
Analysis from: 10/1/2019 9:12:03 PM
File: _2019-10-01_

Last save: 10/1/2019 9:40:03 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22650

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 11
Volume: 20.0 µL
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No	Retention min	Width/2 min	Height mV	Height %	Area mV*sec	Area %
1	5.61	0.478	65.30	5.47	2351.377	7.19
2	8.10	0.223	148.43	12.44	2357.462	7.21
3	9.58	0.309	256.82	21.52	5374.691	16.43
4	11.93	0.285	165.90	13.90	3509.314	10.73
5	13.51	0.347	193.76	16.24	4946.162	15.12
6	21.02	0.486	129.07	10.82	5329.197	16.29
7	24.52	0.473	234.01	19.61	8836.662	27.02
7	28.00	0.372	1193.28	100.00	32704.864	100.00

This report has been created by IC Net
METROHM LTD

Report date: 10/15/2019 1:35:48 PM
Printed by: wet


Ident: CCB
Analysis from: 10/1/2019 9:43:02 PM
File: _2019-10-01_

Last save: 10/1/2019 10:11:02 PM

Method: AnionsIC2-100119.mtw
Run operator: wet
Analysis number: 22651

Last save: 10/1/2019 3:21:

SAMPLE:
: CT/AP
Vial number: 12
Volume: 20.0 μ L
Dilution: 1.00
Amount: 1.0000



Quantitation method: Custom

No peaks

This report has been created by IC Net
METROHM LTD