



Analytical Summary Report

Analysis Method: 9045D

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105719

Slope : 98.7

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	10/15/2019	17:00
2	CAL2	1	Water	NA	NA	20.2	7.01	10/15/2019	17:01
3	CAL3	1	Water	NA	NA	20.2	10.03	10/15/2019	17:02
4	ICV	1	Water	NA	NA	20.1	7.00	10/15/2019	17:05
5	CCV1	1	Water	NA	NA	20.1	2.02	10/15/2019	17:06
6	K5152-09	1	Solid	20.02	20	21.9	8.18	10/15/2019	17:08
7	K5392-04	1	Solid	20.03	20	21.3	5.42	10/15/2019	17:10
8	K5392-08	1	Solid	20.04	20	20.8	5.28	10/15/2019	17:11
9	K5392-08DUP	1	Solid	20.03	20	20.9	5.29	10/15/2019	17:12
10	CCV2	1	Water	NA	NA	20.1	12.02	10/15/2019	17:15

13105719

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5152 WorkList ID : 132154 Department : Wet-Chemistry Date : 10-15-2019 14:54:46

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/15/2019	Solid	K5152-09-17	Corrosivity	Cool 4 deg C	PSEG05		SU-01-101519	10/15/2019	9045D
10/15/2019	Solid	K5392-04-17	Corrosivity	Cool 4 deg C	PSEG01	N41	TP-22A-S	10/15/2019	9045D
10/15/2019	Solid	K5392-08-17	Corrosivity	Cool 4 deg C	PSEG01	N41	TP-22A-D	10/15/2019	9045D

Date/Time 10.15.19 15:00
 Received by: J
 Relinquished by: [Signature]

Date/Time 10.15.19 16:30
 Received by: [Signature]
 Relinquished by: J