

LB105730

CHEMTECH
 284 Sheffield Street,
 Mountainside, NJ 07092
 Reviewed by : CT Instrument ID : Konelab

10/16/2019 9:05

Test CNEPA-NEW


Accepted 10/11/2019 9:59

Factor (Slope) 1088 0.0009191
 Bias (Y-Intercept) 0.008

CT
10/16/19

Coeff. of det. 0.999519

Errors



	Calibrator	Response	Calc. con.	Conc.	Errors
1	0.0PPBCN 50.0	0.006	-1.5024	0.0000	
2	5.0PPBCN 55.0	0.011	3.9577	5.0000	79%
3	10PPBCN 510.0	0.016	8.4268	10.0000	84%
4	100PPBCN 5100.0	0.100	99.7904	100.0000	99%
5	250PPBCN 5250.0	0.245	258.4875	250.0000	103%
6	500PPBCN 5500.0	0.464	495.8401	500.0000	99%

Used reagent lots:

Reagent	Lot
CN-BUFFER	default
CN-CHLOR T	default
CN-PYR BAR	default

CT
10/16/19

CHEMTECH
284 Sheffield Street,
Mountainside, NJ 07092

Reviewed by : CT

Instrument ID : Konelab

10/11/2019 11:14

Test: CNEPA-NEW

Sample Id	Result	Dil. 1 +	Response	Errors
ICV001 ICV001	100.225	0.0	0.100	
ICB001 ICB001	-0.402	0.0	0.007	
CCV001 CCV001	247.904	0.0	0.236	
CCB001 CCB001	0.091	0.0	0.008	
PB122371BL PBW37	-0.192	0.0	0.008	
K4374-10 MX2H70	48.268	0.0	0.052	
K4374-11 MX2H70D	48.388	0.0	0.052	
K4374-12 MX2H70S	136.828	0.0	0.134	
CCV002 CCV002	249.531	0.0	0.237	
CCB002 CCB002	0.107	0.0	0.008	

N	10
Mean	83.075
SD	98.9648
CV%	119.13

10/16/19
CT

Aquakem v. 7.2AQ1

Results from time period:

Fri Oct 11 08:24:46 2019

Fri Oct 11 11:12:48 2019

Sample Id	Sam/Ctr/c	Test short nam	Test type	Result	Result unit	Result date and time
S0.0	A	CNEPA-NEW	P	-1.5024	µg/l	10/11/2019 9:57:59
S5.0	A	CNEPA-NEW	P	3.9577	µg/l	10/11/2019 9:58:00
S10.0	A	CNEPA-NEW	P	8.4268	µg/l	10/11/2019 9:58:01
S100.0	A	CNEPA-NEW	P	99.7904	µg/l	10/11/2019 9:58:02
S250.0	A	CNEPA-NEW	P	258.4875	µg/l	10/11/2019 9:58:03
S500.0	A	CNEPA-NEW	P	495.8401	µg/l	10/11/2019 9:58:04
ICV001 ICV001	S	CNEPA-NEW	P	100.2248	µg/l	10/11/2019 11:12:39
ICB001 ICB001	S	CNEPA-NEW	P	-0.4023	µg/l	10/11/2019 11:12:40
CCV001 CCV001	S	CNEPA-NEW	P	247.9041	µg/l	10/11/2019 11:12:41
CCB001 CCB001	S	CNEPA-NEW	P	0.0908	µg/l	10/11/2019 11:12:42
PB122371BL PBW37	S	CNEPA-NEW	P	-0.1917	µg/l	10/11/2019 11:12:43
K4374-10 MX2H70	S	CNEPA-NEW	P	48.2679	µg/l	10/11/2019 11:12:44
K4374-11 MX2H70D	S	CNEPA-NEW	P	48.3885	µg/l	10/11/2019 11:12:45
K4374-12 MX2H70S	S	CNEPA-NEW	P	136.8282	µg/l	10/11/2019 11:12:46
CCV002 CCV002	S	CNEPA-NEW	P	249.531	µg/l	10/11/2019 11:12:47
CCB002 CCB002	S	CNEPA-NEW	P	0.1072	µg/l	10/11/2019 11:12:48