


LB105T36

ISM02.4

INSTRUMENT ID : CUI



Linear

A= 0.0000e+000

B= 2.0463e-004

C= -1.1760e-002

Rho= 0.9997906

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.008	0.008	98	0.000	98					—
0.05	0.050										—
0.20	0.200	0.198	-0.002	1026	0.0 %	1026					-1
2.50	2.500	2.428	-0.072	11925	0.0 %	11925					-3
5.00	5.000	5.035	0.035	24663	0.0 %	24663					1
7.50	7.500	7.632	0.132	37353	0.0 %	37353					2
10.0	10.000	9.898	-0.102	48427	0.0 %	48427					-1

RS 10.16.19

LB105736
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	98	-	0	ISM02.4	PPB	10/16/2019 10:48	Std
0.2 S0.2	1026	-	0.2	ISM02.4	PPB	10/16/2019 10:51	Std
2.5 S2.5	11925	-	2.5	ISM02.4	PPB	10/16/2019 10:53	Std
5 S5.0	24663	-	5	ISM02.4	PPB	10/16/2019 10:55	Std
7.5 S7.5	37353	-	7.5	ISM02.4	PPB	10/16/2019 11:01	Std
10 S10.0	48427	-	10	ISM02.4	PPB	10/16/2019 11:03	Std
ICV	ICV	20519	4.1871 -	ISM02.4	PPB	10/16/2019 11:06	SMPL
ICB	ICB	32	-0.0052 -	ISM02.4	PPB	10/16/2019 11:08	SMPL
CCV046	CCV046	24506	5.003 -	ISM02.4	PPB	10/16/2019 11:10	SMPL
CCB046	CCB046	47	-0.0021 -	ISM02.4	PPB	10/16/2019 11:12	SMPL
PB124025E	PBS025	71	0.0028 -	ISM02.4	PPB	10/16/2019 11:15	SMPL
K5163-01	MH0GK5	3366	0.677 -	ISM02.4	PPB	10/16/2019 11:17	SMPL
K5163-02	MH0GK6	1252	0.2444 -	ISM02.4	PPB	10/16/2019 11:19	SMPL
K5163-03	MH0GK7	413	0.0728 -	ISM02.4	PPB	10/16/2019 11:21	SMPL
K5163-04	MH0GK8	3032	0.6087 -	ISM02.4	PPB	10/16/2019 11:24	SMPL
K5163-05	MH0GK8D	3044	0.6111 -	ISM02.4	PPB	10/16/2019 11:26	SMPL
K5163-06	MH0GK8S	14696	2.9955 -	ISM02.4	PPB	10/16/2019 11:28	SMPL
PB124026E	PBS026	38	-0.004 -	ISM02.4	PPB	10/16/2019 11:31	SMPL
K5182-01	MBEPG7	1803	0.3572 -	ISM02.4	PPB	10/16/2019 11:33	SMPL
K5182-02	MBEPJ9	1629	0.3216 -	ISM02.4	PPB	10/16/2019 11:35	SMPL
K5182-03	MBEPK0	1572	0.3099 -	ISM02.4	PPB	10/16/2019 11:37	SMPL
K5182-04	MBEP93	867	0.1657 -	ISM02.4	PPB	10/16/2019 11:40	SMPL
K5182-05	MBEP94	15860	3.2337 -	ISM02.4	PPB	10/16/2019 11:42	SMPL
K5182-06	MBEP95	17957	3.6628 -	ISM02.4	PPB	10/16/2019 11:44	SMPL
K5182-07	MBEP96	15847	3.2311 -	ISM02.4	PPB	10/16/2019 11:47	SMPL
K5182-08	MBEP97	22145	4.5198 -	ISM02.4	PPB	10/16/2019 11:49	SMPL
K5182-09	MBEP98	8516	1.7309 -	ISM02.4	PPB	10/16/2019 11:51	SMPL
K5182-10	MBEP99	9901	2.0143 -	ISM02.4	PPB	10/16/2019 11:53	SMPL
K5182-11	MBEPA0	51	-0.0013 -	ISM02.4	PPB	10/16/2019 11:56	SMPL
K5182-12	MBEPA1	131	0.015 -	ISM02.4	PPB	10/16/2019 11:58	SMPL
CCV047	CCV047	24634	5.0292 -	ISM02.4	PPB	10/16/2019 12:00	SMPL
CCB047	CCB047	62	0.0009 -	ISM02.4	PPB	10/16/2019 12:03	SMPL
K5182-13	MBEPA2	3188	0.6406 -	ISM02.4	PPB	10/16/2019 12:05	SMPL
K5182-14	MBEPA3	3215	0.6461 -	ISM02.4	PPB	10/16/2019 12:07	SMPL
K5182-15	MBEPK1	66	0.0017 -	ISM02.4	PPB	10/16/2019 12:09	SMPL
K5182-16	MBEPK2	31	-0.0054 -	ISM02.4	PPB	10/16/2019 12:12	SMPL
K5182-17	MBEPE7	2217	0.4419 -	ISM02.4	PPB	10/16/2019 12:14	SMPL
K5182-18	MBEPE8	1769	0.3502 -	ISM02.4	PPB	10/16/2019 12:16	SMPL
K5182-19	MBEPD3	2070	0.4118 -	ISM02.4	PPB	10/16/2019 12:18	SMPL
K5182-20	MBEPD4	1731	0.3425 -	ISM02.4	PPB	10/16/2019 12:21	SMPL
K5182-21	MBEPD4D	1866	0.3701 -	ISM02.4	PPB	10/16/2019 12:23	SMPL
K5182-22	MBEPD4S	16391	3.3424 -	ISM02.4	PPB	10/16/2019 12:25	SMPL
PB124027E	PBS027	23	-0.0071 -	ISM02.4	PPB	10/16/2019 12:28	SMPL
K5183-01	MBEPE9	2734	0.5477 -	ISM02.4	PPB	10/16/2019 12:30	SMPL
K5183-02	MBEPF0	4140	0.8354 -	ISM02.4	PPB	10/16/2019 12:32	SMPL
K5183-03	MBEPF1	4090	0.8252 -	ISM02.4	PPB	10/16/2019 12:34	SMPL

K5183-04	MBEPF2	9284	1.8881 -	ISM02.4	PPB	10/16/2019 12:37	SMPL
K5183-05	MBEPF3	134467	27.5046 -	ISM02.4	PPB	10/16/2019 12:39	SMPL
K5183-06	MBEPF4	92861	18.9907 -	ISM02.4	PPB	10/16/2019 12:41	SMPL
K5183-07	MBEPD5	4435	0.8958 -	ISM02.4	PPB	10/16/2019 12:43	SMPL
K5183-08	MBEPC6	11092	2.258 -	ISM02.4	PPB	10/16/2019 12:46	SMPL
K5183-09	MBEPC7	11828	2.4086 -	ISM02.4	PPB	10/16/2019 12:48	SMPL
CCV048	CCV048	24513	5.0044 -	ISM02.4	PPB	10/16/2019 12:50	SMPL
CCB048	CCB048	45	-0.0026 -	ISM02.4	PPB	10/16/2019 12:53	SMPL
K5183-10	MBEPC8	1161	0.2258 -	ISM02.4	PPB	10/16/2019 12:55	SMPL
K5183-11	MBEPC8D	1167	0.227 -	ISM02.4	PPB	10/16/2019 12:57	SMPL
K5183-13	MBEPC9	1109	0.2152 -	ISM02.4	PPB	10/16/2019 13:02	SMPL
K5183-14	MBEPD0	980	0.1888 -	ISM02.4	PPB	10/16/2019 13:04	SMPL
K5183-15	MBEPD1	1173	0.2283 -	ISM02.4	PPB	10/16/2019 13:06	SMPL
K5183-16	MBEPD2	13303	2.7105 -	ISM02.4	PPB	10/16/2019 13:09	SMPL
K5183-17	MBEPH9	136634	27.9481 -	ISM02.4	PPB	10/16/2019 13:11	SMPL
K5183-18	MBEPJ0	4368	0.8821 -	ISM02.4	PPB	10/16/2019 13:13	SMPL
K5183-19	MBEPJ1	475	0.0854 -	ISM02.4	PPB	10/16/2019 13:16	SMPL
K5183-20	MBEPJ2	483	0.0871 -	ISM02.4	PPB	10/16/2019 13:18	SMPL
K5183-21	MBEPJ3	17917	3.6547 -	ISM02.4	PPB	10/16/2019 13:20	SMPL
K5183-22	MBEPJ4	15965	3.2552 -	ISM02.4	PPB	10/16/2019 13:22	SMPL
PB124028E	PBS028	61	0.0007 -	ISM02.4	PPB	10/16/2019 13:25	SMPL
K5200-01	MBEPJ5	16374	3.3389 -	ISM02.4	PPB	10/16/2019 13:27	SMPL
K5200-02	MBEPJ6	15181	3.0948 -	ISM02.4	PPB	10/16/2019 13:29	SMPL
K5200-03	MBEPJ7	1190	0.2318 -	ISM02.4	PPB	10/16/2019 13:32	SMPL
K5200-04	MBEPJ8	1587	0.313 -	ISM02.4	PPB	10/16/2019 13:34	SMPL
K5200-05	MBEP72	22301	4.5518 -	ISM02.4	PPB	10/16/2019 13:36	SMPL
K5200-06	MBEP73	27142	5.5424 -	ISM02.4	PPB	10/16/2019 13:38	SMPL
CCV049	CCV049	25283	5.162 -	ISM02.4	PPB	10/16/2019 13:41	SMPL
CCB049	CCB049	73	0.0032 -	ISM02.4	PPB	10/16/2019 13:43	SMPL
K5200-07	MBEP74	3683	0.7419 -	ISM02.4	PPB	10/16/2019 13:45	SMPL
K5200-08	MBEP75	3687	0.7427 -	ISM02.4	PPB	10/16/2019 13:47	SMPL
K5200-09	MBEP76	21325	4.352 -	ISM02.4	PPB	10/16/2019 13:50	SMPL
K5200-10	MBEP77	2243	0.4472 -	ISM02.4	PPB	10/16/2019 13:52	SMPL
K5200-11	MBEP78	3318	0.6672 -	ISM02.4	PPB	10/16/2019 13:54	SMPL
K5200-12	MBEP79	25920	5.2923 -	ISM02.4	PPB	10/16/2019 13:57	SMPL
K5200-13	MBEP80	3022	0.6066 -	ISM02.4	PPB	10/16/2019 13:59	SMPL
K5200-14	MBEP81	2443	0.4882 -	ISM02.4	PPB	10/16/2019 14:01	SMPL
K5200-15	MBEP82	2013	0.4002 -	ISM02.4	PPB	10/16/2019 14:04	SMPL
K5200-16	MBEPG8	31919	6.5199 -	ISM02.4	PPB	10/16/2019 14:06	SMPL
K5200-17	MBEPG8D	32073	6.5514 -	ISM02.4	PPB	10/16/2019 14:08	SMPL
K5200-18	MBEPG8S	44434	9.0809 -	ISM02.4	PPB	10/16/2019 14:14	SMPL
K5200-19	MBEPG9	51716	10.571 -	ISM02.4	PPB	10/16/2019 14:16	SMPL
K5200-20	MBEPH0	19647	4.0087 -	ISM02.4	PPB	10/16/2019 14:18	SMPL
K5200-21	MBEPH1	24559	5.0138 -	ISM02.4	PPB	10/16/2019 14:20	SMPL
K5200-22	MBEPH2	7641	1.5518 -	ISM02.4	PPB	10/16/2019 14:23	SMPL
PB124029E	PBS029	65	0.0015 -	ISM02.4	PPB	10/16/2019 14:25	SMPL
K5224-01	MBEPH3	11249	2.2902 -	ISM02.4	PPB	10/16/2019 14:27	SMPL

INSTRUMENT ID : CV1

K5224-02	MBEPD6	31688	6.4727 -	ISM02.4	PPB	10/16/2019 14:30 SMPL
K5224-03	MBEPD7	22453	4.5829 -	ISM02.4	PPB	10/16/2019 14:32 SMPL
CCV050	CCV050	25506	5.2076 -	ISM02.4	PPB	10/16/2019 14:34 SMPL
CCB050	CCB050	59	0.0003 -	ISM02.4	PPB	10/16/2019 14:36 SMPL
K5224-04	MBEPD8	37991	7.7625 -	ISM02.4	PPB	10/16/2019 14:39 SMPL
K5224-05	MBEPD9	20912	4.2675 -	ISM02.4	PPB	10/16/2019 14:41 SMPL
K5224-06	MBEPE0	33969	6.9394 -	ISM02.4	PPB	10/16/2019 14:43 SMPL
K5224-07	MBEPE3	18447	3.7631 -	ISM02.4	PPB	10/16/2019 14:46 SMPL
K5224-08	MBEPE4	6107	1.2379 -	ISM02.4	PPB	10/16/2019 14:48 SMPL
K5224-09	MBEPE1	16532	3.3712 -	ISM02.4	PPB	10/16/2019 14:50 SMPL
K5224-10	MBEPE2	37611	7.6847 -	ISM02.4	PPB	10/16/2019 14:52 SMPL
K5224-11	MBEPE5	6319	1.2813 -	ISM02.4	PPB	10/16/2019 14:55 SMPL
K5224-12	MBEPH4	971	0.1869 -	ISM02.4	PPB	10/16/2019 14:57 SMPL
K5224-13	MBEPH5	899	0.1722 -	ISM02.4	PPB	10/16/2019 14:59 SMPL
K5224-14	MBEPH6	1061	0.2054 -	ISM02.4	PPB	10/16/2019 15:02 SMPL
K5224-15	MBEPH7	1041	0.2013 -	ISM02.4	PPB	10/16/2019 15:04 SMPL
K5224-16	MBEPA4	14618	2.9796 -	ISM02.4	PPB	10/16/2019 15:06 SMPL
K5224-17	MBEPA5	13955	2.8439 -	ISM02.4	PPB	10/16/2019 15:08 SMPL
K5224-18	MBEPA6	21830	4.4554 -	ISM02.4	PPB	10/16/2019 15:11 SMPL
K5224-19	MBEPA7	31055	6.3431 -	ISM02.4	PPB	10/16/2019 15:13 SMPL
K5224-20	MBEPA8	25016	5.1073 -	ISM02.4	PPB	10/16/2019 15:15 SMPL
K5224-21	MBEPA8D	25217	5.1485 -	ISM02.4	PPB	10/16/2019 15:18 SMPL
K5224-22	MBEPA8S	35860	7.3264 -	ISM02.4	PPB	10/16/2019 15:20 SMPL
K5183-05DLX5		25579	5.2226 -	ISM02.4	PPB	10/16/2019 15:22 SMPL
CCV051	CCV051	25199	5.1448 -	ISM02.4	PPB	10/16/2019 15:25 SMPL
CCB051	CCB051	31	-0.0054 -	ISM02.4	PPB	10/16/2019 15:27 SMPL
K5183-06DLX5		17425	3.554 -	ISM02.4	PPB	10/16/2019 15:29 SMPL
K5183-12	MBEPC8S	13098	2.6685 -	ISM02.4	PPB	10/16/2019 15:31 SMPL
K5183-17DLX5		25864	5.2809 -	ISM02.4	PPB	10/16/2019 15:34 SMPL
K5200-19DLX2		24644	5.0312 -	ISM02.4	PPB	10/16/2019 15:36 SMPL
CCV052	CCV052	25212	5.1475 -	ISM02.4	PPB	10/16/2019 15:38 SMPL
CCB052	CCB052	74	0.0034 -	ISM02.4	PPB	10/16/2019 15:41 SMPL