



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105742

Slope : 98.7

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/16/2019	11:10
2	CAL2	1	Water	NA	NA	20.1	7.00	10/16/2019	11:11
3	CAL3	1	Water	NA	NA	20.2	10.03	10/16/2019	11:12
4	ICV	1	Water	NA	NA	20.1	6.99	10/16/2019	11:15
5	CCV1	1	Water	NA	NA	20.1	2.02	10/16/2019	11:16
6	K5397-01	1	Water	NA	NA	19.8	5.34	10/16/2019	11:20
7	K5397-01DUP	1	Water	NA	NA	19.9	5.35	10/16/2019	11:21
8	CCV2	1	Water	NA	NA	20.1	12.01	10/16/2019	11:25

LB 105742

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5397 WorkList ID : 132200 Department : Wet-Chemistry Date : 10-16-2019 10:48:23

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/11/2019	Water	K5397-01	PH	Cool 4 deg C	STER08	P31	POOL-WATER	10/11/2019	SM 4500-H B

Date/Time 10-16-19 11:00
Received by: J
Relinquished by: CS

Date/Time 10-16-19 12:00
Received by: CS
Relinquished by: J