



## Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105749

Slope : 98.3

pH Meter ID : WC PH METER-1

| Calibration Standards          | Chemtech Log# |
|--------------------------------|---------------|
| PH 4 BUFFER SOLUTION           | W2568         |
| buffer solution pH 7 yellow    | W2447         |
| PH 10.01 BUFFER,COLOR CD 475ML | W2570         |
| buffer solution pH 7 yellow    | W2446         |
| Buffer Solution, PH2 (500ml)   | W2592         |
| Buffer Solution, PH12 (500ml)  | W2467         |

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

| Seq | LabID       | DF | Matrix | Weight (gm) | Volume (ml) | Temperature (°C) | Result (pH) | Anal Date  | Anal Time |
|-----|-------------|----|--------|-------------|-------------|------------------|-------------|------------|-----------|
| 1   | CAL1        | 1  | Water  | NA          | NA          | 20.2             | 4.01        | 10/16/2019 | 17:10     |
| 2   | CAL2        | 1  | Water  | NA          | NA          | 20.2             | 7.00        | 10/16/2019 | 17:11     |
| 3   | CAL3        | 1  | Water  | NA          | NA          | 20.2             | 10.03       | 10/16/2019 | 17:12     |
| 4   | ICV         | 1  | Water  | NA          | NA          | 20.1             | 6.99        | 10/16/2019 | 17:15     |
| 5   | CCV1        | 1  | Water  | NA          | NA          | 20.1             | 2.02        | 10/16/2019 | 17:19     |
| 6   | K5407-01    | 1  | Water  | NA          | NA          | 17.3             | 9.54        | 10/16/2019 | 17:20     |
| 7   | K5407-01DUP | 1  | Water  | NA          | NA          | 17.4             | 9.55        | 10/16/2019 | 17:21     |
| 8   | CCV2        | 1  | Water  | NA          | NA          | 20.1             | 12.02       | 10/16/2019 | 17:22     |

23105749

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5407      WorkList ID : 132216      Department : Wet-Chemistry      Date : 10-16-2019 16:59:16

| Due Date   | Matrix | Sample      | Test | Preservative | Customer | Storage Location | Customer Sample   | Collect Date | Method |
|------------|--------|-------------|------|--------------|----------|------------------|-------------------|--------------|--------|
| 10/14/2019 | Water  | K5407-01 -F | pH   | Cool 4 deg C | ENVI27   | P31              | DEVELOPMENT-WATER | 10/14/2019   | 9040C  |

Date/Time 10-16-19 17:10:00  
Received by: Jn  
Relinquished by: S

Date/Time 10-16-19 17:40  
Received by: CL  
Relinquished by: Jn