


LB105755

7470A

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.3943e-004

C= -3.7622e-002

Rho= 0.9996414

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.000	0.000	158	0.000	158				
0.2	0.200	0.167	-0.033	856	0.0 %	856				
2.5	2.500	2.500	0.000	10599	0.0 %	10599				
5.0	5.000	5.161	0.161	21714	0.0 %	21714				
7.5	7.500	7.328	-0.172	30763	0.0 %	30763				
10.0	10.000	10.043	0.043	42104	0.0 %	42104				

LB105755 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 17 Oct 2019 08:08:10

Sample ID	Extended ID	µ Abs	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S50-0	158	-	0.0000	7470A	PPB	17 Oct 2019 08:38:34	S	Std
0.2 - 1	S50-2	856	-	0.2000	7470A	PPB	17 Oct 2019 08:40:51	S	Std
2.5 - 1	S52-5	10599	-	2.5000	7470A	PPB	17 Oct 2019 08:43:00	S	Std
5.0 - 1	S55-0	21714	-	5.0000	7470A	PPB	17 Oct 2019 08:45:08	S	Std
7.5 - 1	S57-5	30763	-	7.5000	7470A	PPB	17 Oct 2019 08:47:17	S	Std
10.0 - 1	S10-0	42104	-	10.0000	7470A	PPB	17 Oct 2019 08:49:24	S	Std
ICV59 - 1	ICV59	17636	4.1849	-	7470A	PPB	17 Oct 2019 08:54:29	U	SMPL
ICB59 - 1	ICB59	0	-0.0376	-	7470A	PPB	17 Oct 2019 08:56:35	U	SMPL
CCV53 - 1	CCV53	21918	5.2101	-	7470A	PPB	17 Oct 2019 08:58:42	U	SMPL
CCB53 - 1	CCB53	6	-0.0362	-	7470A	PPB	17 Oct 2019 09:00:48	U	SMPL
CRA - 1	CRA	823	0.1594	-	7470A	PPB	17 Oct 2019 09:02:54	U	SMPL
HighStd - 1	HighStd	41962	10.0092	-	7470A	PPB	17 Oct 2019 09:05:01	U	SMPL
ChkStd - 1	ChkStd	29338	6.9867	-	7470A	PPB	17 Oct 2019 09:07:08	U	SMPL
PB124045BL - 1	PBW	6	-0.0362	-	7470A	PPB	17 Oct 2019 09:09:16	U	SMPL
PB124045BS - 1	LCSW	15691	3.7192	-	7470A	PPB	17 Oct 2019 09:11:24	U	SMPL
K5152-09 - 1	SU-01-101519	496	0.0811	-	7470A	PPB	17 Oct 2019 09:13:31	U	SMPL
K5152-09DUP - 1	SU-01-101519DUP	516	0.0859	-	7470A	PPB	17 Oct 2019 09:15:40	U	SMPL
K5152-09MS - 1	SU-01-101519MS	18704	4.4406	-	7470A	PPB	17 Oct 2019 09:17:48	U	SMPL
K5152-09MSD - 1	SU-01-101519MSD	19392	4.6054	-	7470A	PPB	17 Oct 2019 09:19:56	U	SMPL
K5377-01 - 1	205-166	1583	0.3414	-	7470A	PPB	17 Oct 2019 09:22:03	U	SMPL
CCV54 - 1	CCV54	21778	5.1766	-	7470A	PPB	17 Oct 2019 09:24:10	U	SMPL
CCB54 - 1	CCB54	11	-0.0350	-	7470A	PPB	17 Oct 2019 09:26:17	U	SMPL
K5392-04 - 1	TP-22A-S	1465	0.3131	-	7470A	PPB	17 Oct 2019 09:28:23	U	SMPL
K5392-08 - 1	TP-22A-D	1880	0.4125	-	7470A	PPB	17 Oct 2019 09:30:30	U	SMPL
PB124046BL - 1	PBW	48	-0.0261	-	7470A	PPB	17 Oct 2019 09:32:37	U	SMPL
PB124046BS - 1	LCSW	15578	3.6922	-	7470A	PPB	17 Oct 2019 09:34:44	U	SMPL
K5215-15 - 1	30358-360-363-364	38	-0.0285	-	7470A	PPB	17 Oct 2019 09:36:51	U	SMPL
K5355-07 - 1	COMP-31	54	-0.0247	-	7470A	PPB	17 Oct 2019 09:38:58	U	SMPL
K5355-08 - 1	COMP-32	37	-0.0288	-	7470A	PPB	17 Oct 2019 09:41:06	U	SMPL
K5355-09 - 1	COMP-33	46	-0.0266	-	7470A	PPB	17 Oct 2019 09:43:15	U	SMPL
K5356-02 - 1	72-12017	38	-0.0285	-	7470A	PPB	17 Oct 2019 09:45:23	U	SMPL
K5357-02 - 1	72-11933	33	-0.0297	-	7470A	PPB	17 Oct 2019 09:47:30	U	SMPL
CCV55 - 1	CCV55	21922	5.2111	-	7470A	PPB	17 Oct 2019 09:49:37	U	SMPL
CCB55 - 1	CCB55	34	-0.0295	-	7470A	PPB	17 Oct 2019 09:51:44	U	SMPL
K5357-02DUP - 1	72-11933DUP	46	-0.0266	-	7470A	PPB	17 Oct 2019 09:53:50	U	SMPL
K5357-02MS - 1	72-11933MS	17690	4.1978	-	7470A	PPB	17 Oct 2019 09:55:57	U	SMPL
K5357-02MSD - 1	72-11933MSD	14778	3.5006	-	7470A	PPB	17 Oct 2019 09:58:04	U	SMPL
PB124047BL - 1	PBW	-2	-0.0381	-	7470A	PPB	17 Oct 2019 10:00:11	U	SMPL
PB124047BS - 1	LCSW	15667	3.7135	-	7470A	PPB	17 Oct 2019 10:02:20	U	SMPL
K5358-04 - 1	20190897-AMENDED-COMP-KM-CA	35753	8.5226	-	7470A	PPB	17 Oct 2019 10:04:27	U	SMPL
K5358-04DUP - 1	20190897-AMENDED-COMP-KM-CADUP	35884	8.5540	-	7470A	PPB	17 Oct 2019 10:06:35	U	SMPL
K5358-04MS - 1	20190897-AMENDED-COMP-KM-CAMS	51495	12.2917	-	7470A	PPB	17 Oct 2019 10:08:43	U	SMPL
K5358-04MSD - 1	20190897-AMENDED-COMP-KM-CAMSD	57330	13.6887	-	7470A	PPB	17 Oct 2019 10:10:52	U	SMPL
PB124048BL - 1	PBW	-31	-0.0450	-	7470A	PPB	17 Oct 2019 10:12:59	U	SMPL
CCV56 - 1	CCV56	22559	5.3636	-	7470A	PPB	17 Oct 2019 10:15:06	U	SMPL
CCB56 - 1	CCB56	18	-0.0333	-	7470A	PPB	17 Oct 2019 10:17:13	U	SMPL
PB124048BS - 1	LCSW	16625	3.9429	-	7470A	PPB	17 Oct 2019 10:19:21	U	SMPL
K5337-02 - 1	HHT-WATER	-18	-0.0419	-	7470A	PPB	17 Oct 2019 10:21:28	U	SMPL
K5337-02DUP - 1	HHT-WATERDUP	17	-0.0336	-	7470A	PPB	17 Oct 2019 10:23:35	U	SMPL
K5337-02MS - 1	HHT-WATERMS	17450	4.1404	-	7470A	PPB	17 Oct 2019 10:25:42	U	SMPL
K5337-02MSD - 1	HHT-WATERMSD	17979	4.2670	-	7470A	PPB	17 Oct 2019 10:27:50	U	SMPL
PB124058BL - 1	PBW	9	-0.0355	-	7470A	PPB	17 Oct 2019 10:29:57	U	SMPL
PB124058BS - 1	LCSW	17146	4.0676	-	7470A	PPB	17 Oct 2019 10:32:05	U	SMPL
K5407-01 - 1	DEVELOPMENT-WATER	136	-0.0051	-	7470A	PPB	17 Oct 2019 10:34:14	U	SMPL
K5407-01DUP - 1	DEVELOPMENT-WATERDUP	215	0.0139	-	7470A	PPB	17 Oct 2019 10:36:22	U	SMPL
K5407-01MS - 1	DEVELOPMENT-WATERMS	17763	4.2153	-	7470A	PPB	17 Oct 2019 10:38:29	U	SMPL
CCV57 - 1	CCV57	21866	5.1977	-	7470A	PPB	17 Oct 2019 10:40:36	U	SMPL
CCB57 - 1	CCB57	25	-0.0316	-	7470A	PPB	17 Oct 2019 10:42:43	U	SMPL
K5407-01MSD - 1	DEVELOPMENT-WATERMSD	16863	3.9998	-	7470A	PPB	17 Oct 2019 10:44:50	U	SMPL
PB124045TB - 1	PB12498TB	6	-0.0362	-	7470A	PPB	17 Oct 2019 10:46:56	U	SMPL
PB124046TB - 1	PB124952TB	25	-0.0316	-	7470A	PPB	17 Oct 2019 10:49:04	U	SMPL
PB124047TB - 1	PB124964TB	23	-0.0321	-	7470A	PPB	17 Oct 2019 10:51:10	U	SMPL
PB124048TB - 1	PB124954TB	28	-0.0309	-	7470A	PPB	17 Oct 2019 10:53:18	U	SMPL
K5152-09LX5 - 1		122	-0.0084	-	7470A	PPB	17 Oct 2019 10:55:25	U	SMPL
K5152-09A - 1		19216	4.5632	-	7470A	PPB	17 Oct 2019 10:57:33	U	SMPL
K5357-02LX5 - 1		9	-0.0355	-	7470A	PPB	17 Oct 2019 10:59:41	U	SMPL
K5357-02A - 1		16598	3.9364	-	7470A	PPB	17 Oct 2019 11:01:49	U	SMPL
K5358-04LX5 - 1		6688	1.5637	-	7470A	PPB	17 Oct 2019 11:03:57	U	SMPL
CCV58 - 1	CCV58	19948	4.7385	-	7470A	PPB	17 Oct 2019 11:06:05	U	SMPL
CCB58 - 1	CCB58	7	-0.0359	-	7470A	PPB	17 Oct 2019 11:08:12	U	SMPL
K5358-04A - 1		54063	12.9065	-	7470A	PPB	17 Oct 2019 11:10:20	U	SMPL
K5337-02LX5 - 1		-60	-0.0520	-	7470A	PPB	17 Oct 2019 11:12:27	U	SMPL
K5337-02A - 1		17768	4.2165	-	7470A	PPB	17 Oct 2019 11:14:35	U	SMPL
K5407-01LX5 - 1		19	-0.0331	-	7470A	PPB	17 Oct 2019 11:16:42	U	SMPL
K5407-01A - 1		17076	4.0508	-	7470A	PPB	17 Oct 2019 11:18:50	U	SMPL
K5358-04DLX2 - 1		17902	4.2486	-	7470A	PPB	17 Oct 2019 11:20:59	U	SMPL
K5358-04DUPDLX2 - 1		17602	4.1768	-	7470A	PPB	17 Oct 2019 11:23:07	U	SMPL
K5358-04MSDLX2 - 1		24391	5.8022	-	7470A	PPB	17 Oct 2019 11:25:15	U	SMPL
K5358-04MSDDLX2 - 1		28019	6.6709	-	7470A	PPB	17 Oct 2019 11:27:44	U	SMPL
K5358-04DL2X10 - 1		3185	0.7250	-	7470A	PPB	17 Oct 2019 11:29:51	U	SMPL
CCV59 - 1	CCV59	19051	4.5237	-	7470A	PPB	17 Oct 2019 11:31:59	U	SMPL
CCB59 - 1	CCB59	-32	-0.0453	-	7470A	PPB	17 Oct 2019 11:34:06	U	SMPL
K5358-04ADLX2 - 1		26163	6.2265	-	7470A	PPB	17 Oct 2019 11:36:14	U	SMPL
CCV60 - 1	CCV60	21454	5.0990	-	7470A	PPB	17 Oct 2019 11:46:03	U	SMPL
CCB60 - 1	CCB60	53	-0.0249	-	7470A	PPB	17 Oct 2019 11:48:23	U	SMPL