



Analytical Summary Report

Analysis Method: SM2130 B

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105764

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP79076
Turbidity Calibration std, 0NTU	WP79079
Turbidity Calibration std, 20NTU	WP79072
Turbidity Calibration std, 80NTU	WP79070
Turbidity Calibration std, 40NTU	WP79071
Turbidity Calibration std, 1NTU	WP79075
Turbidity Calibration std, 10NTU	WP79073
Turbidity Calibration std, 5NTU	WP79074
TURBIDITY LOD STD, 0.5NTU	WP79077
Turbidity LOQ std, 1.0NTU	WP79078
10 NTU Standard 500 ml	W2489

Intercept: 0.1258 Slope: 0.9352 Regression: 0.9998						
Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.144	10/17/2019	10:14
2	CAL2	1	1	0.992	10/17/2019	10:17
3	CAL3	5	1	4.860	10/17/2019	10:21
4	CAL4	10	1	9.850	10/17/2019	10:25
5	CAL5	20	1	19.200	10/17/2019	10:28
6	CAL6	40	1	38.800	10/17/2019	10:31
7	CAL7	80	1	72.100	10/17/2019	10:35
8	CAL8	200	1	188.000	10/17/2019	10:38



Analytical Summary Report

Analysis Method: SM2130 B
Parameter: Turbidity
Run Number: LB105764

ANALYST: AMANDEEP
SUPERVISOR REVIEW BY: apatel

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.120	9.618	10/17/2019	10:33
2	ICB		1	0.150	0.026	10/17/2019	10:37
3	CCV1	10	1	9.650	10.185	10/17/2019	11:08
4	CCB1		1	0.143	0.019	10/17/2019	11:11
5	LB105764BL		1	0.145	0.021	10/17/2019	11:15
6	K5111-10		1	0.568	0.473	10/17/2019	11:25
7	K5111-11		1	1.110	1.053	10/17/2019	11:28
8	K5416-01		1	106.000	113.211	10/17/2019	11:30
9	K5417-01		1	70.100	74.823	10/17/2019	11:32
10	K5418-01		1	39.100	41.675	10/17/2019	11:35
11	K5418-01DUP		1	39.000	41.568	10/17/2019	11:37
12	CCV2	10	1	9.610	10.142	10/17/2019	11:39
13	CCB2		1	0.152	0.029	10/17/2019	11:43

WORKLIST(Hardcopy Internal Chain)

CB105764

WorkList Name : TUR101719

WorkList ID : 132255

Department : Wet-Chemistry

Date : 10-17-2019 10:56:20

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/18/2019	Water	K5416-01	Turbidity	Cool 4 deg C	LIRO01	D11	UDS-2-01	10/16/2019	SM2130 B
10/18/2019	Water	K5417-01	Turbidity	Cool 4 deg C	LIRO01	D11	UDS-4-16A	10/16/2019	SM2130 B
10/18/2019	Water	K5418-01	Turbidity	Cool 4 deg C	LIRO01	D11	UDS-4-26	10/16/2019	SM2130 B

Date/Time 10-17-19 11:20
 Received by: AV
 Relinquished by: JF

Date/Time 10-17-19 11:50
 Received by: SP
 Relinquished by: AV