



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105765

Reagent/Standard	Lot/Loq #
TURBIDITY CCV 10 NTU	WP79076
TURBIDITY LOD STD, 0.5NTU	WP79077
Turbidity Calibration std, 0NTU	WP79079
Turbidity Calibration std, 20NTU	WP79072
Turbidity Calibration std, 40NTU	WP79071
Turbidity Calibration std, 1NTU	WP79075
Turbidity Calibration std, 10NTU	WP79073
Turbidity Calibration std, 5NTU	WP79074
Turbidity LOQ std, 1.0NTU	WP79078
10 NTU Standard 500 ml	W2489

Intercept: 0.0648

Slope: 0.9665

Regression: 1

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.144	10/17/2019	10:14
2	CAL2	1	1	0.992	10/17/2019	10:17
3	CAL3	5	1	4.860	10/17/2019	10:21
4	CAL4	10	1	9.850	10/17/2019	10:25
5	CAL5	20	1	19.200	10/17/2019	10:28
6	CAL6	40	1	38.800	10/17/2019	10:31



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105765

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.120	9.369	10/17/2019	10:33
2	ICB		1	0.150	0.088	10/17/2019	10:37
3	CCV1	10	1	9.600	9.866	10/17/2019	10:40
4	CCB1		1	0.148	0.086	10/17/2019	10:52
5	LB105765BL		1	0.149	0.087	10/17/2019	10:55
6	K5111-07		1	0.565	0.518	10/17/2019	10:58
7	K5111-08		1	1.120	1.092	10/17/2019	11:04
8	CCV2	10	1	9.650	9.917	10/17/2019	11:08
9	CCB2		1	0.143	0.081	10/17/2019	11:11