


LB105788

ISM02.4

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.3211e-004

C= -1.5024e-002

Rho= 0.9999577

Accept=Accepted

Slope
y-intercept
Slope
y-intercept

RS
10-12-19

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.009	0.009	105	0.000	105					
0.05	0.050	0.049	-0.001	276	0.0 %	276					
0.20	0.200	0.219	0.019	1010	0.0 %	1010					
2.50	2.500	2.498	-0.002	10829	0.0 %	10829					
5.00	5.000	4.981	-0.019	21526	0.0 %	21526					
7.50	7.500	7.436	-0.064	32101	0.0 %	32101					
10.0	10.000	10.057	0.057	43393	0.0 %	43393					

RS 10-12-19

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0 S0.0	105	-	0	ISM02.4	PPB	10/18/2019 10:56	Std	
0.05 S0.05	276	-	0.05	ISM02.4	PPB	10/18/2019 10:58	Std	
0.2 S0.2	1010	-	0.2	ISM02.4	PPB	10/18/2019 11:00	Std	
2.5 S2.5	10829	-	2.5	ISM02.4	PPB	10/18/2019 11:02	Std	
5 S5.0	21526	-	5	ISM02.4	PPB	10/18/2019 11:05	Std	
7.5 S7.5	32101	-	7.5	ISM02.4	PPB	10/18/2019 11:08	Std	
10 S10.0	43393	-	10	ISM02.4	PPB	10/18/2019 11:10	Std	
ICV	ICV	16676	3.8556	-	ISM02.4	PPB	10/18/2019 11:14	SMPL
ICB	ICB	32	-0.0076	-	ISM02.4	PPB	10/18/2019 11:16	SMPL
CCV072	CCV072	21331	4.936	-	ISM02.4	PPB	10/18/2019 11:18	SMPL
CCB072	CCB072	19	-0.0106	-	ISM02.4	PPB	10/18/2019 11:21	SMPL
PB124113E	PBW113	65	0.0001	-	ISM02.4	PPB	10/18/2019 11:23	SMPL
K5274-01	MH0B02	103	0.0089	-	ISM02.4	PPB	10/18/2019 11:25	SMPL
K5274-02	MH0B03	124	0.0138	-	ISM02.4	PPB	10/18/2019 11:27	SMPL
K5274-03	MH0B05	116	0.0119	-	ISM02.4	PPB	10/18/2019 11:30	SMPL
K5274-04	MH0B06	146	0.0189	-	ISM02.4	PPB	10/18/2019 11:32	SMPL
K5274-05	MH0B07	183	0.0275	-	ISM02.4	PPB	10/18/2019 11:34	SMPL
K5274-06	MH0B08	86	0.0049	-	ISM02.4	PPB	10/18/2019 11:36	SMPL
K5274-07	MH0B09	87	0.0052	-	ISM02.4	PPB	10/18/2019 11:39	SMPL
K5274-08	MH0B10	108	0.01	-	ISM02.4	PPB	10/18/2019 11:41	SMPL
K5274-09	MH0B11	82	0.004	-	ISM02.4	PPB	10/18/2019 11:43	SMPL
K5274-10	MH0B12	74	0.0022	-	ISM02.4	PPB	10/18/2019 11:46	SMPL
K5274-11	MH0B13	88	0.0054	-	ISM02.4	PPB	10/18/2019 11:48	SMPL
K5274-12	MH0B14	324	0.0602	-	ISM02.4	PPB	10/18/2019 11:53	SMPL
K5274-13	MH0B16	340	0.0639	-	ISM02.4	PPB	10/18/2019 11:55	SMPL
K5274-14	MH0B16D	332	0.062	-	ISM02.4	PPB	10/18/2019 11:58	SMPL
K5274-15	MH0B16S	4573	1.0464	-	ISM02.4	PPB	10/18/2019 12:00	SMPL
K5274-16	MH0B17	405	0.079	-	ISM02.4	PPB	10/18/2019 12:02	SMPL
K5274-17	MH0B18	211	0.0339	-	ISM02.4	PPB	10/18/2019 12:04	SMPL
K5274-18	MH0B20	229	0.0381	-	ISM02.4	PPB	10/18/2019 12:07	SMPL
K5274-19	MH0B21	311	0.0572	-	ISM02.4	PPB	10/18/2019 12:09	SMPL
CCV073	CCV073	21177	4.9003	-	ISM02.4	PPB	10/18/2019 12:17	SMPL
CCB073	CCB073	77	0.0028	-	ISM02.4	PPB	10/18/2019 12:20	SMPL
K5274-20	MH0B22	108	0.01	-	ISM02.4	PPB	10/18/2019 12:31	SMPL
K5274-21	MH0B23	181	0.027	-	ISM02.4	PPB	10/18/2019 12:33	SMPL
K5274-22	MH0B24	138	0.017	-	ISM02.4	PPB	10/18/2019 12:36	SMPL
PB124114E	PBW114	57	-0.0018	-	ISM02.4	PPB	10/18/2019 12:38	SMPL
K5275-01	MH0B25	142	0.0179	-	ISM02.4	PPB	10/18/2019 12:40	SMPL
K5275-02	MH0B26	216	0.0351	-	ISM02.4	PPB	10/18/2019 12:43	SMPL
K5275-03	MH0B27	228	0.0379	-	ISM02.4	PPB	10/18/2019 12:46	SMPL
K5275-04	MH0B28	139	0.0172	-	ISM02.4	PPB	10/18/2019 12:48	SMPL
K5275-05	MH0B28D	115	0.0117	-	ISM02.4	PPB	10/18/2019 12:50	SMPL
K5275-06	MH0B28S	4308	0.9849	-	ISM02.4	PPB	10/18/2019 12:52	SMPL
K5275-07	MH0B29	237	0.04	-	ISM02.4	PPB	10/18/2019 12:55	SMPL
K5275-08	MH0B32	296	0.0537	-	ISM02.4	PPB	10/18/2019 12:57	SMPL
K5275-09	MH0B33	83	0.0042	-	ISM02.4	PPB	10/18/2019 12:59	SMPL

INSTRUMENT ID : CV1

K5275-10	MH0B55	53	-0.0027 -	ISM02.4	PPB	10/18/2019 13:02 SMPL
PB124115E	PBW115	79	0.0033 -	ISM02.4	PPB	10/18/2019 13:04 SMPL
K5294-01	MBF4K6	199	0.0312 -	ISM02.4	PPB	10/18/2019 13:06 SMPL
K5294-02	MBF4K6D	199	0.0312 -	ISM02.4	PPB	10/18/2019 13:08 SMPL
K5294-03	MBF4K6S	4451	1.0181 -	ISM02.4	PPB	10/18/2019 13:11 SMPL
K5294-04	MBF546	247	0.0423 -	ISM02.4	PPB	10/18/2019 13:13 SMPL
K5294-05	MBF4K8	214	0.0346 -	ISM02.4	PPB	10/18/2019 13:15 SMPL
CCV074	CCV074	21085	4.8789 -	ISM02.4	PPB	10/18/2019 13:17 SMPL
CCB074	CCB074	15	-0.0115 -	ISM02.4	PPB	10/18/2019 13:20 SMPL
K5294-06	MBF530	198	0.0309 -	ISM02.4	PPB	10/18/2019 13:22 SMPL
K5294-07	MBF531	546	0.1117 -	ISM02.4	PPB	10/18/2019 13:24 SMPL
K5294-08	MBF4K2	239	0.0404 -	ISM02.4	PPB	10/18/2019 13:27 SMPL
K5294-09	MBF533	299	0.0544 -	ISM02.4	PPB	10/18/2019 13:29 SMPL
K5294-10	MBF558	136	0.0165 -	ISM02.4	PPB	10/18/2019 13:31 SMPL
K5294-11	MBF532	158	0.0216 -	ISM02.4	PPB	10/18/2019 13:33 SMPL
K5294-12	MBF534	170	0.0244 -	ISM02.4	PPB	10/18/2019 13:36 SMPL
K5294-13	MBF549	145	0.0186 -	ISM02.4	PPB	10/18/2019 13:38 SMPL
K5294-14	MBF4K4	113	0.0112 -	ISM02.4	PPB	10/18/2019 13:40 SMPL
K5294-15	MBF4K5	102	0.0087 -	ISM02.4	PPB	10/18/2019 13:43 SMPL
K5294-16	MBF548	126	0.0142 -	ISM02.4	PPB	10/18/2019 13:45 SMPL
K5294-17	MBF4K7	269	0.0474 -	ISM02.4	PPB	10/18/2019 13:47 SMPL
K5294-18	MBF4K9	132	0.0156 -	ISM02.4	PPB	10/18/2019 13:50 SMPL
CCV075	CCV075	20703	4.7902 -	ISM02.4	PPB	10/18/2019 13:52 SMPL
CCB075	CCB075	41	-0.0055 -	ISM02.4	PPB	10/18/2019 13:54 SMPL