



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/21/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 10/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:13
Out Date: 10/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105792

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5443-01	MH13K9	1	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-02	MH13L0	2	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-03	MH13L1	3	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-04	MH13L2	4	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-05	MH13L3	5	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-06	MH13L4	6	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-07	MH13L5	7	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-08	MH13L6	8	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-09	MH13L7	9	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-10	MH13L8	10	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-11	MH13L9	11	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-12	MH13M0	12	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-13	MH13M1	13	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-14	MH13M2	14	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-15	MH13M3	15	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-16	MH13M4	16	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-17	MH13M5	17	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-18	MH13M6	18	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-19	MH13M7	19	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-20	MH13M8	20	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-21	MH13M8D	21	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS
K5443-22	MH13M8S	22	1.00	2.00	2.00	100.0	DUST SOIL 100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$