



PERCENT SOLID

Supervisor: apatel
Analyst: JIGNESH
Date: 10/21/2019

OVENTEMP IN Celsius(°C): 107
Time IN: 13:30
In Date: 10/18/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN-1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:32
Out Date: 10/19/2019
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: %SOLIDS-OVEN

QC:LB105794

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
K5445-01	MH12H7	1	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-02	MH12H7D	2	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-03	MH12H7S	3	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-04	MH12S7	4	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-05	MH12S8	5	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-06	MH12S9	6	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-07	MH12T0	7	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-08	MH12T1	8	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-09	MH12T2	9	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-10	MH12T3	10	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-11	MH12T4	11	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-12	MH12T5	12	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-13	MH12T6	13	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-14	MH1369	14	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-15	MH1370	15	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-16	MH1371	16	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-17	MH1372	17	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-18	MH1373	18	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-19	MH1374	19	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-20	MH1375	20	1.00	2.00	2.00	100.0	100% SOLIDS
K5445-21	MH1376	21	1.00	2.00	2.00	100.0	100% SOLIDS

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$