

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus, TotalSUPERVISOR REVIEW BY: apatelRun Number: LB105797

BALANCE ID: _____

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP76600
calibration std. phosphate 1 ppm	WP79141
calibration std. phosphate 0.5 ppm	WP79140
calibration std. phosphate 0.3 ppm	WP79139
calibration std. 0 ppm	WP79136
phosphate ICV std.	WP79143
calibration std. phosphate 0.1 ppm	WP79138
calibration std. phosphate 0.05 ppm	WP79137
Combined reagent	WP79146
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314
phosphate CCV std.	WP79142
PHOSPHATE LOD STD, 0.01PPM	WP79147

Intercept: -0.0006Slope: 0.6487Regression: 0.999856

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/18/2019	14:20
2	CAL2	0.05	1	50	50	0.033	10/18/2019	14:20
3	CAL3	0.10	1	50	50	0.066	10/18/2019	14:21
4	CAL4	0.30	1	50	50	0.186	10/18/2019	14:22
5	CAL5	0.50	1	50	50	0.328	10/18/2019	14:22
6	CAL6	1.00	1	50	50	0.648	10/18/2019	14:23



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus, Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB105797

BALANCE ID: _____

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.497	10/18/2019	14:23
2	ICB		1	50	50	0.000	0.001	10/18/2019	14:24
3	CCV1	0.50	1	50	50	0.327	0.505	10/18/2019	14:25
4	CCB1		1	50	50	0.000	0.001	10/18/2019	14:25
5	PB124120BL		1	1.01	50	0.000	0.001	10/18/2019	14:26
6	PB124120BS	0.50	1	1.00	50	0.320	0.494	10/18/2019	14:26
7	K5111-03		1	1.01	50	0.012	0.019	10/18/2019	14:27
8	CCV2	0.50	1	50	50	0.329	0.508	10/18/2019	14:28
9	CCB2		1	50	50	0.000	0.001	10/18/2019	14:28