

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB105798

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP76600
calibration std. phosphate 1 ppm	WP79141
calibration std. phosphate 0.5 ppm	WP79140
calibration std. phosphate 0.3 ppm	WP79139
calibration std. 0 ppm	WP79136
phosphate ICV std.	WP79143
calibration std. phosphate 0.1 ppm	WP79138
calibration std. phosphate 0.05 ppm	WP79137
Combined reagent	WP79146
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314
phosphate CCV std.	WP79142

Intercept: -0.0006Slope: 0.6487Regression: 0.999856

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/18/2019	14:20
2	CAL2	0.05	1	50	50	0.033	10/18/2019	14:20
3	CAL3	0.10	1	50	50	0.066	10/18/2019	14:21
4	CAL4	0.30	1	50	50	0.186	10/18/2019	14:22
5	CAL5	0.50	1	50	50	0.328	10/18/2019	14:22
6	CAL6	1.00	1	50	50	0.648	10/18/2019	14:23



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB105798

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.322	0.497	10/18/2019	14:23
2	ICB		1	50	50	0.000	0.001	10/18/2019	14:24
3	CCV1	0.50	1	50	50	0.329	0.508	10/18/2019	14:28
4	CCB1		1	50	50	0.000	0.001	10/18/2019	14:28
5	PB124119BL		1	50	50	0.000	0.001	10/18/2019	14:29
6	PB124119BS	0.50	1	50	50	0.321	0.496	10/18/2019	14:30
7	K5410-04		1	50	50	0.303	0.468	10/18/2019	14:31
8	K5410-07		1	50	50	0.346	0.534	10/18/2019	14:31
9	K5411-04		1	50	50	0.538	0.830	10/18/2019	14:33
10	K5411-05		1	50	50	0.569	0.878	10/18/2019	14:33
11	K5411-05DUP		1	50	50	0.566	0.873	10/18/2019	14:34
12	K5411-05MS	0.50	1	50	50	0.639	0.986	10/18/2019	14:34
13	K5411-05MSD	0.50	1	50	50	0.645	0.995	10/18/2019	14:35
14	K5456-01		1	50	50	2.299	3.545	10/18/2019	14:36
15	CCV2	0.50	1	50	50	0.325	0.502	10/18/2019	14:37
16	CCB2		1	50	50	0.000	0.001	10/18/2019	14:37
17	K5456-01		50	50	50	0.471	0.727	10/18/2019	14:38
18	CCV3	0.50	1	50	50	0.326	0.503	10/18/2019	14:38
19	CCB3		1	50	50	0.000	0.001	10/18/2019	14:39