


7470A

LB105803
INSTRUMENT ID: CV1

Linear

 μ Abs.:

40775



A= 0.0000e+000

B= 2.4288e-004

C= -7.3824e-002

Rho= 0.9995662

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.050	-0.050	97	0.000	97				
0.2	0.200	0.168	-0.032	995	0.0 %	995				
2.5	2.500	2.478	-0.022	10507	0.0 %	10507				
5.0	5.000	5.146	0.146	21493	0.0 %	21493				
7.5	7.500	7.628	0.128	31710	0.0 %	31710				
10.0	10.000	9.830	-0.170	40775	0.0 %	40775				

LB105803 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 18 Oct 2019 14:28:34

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	S0.0	97	-	0.0000	7470A	PPB	18 Oct 2019 14:37:41	S	Std
0.2 - 1	S0.2	995	-	0.2000	7470A	PPB	18 Oct 2019 14:39:51	S	Std
2.5 - 1	S2.5	10507	-	2.5000	7470A	PPB	18 Oct 2019 14:42:00	S	Std
5.0 - 1	S5.0	21493	-	5.0000	7470A	PPB	18 Oct 2019 14:44:08	S	Std
7.5 - 1	S7.5	31710	-	7.5000	7470A	PPB	18 Oct 2019 14:46:16	S	Std
10.0 - 1	S10.0	40775	-	10.0000	7470A	PPB	18 Oct 2019 14:48:24	S	Std
ICV64 - 1	ICV64	16088	3.8337	-	7470A	PPB	18 Oct 2019 14:54:21	U	SMPL
ICB64 - 1	ICB64	50	-0.0617	-	7470A	PPB	18 Oct 2019 14:56:27	U	SMPL
CCV76 - 1	CCV76	21223	5.0809	-	7470A	PPB	18 Oct 2019 14:58:33	U	SMPL
CCB76 - 1	CCB76	44	-0.0631	-	7470A	PPB	18 Oct 2019 15:00:40	U	SMPL
CRA - 1	CRA	974	0.1627	-	7470A	PPB	18 Oct 2019 15:02:46	U	SMPL
HighStd - 1	HighStd	42146	10.1628	-	7470A	PPB	18 Oct 2019 15:04:54	U	SMPL
ChkStd - 1	ChkStd	29579	7.1104	-	7470A	PPB	18 Oct 2019 15:07:00	U	SMPL
PB124127BL - 1	PBW	-1	-0.0741	-	7470A	PPB	18 Oct 2019 15:09:08	U	SMPL
PB124127BS - 1	LCSW	16724	3.9882	-	7470A	PPB	18 Oct 2019 15:11:17	U	SMPL
K5153-04 - 1	TR-04-101619	96	-0.0505	-	7470A	PPB	18 Oct 2019 15:13:25	U	SMPL
K5153-04DUP - 1	TR-04-101619DUP	108	-0.0476	-	7470A	PPB	18 Oct 2019 15:15:34	U	SMPL
K5153-04MS - 1	TR-04-101619MS	17286	4.1247	-	7470A	PPB	18 Oct 2019 15:17:43	U	SMPL
K5153-04MSD - 1	TR-04-101619MSD	16526	3.9401	-	7470A	PPB	18 Oct 2019 15:19:52	U	SMPL
K5393-20 - 1	T-6-D1-D3-(0-46)	78	-0.0549	-	7470A	PPB	18 Oct 2019 15:21:59	U	SMPL
CCV77 - 1	CCV77	21801	5.2213	-	7470A	PPB	18 Oct 2019 15:24:05	U	SMPL
CCB77 - 1	CCB77	18	-0.0695	-	7470A	PPB	18 Oct 2019 15:26:12	U	SMPL
K5393-22 - 1	T-9-U1-U3-(0-28)	82	-0.0539	-	7470A	PPB	18 Oct 2019 15:28:19	U	SMPL
K5393-24 - 1	T-13-V1-(0-24)	48	-0.0622	-	7470A	PPB	18 Oct 2019 15:30:26	U	SMPL
K5393-26 - 1	T-16-G1-(0-27)	60	-0.0593	-	7470A	PPB	18 Oct 2019 15:32:33	U	SMPL
K5394-07 - 1	T-17-F1-(0-30)	66	-0.0578	-	7470A	PPB	18 Oct 2019 15:34:41	U	SMPL
K5394-09 - 1	T-15-E1-(0-20)	68	-0.0573	-	7470A	PPB	18 Oct 2019 15:36:49	U	SMPL
PB124127TB - 1	PB124031TB	86	-0.0529	-	7470A	PPB	18 Oct 2019 15:38:57	U	SMPL
K5153-04LX5 - 1		21	-0.0687	-	7470A	PPB	18 Oct 2019 15:41:05	U	SMPL
K5153-04A - 1		16752	3.9950	-	7470A	PPB	18 Oct 2019 15:43:14	U	SMPL
CCV78 - 1	CCV78	21602	5.1730	-	7470A	PPB	18 Oct 2019 15:45:23	U	SMPL
CCB78 - 1	CCB78	20	-0.0690	-	7470A	PPB	18 Oct 2019 15:47:30	U	SMPL