



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105826

Reagent/Standard	Lot/Log #
TURBIDITY CCV 10 NTU	WP79186
TURBIDITY LOD STD, 0.5NTU	WP79189
Turbidity Calibration std, 0NTU	WP79185
Turbidity Calibration std, 20NTU	WP79181
Turbidity Calibration std, 40NTU	WP79180
Turbidity Calibration std, 1NTU	WP79184
Turbidity Calibration std, 5NTU	WP79183

Intercept: -0.003

Slope: 0.9606

Regression: 0.9998

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Anal Date	Anal Time
1	CAL1	0	1	0.148	10/21/2019	10:24
2	CAL2	1	1	1.020	10/21/2019	10:28
3	CAL3	5	1	4.840	10/21/2019	10:33
4	CAL4	10	1	9.680	10/21/2019	10:36
5	CAL5	20	1	18.600	10/21/2019	10:40
6	CAL6	40	1	38.700	10/21/2019	10:43



Analytical Summary Report

Analysis Method: 180.1

ANALYST: AMANDEEP

Parameter: Turbidity

SUPERVISOR REVIEW BY: apatel

Run Number: LB105826

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	9.250	9.633	10/21/2019	10:47
2	ICB		1	0.150	0.159	10/21/2019	10:50
3	CCV1	10	1	9.450	9.841	10/21/2019	10:53
4	CCB1		1	0.145	0.154	10/21/2019	10:56
5	LB105826BL		1	0.148	0.157	10/21/2019	10:59
6	K5111-09		1	0.506	0.530	10/21/2019	11:05
7	CCV2	10	1	9.520	9.914	10/21/2019	11:07
8	CCB2		1	0.149	0.158	10/21/2019	11:10