


LB105852

7470A

INSTRUMENT ID: CV1



Linear

A= 0.0000e+000

B= 2.5496e-004

C= -2.5760e-002

Rho= 0.9999672

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.011	-0.011	57	0.000	57				
0.2	0.200	0.203	0.003	898	0.0 %	898				
2.5	2.500	2.510	0.010	9947	0.0 %	9947				
5.0	5.000	4.973	-0.027	19606	0.0 %	19606				
7.5	7.500	7.558	0.058	29743	0.0 %	29743				
10.0	10.000	9.967	-0.033	39194	0.0 %	39194				

LB105852 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 22 Oct 2019 10:09:39			
Sample ID	Extended ID	μ Abs.	Conc.	Std ConcMethod	Units	Date	Type
0.0 - 1	50.0	57	-	0.00007470A	PPB	22 Oct 2019 10:54:11	S Std
0.2 - 1	50.2	898	-	0.20007470A	PPB	22 Oct 2019 10:56:27	S Std
2.5 - 1	52.5	9947	-	2.50007470A	PPB	22 Oct 2019 10:58:36	S Std
5.0 - 1	55.0	19606	-	5.00007470A	PPB	22 Oct 2019 11:00:44	S Std
7.5 - 1	57.5	29743	-	7.50007470A	PPB	22 Oct 2019 11:02:52	S Std
10.0 - 1	510.0	39194	-	10.00007470A	PPB	22 Oct 2019 11:05:00	S Std
ICV65 - 1	ICV65	16109	4.0814	-7470A	PPB	22 Oct 2019 11:07:48	U SMPL
ICB65 - 1	ICB65	21	-0.0204	-7470A	PPB	22 Oct 2019 11:09:54	U SMPL
CCV79 - 1	CCV79	19418	4.9251	-7470A	PPB	22 Oct 2019 11:12:01	U SMPL
CCB79 - 1	CCB79	1	-0.0255	-7470A	PPB	22 Oct 2019 11:14:07	U SMPL
CRA - 1	CRA	868	0.1955	-7470A	PPB	22 Oct 2019 11:16:14	U SMPL
HighStd - 1	HighStd	38429	9.7721	-7470A	PPB	22 Oct 2019 11:18:20	U SMPL
ChkStd - 1	ChkStd	26939	6.8426	-7470A	PPB	22 Oct 2019 11:20:27	U SMPL
PB124184BL - 1	PBW	131	0.0076	-7470A	PPB	22 Oct 2019 11:22:35	U SMPL
PB124184BS - 1	LCSW	16927	4.2900	-7470A	PPB	22 Oct 2019 11:24:43	U SMPL
K5146-14 - 1	CL-02-101819	143	0.0107	-7470A	PPB	22 Oct 2019 11:26:51	U SMPL
K5146-14DUP - 1	CL-02-101819DUP	157	0.0143	-7470A	PPB	22 Oct 2019 11:28:59	U SMPL
K5146-14MS - 1	CL-02-101819MS	16334	4.1388	-7470A	PPB	22 Oct 2019 11:31:07	U SMPL
K5146-14MSD - 1	CL-02-101819MSD	16184	4.1005	-7470A	PPB	22 Oct 2019 11:35:52	U SMPL
K5147-07 - 1	OK-02-101819	115	0.0036	-7470A	PPB	22 Oct 2019 11:37:58	U SMPL
CCV80 - 1	CCV80	19269	4.8871	-7470A	PPB	22 Oct 2019 11:40:05	U SMPL
CCB80 - 1	CCB80	28	-0.0186	-7470A	PPB	22 Oct 2019 11:42:12	U SMPL
K5150-09 - 1	PL-01-101819	168	0.0171	-7470A	PPB	22 Oct 2019 11:44:20	U SMPL
K5436-03 - 1	3425-28	156	0.0140	-7470A	PPB	22 Oct 2019 11:46:27	U SMPL
K5436-06 - 1	3860-50	165	0.0163	-7470A	PPB	22 Oct 2019 11:48:35	U SMPL
K5439-02 - 1	33RD-ST-7TH-AVE-EAST-COMP	168	0.0171	-7470A	PPB	22 Oct 2019 11:50:42	U SMPL
K5439-03 - 1	33RD-ST-7TH-AVE-WEST-COMP	71	-0.0077	-7470A	PPB	22 Oct 2019 11:52:49	U SMPL
K5447-02 - 1	RT5124	156	0.0140	-7470A	PPB	22 Oct 2019 11:54:57	U SMPL
K5448-06 - 1	SP	131	0.0076	-7470A	PPB	22 Oct 2019 11:57:05	U SMPL
K5453-02 - 1	RT1496	131	0.0076	-7470A	PPB	22 Oct 2019 11:59:12	U SMPL
PB124184TB - 1	PB124116TB	139	0.0097	-7470A	PPB	22 Oct 2019 12:01:21	U SMPL
K5146-14LX5 - 1		32	-0.0176	-7470A	PPB	22 Oct 2019 12:03:28	U SMPL
CCV81 - 1	CCV81	19250	4.8822	-7470A	PPB	22 Oct 2019 12:05:35	U SMPL
CCB81 - 1	CCB81	38	-0.0161	-7470A	PPB	22 Oct 2019 12:07:42	U SMPL
K5146-14A - 1		16442	4.1663	-7470A	PPB	22 Oct 2019 12:09:49	U SMPL
CCV82 - 1	CCV82	19935	5.0569	-7470A	PPB	22 Oct 2019 12:11:56	U SMPL
CCB82 - 1	CCB82	38	-0.0161	-7470A	PPB	22 Oct 2019 12:14:03	U SMPL