

LB105854


ISM02.4

INSTRUMENT ID: CU1

Linear

 μ Abs.:

49687



A= 0.0000e+000

B= 2.0344e-004

C= -1.1335e-002

Rho= 0.9998749

Accept=Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.006	0.006	87	0.000	87					
0.05	0.050	0.058	0.008	339	0.0 %	339					
0.20	0.200	0.234	0.034	1204	0.0 %	1204					
2.50	2.500	2.516	0.016	12423	0.0 %	12423					
5.00	5.000	4.935	-0.065	24312	0.0 %	24312					
7.50	7.500	7.405	-0.095	36456	0.0 %	36456					
10.0	10.000	10.097	0.097	49687	0.0 %	49687					

RS 10-22-19

LB105854

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0.0	87	-		0 ISM02.4	PPB	10/22/2019 13:20	Std
	0.05 S0.05	339	-	0.05	ISM02.4	PPB	10/22/2019 13:22	Std
	0.2 S0.2	1204	-	0.2	ISM02.4	PPB	10/22/2019 13:24	Std
	2.5 S2.5	12423	-	2.5	ISM02.4	PPB	10/22/2019 13:27	Std
	5 S5.0	24312	-	5	ISM02.4	PPB	10/22/2019 13:29	Std
	7.5 S7.5	36456	-	7.5	ISM02.4	PPB	10/22/2019 13:31	Std
	10 S10.0	49687	-	10	ISM02.4	PPB	10/22/2019 13:43	Std
ICV	ICV	19662	3.9886 -		ISM02.4	PPB	10/22/2019 13:46	SMPL
ICB	ICB	38	-0.0036 -		ISM02.4	PPB	10/22/2019 13:48	SMPL
CCV083	CCV083	24566	4.9863 -		ISM02.4	PPB	10/22/2019 13:51	SMPL
CCB083	CCB083	57	0.0003 -		ISM02.4	PPB	10/22/2019 13:53	SMPL
PB124190BL	PBS190	75	0.0039 -		ISM02.4	PPB	10/22/2019 13:55	SMPL
K5237-01	MBEPC2	15105	3.0616 -		ISM02.4	PPB	10/22/2019 13:57	SMPL
K5237-02	MBEPC3	26090	5.2963 -		ISM02.4	PPB	10/22/2019 14:00	SMPL
K5237-03	MBEPL2	68438	13.9114 -		ISM02.4	PPB	10/22/2019 14:02	SMPL
K5237-04	MBEPL3	83911	17.0591 -		ISM02.4	PPB	10/22/2019 14:04	SMPL
K5237-05	MBEPL5	59036	11.9987 -		ISM02.4	PPB	10/22/2019 14:07	SMPL
K5237-06	MBEPL6	18325	3.7166 -		ISM02.4	PPB	10/22/2019 14:09	SMPL
K5237-07	MBEPL8	52851	10.7404 -		ISM02.4	PPB	10/22/2019 14:11	SMPL
K5237-08	MBEPL8D	53753	10.9239 -		ISM02.4	PPB	10/22/2019 14:13	SMPL
K5237-09	MBEPL8S	53718	10.9168 -		ISM02.4	PPB	10/22/2019 14:16	SMPL
K5237-10	MBEPL9	3045	0.6081 -		ISM02.4	PPB	10/22/2019 14:18	SMPL
K5237-11	MBEPM1	19301	3.9152 -		ISM02.4	PPB	10/22/2019 14:20	SMPL
K5237-12	MBEPM2	3488	0.6982 -		ISM02.4	PPB	10/22/2019 14:23	SMPL
K5237-13	MBEPM4	16006	3.2449 -		ISM02.4	PPB	10/22/2019 14:25	SMPL
K5237-14	MBEPM5	12685	2.5692 -		ISM02.4	PPB	10/22/2019 14:27	SMPL
K5237-15	MBEPM7	8958	1.811 -		ISM02.4	PPB	10/22/2019 14:29	SMPL
K5237-16	MBEPM8	22918	4.651 -		ISM02.4	PPB	10/22/2019 14:32	SMPL
K5237-17	MBEPN0	3233	0.6464 -		ISM02.4	PPB	10/22/2019 14:34	SMPL
PB124191BL	PBS191	36	-0.004 -		ISM02.4	PPB	10/22/2019 14:36	SMPL
K5244-01	MH0AZ6	63121	12.8297 -		ISM02.4	PPB	10/22/2019 14:38	SMPL
CCV084	CCV084	24292	4.9305 -		ISM02.4	PPB	10/22/2019 14:41	SMPL
CCB084	CCB084	0	-0.0113 -		ISM02.4	PPB	10/22/2019 14:43	SMPL
K5244-02	MH0AZ7	438	0.0778 -		ISM02.4	PPB	10/22/2019 14:45	SMPL
K5244-03	MH0AZ8	4277	0.8588 -		ISM02.4	PPB	10/22/2019 14:48	SMPL
K5244-04	MH0AZ9	641	0.1191 -		ISM02.4	PPB	10/22/2019 14:50	SMPL
K5244-05	MH0B00	4805	0.9662 -		ISM02.4	PPB	10/22/2019 14:52	SMPL
K5244-06	MH0B01	1215	0.2358 -		ISM02.4	PPB	10/22/2019 14:54	SMPL
K5244-07	MH0B34	3394	0.6791 -		ISM02.4	PPB	10/22/2019 14:57	SMPL
K5244-08	MH0B36	10936	2.2134 -		ISM02.4	PPB	10/22/2019 14:59	SMPL
K5244-09	MH0B37	1152	0.223 -		ISM02.4	PPB	10/22/2019 15:01	SMPL
K5244-10	MH0B38	732836	HIGH -		ISM02.4	PPB	10/22/2019 15:04	
K5244-11	MH0B40	308613	62.7715 -		ISM02.4	PPB	10/22/2019 15:05	SMPL
K5244-12	MH0B41	121012	24.6068 -		ISM02.4	PPB	10/22/2019 15:08	SMPL
K5244-13	MH0B42	34096	6.925 -		ISM02.4	PPB	10/22/2019 15:10	SMPL
K5244-14	MH0B43	65070	13.2262 -		ISM02.4	PPB	10/22/2019 15:12	SMPL

INSTRUMENT ID : CV1

K5244-15	MHOB43D	65200	13.2527 -	ISM02.4	PPB	10/22/2019 15:15	SMPL
K5244-16	MHOB43S	83552	16.9861 -	ISM02.4	PPB	10/22/2019 15:17	SMPL
K5244-17	MHOB44	1732	0.341 -	ISM02.4	PPB	10/22/2019 15:19	SMPL
K5244-18	MHOB45	46404	9.4289 -	ISM02.4	PPB	10/22/2019 15:21	SMPL
K5244-19	MHOB46	85478	17.3779 -	ISM02.4	PPB	10/22/2019 15:24	SMPL
K5244-20	MHOB47	89322	18.1599 -	ISM02.4	PPB	10/22/2019 15:26	SMPL
K5244-21	MHOB48	74466	15.1377 -	ISM02.4	PPB	10/22/2019 15:28	SMPL
CCV085	CCV085	23089	4.6858 -	ISM02.4	PPB	10/22/2019 15:31	SMPL
CCB085	CCB085	13	-0.0087 -	ISM02.4	PPB	10/22/2019 15:33	SMPL
K5244-22	MHOB49	22983	4.6642 -	ISM02.4	PPB	10/22/2019 15:35	SMPL
PB124192BL	PBS192	30	-0.0052 -	ISM02.4	PPB	10/22/2019 15:37	SMPL
K5273-01	MHOB50	20396	4.1379 -	ISM02.4	PPB	10/22/2019 15:40	SMPL
K5273-02	MHOB51	27247	5.5317 -	ISM02.4	PPB	10/22/2019 15:42	SMPL
K5273-03	MHOB52	4229	0.849 -	ISM02.4	PPB	10/22/2019 15:44	SMPL
K5273-04	MHOB53	62947	12.7943 -	ISM02.4	PPB	10/22/2019 15:47	SMPL
K5273-05	MHOB53D	68499	13.9238 -	ISM02.4	PPB	10/22/2019 15:49	SMPL
K5273-06	MHOB53S	77632	15.7818 -	ISM02.4	PPB	10/22/2019 15:51	SMPL
K5273-07	MHOB54	18530	3.7583 -	ISM02.4	PPB	10/22/2019 15:53	SMPL
K5237-03DLX3		23029	4.6736 -	ISM02.4	PPB	10/22/2019 15:56	SMPL
K5237-04DLX3		27717	5.6273 -	ISM02.4	PPB	10/22/2019 15:58	SMPL
K5237-05DLX3		19871	4.0311 -	ISM02.4	PPB	10/22/2019 16:00	SMPL
K5237-07DLX3		17823	3.6145 -	ISM02.4	PPB	10/22/2019 16:03	SMPL
K5237-08DLX3		17258	3.4996 -	ISM02.4	PPB	10/22/2019 16:05	SMPL
K5237-09DLX3		17563	3.5616 -	ISM02.4	PPB	10/22/2019 16:07	SMPL
K5244-01DLX3		19665	3.9892 -	ISM02.4	PPB	10/22/2019 16:09	SMPL
K5244-10DLX10		125263	25.4716 -	ISM02.4	PPB	10/22/2019 16:12	SMPL
K5244-11DLX10		30393	6.1717 -	ISM02.4	PPB	10/22/2019 16:14	SMPL
K5244-12DLX5		22387	4.543 -	ISM02.4	PPB	10/22/2019 16:16	SMPL
K5244-14DLX4		15178	3.0764 -	ISM02.4	PPB	10/22/2019 16:19	SMPL
CCV086	CCV086	23868	4.8443 -	ISM02.4	PPB	10/22/2019 16:21	SMPL
CCB086	CCB086	195	0.0283 -	ISM02.4	PPB	10/22/2019 16:23	SMPL
K5244-15DLX4		15425	3.1267 -	ISM02.4	PPB	10/22/2019 16:25	SMPL
K5244-16DLX4		21158	4.293 -	ISM02.4	PPB	10/22/2019 16:28	SMPL
K5244-19DLX5		17200	3.4878 -	ISM02.4	PPB	10/22/2019 16:30	SMPL
K5244-20DLX5		18040	3.6586 -	ISM02.4	PPB	10/22/2019 16:32	SMPL
K5244-21DLX5		14889	3.0176 -	ISM02.4	PPB	10/22/2019 16:35	SMPL
K5273-04DLX4		15676	3.1777 -	ISM02.4	PPB	10/22/2019 16:37	SMPL
K5273-05DLX4		16550	3.3555 -	ISM02.4	PPB	10/22/2019 16:39	SMPL
K5273-06DLX4		18916	3.8369 -	ISM02.4	PPB	10/22/2019 16:41	SMPL
K5244-10DL2X50		23475	4.7643 -	ISM02.4	PPB	10/22/2019 16:44	SMPL
CCV087	CCV087	23485	4.7663 -	ISM02.4	PPB	10/22/2019 16:46	SMPL
CCB087	CCB087	-16	-0.0146 -	ISM02.4	PPB	10/22/2019 16:48	SMPL