

7471B


LB105874

INSTRUMENT ID : CV1

Linear

 μ Abs.:

59333



A= 0.0000e+000

B= 1.7096e-004

C= -2.4098e-003

Rho= 0.9996296

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	0.031	0.031	198	0.000	198				
0.2	0.200	0.215	0.015	1271	0.0 %	1271				
2.5	2.500	2.517	0.017	14736	0.0 %	14736				
5.0	5.000	4.994	-0.006	29225	0.0 %	29225				
7.5	7.500	7.302	-0.198	42728	0.0 %	42728				
10.0	10.000	10.141	0.141	59333	0.0 %	59333				

LB105874 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 23 Oct 2019 07:41:59

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50.0	198	-	0.0000	7471B	PPB	23 Oct 2019 08:02:15	S	Std
0.2 - 1	50.2	1271	-	0.2000	7471B	PPB	23 Oct 2019 08:04:40	S	Std
2.5 - 1	52.5	14736	-	2.5000	7471B	PPB	23 Oct 2019 08:06:58	S	Std
5.0 - 1	55.0	29225	-	5.0000	7471B	PPB	23 Oct 2019 08:09:16	S	Std
7.5 - 1	57.5	42728	-	7.5000	7471B	PPB	23 Oct 2019 08:11:33	S	Std
10.0 - 1	510.0	59333	-	10.0000	7471B	PPB	23 Oct 2019 08:16:55	S	Std
ICV68 - 1	ICV68	23423	4.0019	-	7471B	PPB	23 Oct 2019 08:20:54	U	SMPL
ICB68 - 1	ICB68	169	0.0265	-	7471B	PPB	23 Oct 2019 08:23:11	U	SMPL
CCV88 - 1	CCV88	29316	5.0093	-	7471B	PPB	23 Oct 2019 08:25:27	U	SMPL
CCB88 - 1	CCB88	233	0.0374	-	7471B	PPB	23 Oct 2019 08:27:43	U	SMPL
CRA - 1	CRA	1308	0.2212	-	7471B	PPB	23 Oct 2019 08:29:59	U	SMPL
HighStd - 1	HighStd	58485	9.9959	-	7471B	PPB	23 Oct 2019 08:32:15	U	SMPL
ChkStd - 1	ChkStd	40282	6.8840	-	7471B	PPB	23 Oct 2019 08:34:30	U	SMPL
PB124198BL - 1	PBS	-16	-0.0051	-	7471B	PPB	23 Oct 2019 08:36:48	U	SMPL
PB124198BS - 1	LCSS	20683	3.5335	-	7471B	PPB	23 Oct 2019 08:39:04	U	SMPL
K5139-01 - 1	BU-03-102119	2530	0.4301	-	7471B	PPB	23 Oct 2019 08:41:21	U	SMPL
K5146-08 - 1	CL-02-101819	4237	0.7219	-	7471B	PPB	23 Oct 2019 08:43:38	U	SMPL
K5147-05 - 1	OK-02-101819	2141	0.3636	-	7471B	PPB	23 Oct 2019 08:45:56	U	SMPL
K5150-01 - 1	PL-01-101819	1997	0.3390	-	7471B	PPB	23 Oct 2019 08:48:13	U	SMPL
K5150-01DUP - 1	PL-01-101819DUP	2037	0.3458	-	7471B	PPB	23 Oct 2019 08:50:29	U	SMPL
CCV89 - 1	CCV89	30123	5.1473	-	7471B	PPB	23 Oct 2019 08:52:46	U	SMPL
CCB89 - 1	CCB89	169	0.0265	-	7471B	PPB	23 Oct 2019 08:55:03	U	SMPL
K5150-01MS - 1	PL-01-101819MS	23251	3.9725	-	7471B	PPB	23 Oct 2019 08:57:19	U	SMPL
K5150-01MSD - 1	PL-01-101819MSD	23091	3.9451	-	7471B	PPB	23 Oct 2019 08:59:35	U	SMPL
K5447-01 - 1	RT5124	1538	0.2605	-	7471B	PPB	23 Oct 2019 09:01:52	U	SMPL
K5448-01 - 1	SP	1428	0.2417	-	7471B	PPB	23 Oct 2019 09:04:09	U	SMPL
K5453-01 - 1	RT1496	2074	0.3522	-	7471B	PPB	23 Oct 2019 09:06:25	U	SMPL
K5459-01 - 1	SOIL-PILE-HARVEY-SUB	1023	0.1725	-	7471B	PPB	23 Oct 2019 09:08:44	U	SMPL
K5480-01 - 1	B-1	1888	0.3204	-	7471B	PPB	23 Oct 2019 09:11:01	U	SMPL
PB124199BL - 1	PBS	-16	-0.0051	-	7471B	PPB	23 Oct 2019 09:13:18	U	SMPL
PB124199BS - 1	LCSS	21150	3.6133	-	7471B	PPB	23 Oct 2019 09:15:37	U	SMPL
K5152-10 - 1	SU-03-102219	10361	1.7689	-	7471B	PPB	23 Oct 2019 09:17:53	U	SMPL
CCV90 - 1	CCV90	30104	5.1440	-	7471B	PPB	23 Oct 2019 09:20:09	U	SMPL
CCB90 - 1	CCB90	16	0.0003	-	7471B	PPB	23 Oct 2019 09:22:26	U	SMPL
K5485-03 - 1	TP-24-S	3524	0.6000	-	7471B	PPB	23 Oct 2019 09:24:43	U	SMPL
K5485-07 - 1	TP-24-D	9053	1.5453	-	7471B	PPB	23 Oct 2019 09:26:59	U	SMPL
K5485-11 - 1	TP-23-S	953	0.1605	-	7471B	PPB	23 Oct 2019 09:29:16	U	SMPL
K5489-01 - 1	MAIN-STOCKPILE-SLP	1028	0.1733	-	7471B	PPB	23 Oct 2019 09:31:33	U	SMPL
K5489-04 - 1	SECONDARY-PILE-SLP	3182	0.5416	-	7471B	PPB	23 Oct 2019 09:33:50	U	SMPL
K5493-01 - 1	SLOAN-ROLL-OFF-COMP	1156	0.1952	-	7471B	PPB	23 Oct 2019 09:36:07	U	SMPL
K5494-01 - 1	TP1-2FT	6562	1.1194	-	7471B	PPB	23 Oct 2019 09:38:24	U	SMPL
K5494-01DUP - 1	TP1-2FTDUP	6867	1.1715	-	7471B	PPB	23 Oct 2019 09:40:42	U	SMPL
K5494-02 - 1	K5494-01MS	29738	5.0815	-	7471B	PPB	23 Oct 2019 09:43:00	U	SMPL
K5494-03 - 1	K5494-01MSD	28348	4.8438	-	7471B	PPB	23 Oct 2019 09:45:17	U	SMPL
CCV91 - 1	CCV91	30269	5.1723	-	7471B	PPB	23 Oct 2019 09:47:33	U	SMPL
CCB91 - 1	CCB91	-22	-0.0062	-	7471B	PPB	23 Oct 2019 09:49:50	U	SMPL
K5494-04 - 1	TP1-6FT	6006	1.0244	-	7471B	PPB	23 Oct 2019 09:52:06	U	SMPL
K5494-05 - 1	TP2-2FT	5083	0.8666	-	7471B	PPB	23 Oct 2019 09:54:22	U	SMPL
K5494-06 - 1	TP2-6FT	4528	0.7717	-	7471B	PPB	23 Oct 2019 09:56:39	U	SMPL
K5494-07 - 1	TP92-2FT	6170	1.0524	-	7471B	PPB	23 Oct 2019 09:58:56	U	SMPL
K5494-08 - 1	TP3-2FT	3483	0.5930	-	7471B	PPB	23 Oct 2019 10:01:13	U	SMPL
K5494-09 - 1	TP3-5FT	2209	0.3752	-	7471B	PPB	23 Oct 2019 10:03:30	U	SMPL
K5150-01LX5 - 1		357	0.0586	-	7471B	PPB	23 Oct 2019 10:05:47	U	SMPL
K5150-01A - 1		23220	3.9672	-	7471B	PPB	23 Oct 2019 10:08:05	U	SMPL
K5494-01LX5 - 1		1143	0.1930	-	7471B	PPB	23 Oct 2019 10:10:24	U	SMPL
K5494-01A - 1		29251	4.9982	-	7471B	PPB	23 Oct 2019 10:12:41	U	SMPL
CCV92 - 1	CCV92	30078	5.1396	-	7471B	PPB	23 Oct 2019 10:14:57	U	SMPL
CCB92 - 1	CCB92	70	0.0096	-	7471B	PPB	23 Oct 2019 10:17:14	U	SMPL