


LB105877

ISM02.4

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.0060e-004

C= 8.7259e-003

Rho= 0.9995552

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	0.040	0.040	156	0.000	156				
0.05	0.050									
0.20	0.200	0.230	0.030	1105	0.0 %	1105				
2.50	2.500	2.574	0.074	12786	0.0 %	12786				
5.00	5.000	4.762	-0.238	23693	0.0 %	23693				
7.50	7.500	7.506	0.006	37372	0.0 %	37372				
10.0	10.000	10.089	0.089	50248	0.0 %	50248				

AD

15
3
15
10

AS 10.23.19

LB105877
INSTRUMENT ID : CV1

K5384-08	MH0GR6	272	0.0633 -	ISM02.4	PPB	10/23/2019 14:43 SMPL
K5384-09	MH0GR7	817	0.1726 -	ISM02.4	PPB	10/23/2019 14:45 SMPL
K5384-10	MH0GR8	4712	0.954 -	ISM02.4	PPB	10/23/2019 14:47 SMPL
CCV097	CCV097	24048	4.8328 -	ISM02.4	PPB	10/23/2019 14:49 SMPL
CCB097	CCB097	4	0.0095 -	ISM02.4	PPB	10/23/2019 14:52 SMPL
K5384-11	MH0GR9	3734	0.7578 -	ISM02.4	PPB	10/23/2019 14:54 SMPL
K5384-12	MH0GS0	809	0.171 -	ISM02.4	PPB	10/23/2019 14:56 SMPL
K5384-13	MH0GS1	260205	52.2067 -	ISM02.4	PPB	10/23/2019 14:59 SMPL
K5384-14	MH0GS2	2798	0.57 -	ISM02.4	PPB	10/23/2019 15:01 SMPL
K5384-15	MH0GS3	23316	4.686 -	ISM02.4	PPB	10/23/2019 15:03 SMPL
K5384-16	MH0GS4	1085511	HIGH -	ISM02.4	PPB	10/23/2019 15:05
K5384-17	MH0GS5	1241	0.2577 -	ISM02.4	PPB	10/23/2019 15:07 SMPL
K5384-18	MH0GS6	11626	2.3409 -	ISM02.4	PPB	10/23/2019 15:10 SMPL
K5384-19	MH0GS7	1993	0.4085 -	ISM02.4	PPB	10/23/2019 15:12 SMPL
K5384-20	MH0GS8	23881	4.7993 -	ISM02.4	PPB	10/23/2019 15:14 SMPL
K5384-21	MH0GS8D	24165	4.8563 -	ISM02.4	PPB	10/23/2019 15:17 SMPL
K5384-22	MH0GS8S	41685	8.3709 -	ISM02.4	PPB	10/23/2019 15:19 SMPL
K5381-05DLX5		24187	4.8607 -	ISM02.4	PPB	10/23/2019 15:21 SMPL
K5381-08DLX10		27923	5.6102 -	ISM02.4	PPB	10/23/2019 15:23 SMPL
K5384-13DLX10		25782	5.1807 -	ISM02.4	PPB	10/23/2019 15:26 SMPL
K5384-16DLX10		335529	67.317 -	ISM02.4	PPB	10/23/2019 15:28 SMPL
CCV098	CCV098	24005	4.8242 -	ISM02.4	PPB	10/23/2019 15:38 SMPL
CCB098	CCB098	103	0.0294 -	ISM02.4	PPB	10/23/2019 15:40 SMPL
K5384-16DL2X100		35325	7.095 -	ISM02.4	PPB	10/23/2019 15:46 SMPL
CCV099	CCV099	24305	4.8844 -	ISM02.4	PPB	10/23/2019 15:48 SMPL
CCB099	CCB099	26	0.0139 -	ISM02.4	PPB	10/23/2019 15:50 SMPL

LB105877
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	156	-		0 ISM02.4	PPB	10/23/2019 11:04	Std
0.2 S0.2	1105	-		0.2 ISM02.4	PPB	10/23/2019 11:07	Std
2.5 S2.5	12786	-		2.5 ISM02.4	PPB	10/23/2019 11:09	Std
5 S5.0	23693	-		5 ISM02.4	PPB	10/23/2019 11:11	Std
7.5 S7.5	37372	-		7.5 ISM02.4	PPB	10/23/2019 11:13	Std
10 S10.0	50248	-		10 ISM02.4	PPB	10/23/2019 11:16	Std
ICV	ICV	18841	3.7883 -	ISM02.4	PPB	10/23/2019 11:18	SMPL
ICB	ICB	90	0.0268 -	ISM02.4	PPB	10/23/2019 11:21	SMPL
CCV093	CCV093	24512	4.9259 -	ISM02.4	PPB	10/23/2019 11:23	SMPL
CCB093	CCB093	84	0.0256 -	ISM02.4	PPB	10/23/2019 11:25	SMPL
PB124220E	PBS220	152	0.0392 -	ISM02.4	PPB	10/23/2019 11:27	SMPL
K5381-01	MH0GK9	209	0.0507 -	ISM02.4	PPB	10/23/2019 11:30	SMPL
K5381-02	MH0GL0	19992	4.0192 -	ISM02.4	PPB	10/23/2019 11:32	SMPL
K5381-03	MH0GL1	567	0.1225 -	ISM02.4	PPB	10/23/2019 11:34	SMPL
K5381-04	MH0GL2	429	0.0948 -	ISM02.4	PPB	10/23/2019 11:37	SMPL
K5381-05	MH0GL3	127852	25.6563 -	ISM02.4	PPB	10/23/2019 11:39	SMPL
K5381-06	MH0GL4	249	0.0587 -	ISM02.4	PPB	10/23/2019 11:41	SMPL
K5381-07	MH0GL5	550	0.1191 -	ISM02.4	PPB	10/23/2019 11:43	SMPL
K5381-08	MH0GL6	291504	58.4854 -	ISM02.4	PPB	10/23/2019 11:46	SMPL
K5381-09	MH0GL7	975	0.2043 -	ISM02.4	PPB	10/23/2019 11:48	SMPL
K5381-10	MH0GL8	711	0.1514 -	ISM02.4	PPB	10/23/2019 11:50	SMPL
K5381-11	MH0GL9	364	0.0817 -	ISM02.4	PPB	10/23/2019 11:53	SMPL
K5381-12	MH0GM0	4735	0.9586 -	ISM02.4	PPB	10/23/2019 11:55	SMPL
K5381-13	MH0GM1	990	0.2073 -	ISM02.4	PPB	10/23/2019 11:57	SMPL
K5381-14	MH0GM2	13009	2.6184 -	ISM02.4	PPB	10/23/2019 11:59	SMPL
K5381-15	MH0GM3	2796	0.5696 -	ISM02.4	PPB	10/23/2019 12:02	SMPL
K5381-16	MH0GM4	1189	0.2472 -	ISM02.4	PPB	10/23/2019 12:04	SMPL
K5381-17	MH0GM5	5948	1.2019 -	ISM02.4	PPB	10/23/2019 12:06	SMPL
K5381-18	MH0GM6	2026	0.4151 -	ISM02.4	PPB	10/23/2019 12:09	SMPL
K5381-19	MH0GM7	16892	3.3973 -	ISM02.4	PPB	10/23/2019 12:11	SMPL
CCV094	CCV094	24610	4.9456 -	ISM02.4	PPB	10/23/2019 12:17	SMPL
CCB094	CCB094	141	0.037 -	ISM02.4	PPB	10/23/2019 12:20	SMPL
K5381-20	MH0GM8	3430	0.6968 -	ISM02.4	PPB	10/23/2019 12:22	SMPL
K5381-21	MH0GM8C	3359	0.6826 -	ISM02.4	PPB	10/23/2019 12:24	SMPL
K5381-22	MH0GM8S	16711	3.361 -	ISM02.4	PPB	10/23/2019 12:26	SMPL
PB124221E	PBS221	104	0.0296 -	ISM02.4	PPB	10/23/2019 12:29	SMPL
K5382-01	MH0GM9	852	0.1796 -	ISM02.4	PPB	10/23/2019 12:31	SMPL
K5382-02	MH0GN0	739	0.157 -	ISM02.4	PPB	10/23/2019 12:33	SMPL
K5382-03	MH0GN1	2058	0.4216 -	ISM02.4	PPB	10/23/2019 12:35	SMPL
K5382-04	MH0GN2	1406	0.2908 -	ISM02.4	PPB	10/23/2019 12:38	SMPL
K5382-05	MH0GN3	431	0.0952 -	ISM02.4	PPB	10/23/2019 12:40	SMPL
K5382-06	MH0GN4	502	0.1094 -	ISM02.4	PPB	10/23/2019 12:42	SMPL
K5382-07	MH0GN5	23513	4.7255 -	ISM02.4	PPB	10/23/2019 12:45	SMPL
K5382-08	MH0GN6	602	0.1295 -	ISM02.4	PPB	10/23/2019 12:47	SMPL
K5382-09	MH0GN7	582	0.1255 -	ISM02.4	PPB	10/23/2019 12:49	SMPL
K5382-10	MH0GN8	21759	4.3737 -	ISM02.4	PPB	10/23/2019 12:51	SMPL

LB105877
INSTRUMENT ID : CV1

K5382-11	MH0GN9	626	0.1343 -	ISM02.4	PPB	10/23/2019 12:54 SMPL
K5382-12	MH0GO0	975	0.2043 -	ISM02.4	PPB	10/23/2019 12:56 SMPL
K5382-13	MH0GO1	786	0.1664 -	ISM02.4	PPB	10/23/2019 12:58 SMPL
K5382-14	MH0GO2	3167	0.644 -	ISM02.4	PPB	10/23/2019 13:01 SMPL
K5382-15	MH0GO3	10027	2.0202 -	ISM02.4	PPB	10/23/2019 13:03 SMPL
K5382-16	MH0GO4	346	0.0781 -	ISM02.4	PPB	10/23/2019 13:05 SMPL
CCV095	CCV095	24340	4.8914 -	ISM02.4	PPB	10/23/2019 13:07 SMPL
CCB095	CCB095	-54	-0.0021 -	ISM02.4	PPB	10/23/2019 13:10 SMPL
K5382-17	MH0GO5	1950	0.3999 -	ISM02.4	PPB	10/23/2019 13:12 SMPL
K5382-18	MH0GO6	1825	0.3748 -	ISM02.4	PPB	10/23/2019 13:14 SMPL
K5382-19	MH0GO7	3220	0.6547 -	ISM02.4	PPB	10/23/2019 13:16 SMPL
K5382-20	MH0GO8	11424	2.3004 -	ISM02.4	PPB	10/23/2019 13:20 SMPL
K5382-21	MH0GO8D	11404	2.2964 -	ISM02.4	PPB	10/23/2019 13:23 SMPL
K5382-22	MH0GO8S	24440	4.9115 -	ISM02.4	PPB	10/23/2019 13:25 SMPL
PB124222E	PBS222	116	0.032 -	ISM02.4	PPB	10/23/2019 13:27 SMPL
K5383-01	MH0GO9	1101	0.2296 -	ISM02.4	PPB	10/23/2019 13:29 SMPL
K5383-02	MH0GP0	1222	0.2539 -	ISM02.4	PPB	10/23/2019 13:32 SMPL
K5383-03	MH0GP1	19679	3.9564 -	ISM02.4	PPB	10/23/2019 13:34 SMPL
K5383-04	MH0GP2	8009	1.6154 -	ISM02.4	PPB	10/23/2019 13:36 SMPL
K5383-05	MH0GP3	262	0.0613 -	ISM02.4	PPB	10/23/2019 13:39 SMPL
K5383-06	MH0GP4	357	0.0803 -	ISM02.4	PPB	10/23/2019 13:41 SMPL
K5383-07	MH0GP5	360	0.0809 -	ISM02.4	PPB	10/23/2019 13:43 SMPL
K5383-08	MH0GP6	770	0.1632 -	ISM02.4	PPB	10/23/2019 13:45 SMPL
K5383-09	MH0GP7	1760	0.3618 -	ISM02.4	PPB	10/23/2019 13:48 SMPL
K5383-10	MH0GP8	602	0.1295 -	ISM02.4	PPB	10/23/2019 13:50 SMPL
K5383-11	MH0GP9	400	0.089 -	ISM02.4	PPB	10/23/2019 13:52 SMPL
K5383-12	MH0GQ0	1247	0.2589 -	ISM02.4	PPB	10/23/2019 13:55 SMPL
K5383-13	MH0GQ1	626	0.1343 -	ISM02.4	PPB	10/23/2019 13:57 SMPL
CCV096	CCV096	23945	4.8122 -	ISM02.4	PPB	10/23/2019 13:59 SMPL
CCB096	CCB096	94	0.0276 -	ISM02.4	PPB	10/23/2019 14:01 SMPL
K5383-14	MH0GQ2	394	0.0878 -	ISM02.4	PPB	10/23/2019 14:04 SMPL
K5383-15	MH0GQ3	678	0.1447 -	ISM02.4	PPB	10/23/2019 14:06 SMPL
K5383-16	MH0GQ4	386	0.0862 -	ISM02.4	PPB	10/23/2019 14:08 SMPL
K5383-17	MH0GQ5	677	0.1445 -	ISM02.4	PPB	10/23/2019 14:11 SMPL
K5383-18	MH0GQ6	727	0.1546 -	ISM02.4	PPB	10/23/2019 14:13 SMPL
K5383-19	MH0GQ7	351	0.0791 -	ISM02.4	PPB	10/23/2019 14:15 SMPL
K5383-20	MH0GQ8	1673	0.3443 -	ISM02.4	PPB	10/23/2019 14:17 SMPL
K5383-21	MH0GQ8D	1649	0.3395 -	ISM02.4	PPB	10/23/2019 14:20 SMPL
K5383-22	MH0GQ8S	16027	3.2238 -	ISM02.4	PPB	10/23/2019 14:22 SMPL
PB124223E	PBS223	-37	0.0013 -	ISM02.4	PPB	10/23/2019 14:24 SMPL
K5384-01	MH0GQ9	1682	0.3461 -	ISM02.4	PPB	10/23/2019 14:26 SMPL
K5384-02	MH0GR0	2410	0.4922 -	ISM02.4	PPB	10/23/2019 14:29 SMPL
K5384-03	MH0GR1	387	0.0864 -	ISM02.4	PPB	10/23/2019 14:31 SMPL
K5384-04	MH0GR2	563	0.1217 -	ISM02.4	PPB	10/23/2019 14:33 SMPL
K5384-05	MH0GR3	13700	2.757 -	ISM02.4	PPB	10/23/2019 14:36 SMPL
K5384-06	MH0GR4	257	0.0603 -	ISM02.4	PPB	10/23/2019 14:38 SMPL
K5384-07	MH0GR5	325	0.0739 -	ISM02.4	PPB	10/23/2019 14:40 SMPL