



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : jignesh

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105884

Slope : 98.3

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/23/2019	10:10
2	CAL2	1	Water	NA	NA	20.1	7.01	10/23/2019	10:11
3	CAL3	1	Water	NA	NA	20.1	10.03	10/23/2019	10:12
4	ICV	1	Water	NA	NA	20.1	7.00	10/23/2019	10:13
5	CCV1	1	Water	NA	NA	20.2	2.02	10/23/2019	10:15
6	K5496-01	1	Water	NA	NA	17.9	6.76	10/23/2019	10:16
7	K5496-01DUP	1	Water	NA	NA	17.8	6.77	10/23/2019	10:17
8	CCV2	1	Water	NA	NA	20.1	12.02	10/23/2019	10:20

LB 105884

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5496 WorkList ID : 132480 Department : Wet-Chemistry Date : 10-23-2019 13:16:06

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2019	Water	K5496-01	pH	Cool 4 deg C	APT101	E11	WGGM-19-TSF-WW-037	10/22/2019	SM 4500-H B

Date/Time 10/23/19 10:00
Received by: JS
Relinquished by: JS

Date/Time 10/23/19 11:00
Received by: JS
Relinquished by: JS