



Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105891

Slope : 98.5

pH Meter ID : WC PH MBTER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	3.99	10/23/2019	16:25
2	CAL2	1	Water	NA	NA	20.1	7.01	10/23/2019	16:26
3	CAL3	1	Water	NA	NA	20.2	10.03	10/23/2019	16:27
4	ICV	1	Water	NA	NA	20.1	7.00	10/23/2019	16:30
5	CCV1	1	Water	NA	NA	20.1	2.01	10/23/2019	16:31
6	K5514-01	1	Water	NA	NA	19.0	8.69	10/23/2019	16:33
7	K5514-01DUP	1	Water	NA	NA	19.1	8.70	10/23/2019	16:35
8	CCV2	1	Water	NA	NA	20.1	12.02	10/23/2019	16:36

28105891

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIONITY K5514 W WorkList ID : 132499 Department : Wet-Chemistry Date : 10-23-2019 15:14:35

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/23/2019	Water	K5514-01	Corrosivity	Cool 4 deg C	PSEG01	N51	ROLL-OFF-WATER	10/23/2019	9040C

Date/Time 10/23/19 16:20
Received by: Jn
Relinquished by: L.C

Date/Time 10/23/19 17:10
Received by: L.C
Relinquished by: Jn