


LB105895

7470A

INSTRUMENT ID : CV1

Linear



A= 0.0000e+000

B= 2.5927e-004

C= -4.5395e-002

Rho= 0.9999201

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.032	-0.032	50	0.000	50				
0.2	0.200	0.180	-0.020	871	0.0 %	871				
2.5	2.500	2.514	0.014	9871	0.0 %	9871				
5.0	5.000	5.098	0.098	19838	0.0 %	19838				
7.5	7.500	7.474	-0.026	29003	0.0 %	29003				
10.0	10.000	9.966	-0.034	38614	0.0 %	38614				

LB105895 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Oct 2019 07:57:27

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	50	-	0.0000	7470A	PPB	24 Oct 2019 08:01:41	S	Std
0.2 - 1	50.2	871	-	0.2000	7470A	PPB	24 Oct 2019 08:03:50	S	Std
2.5 - 1	52.5	9871	-	2.5000	7470A	PPB	24 Oct 2019 08:05:59	S	Std
5.0 - 1	55.0	19838	-	5.0000	7470A	PPB	24 Oct 2019 08:08:07	S	Std
7.5 - 1	57.5	29003	-	7.5000	7470A	PPB	24 Oct 2019 08:10:16	S	Std
10.0 - 1	510.0	38614	-	10.0000	7470A	PPB	24 Oct 2019 08:12:23	S	Std
ICV70 - 1	ICV70	15962	4.0931	-	7470A	PPB	24 Oct 2019 08:14:59	U	SMPL
ICB70 - 1	ICB70	23	-0.0394	-	7470A	PPB	24 Oct 2019 08:17:05	U	SMPL
CCV01 - 1	CCV01	20041	5.1506	-	7470A	PPB	24 Oct 2019 08:19:12	U	SMPL
CCB01 - 1	CCB01	21	-0.0400	-	7470A	PPB	24 Oct 2019 08:21:18	U	SMPL
CRA - 1	CRA	884	0.1838	-	7470A	PPB	24 Oct 2019 08:23:25	U	SMPL
HighStd - 1	HighStd	38720	9.9935	-	7470A	PPB	24 Oct 2019 08:25:32	U	SMPL
ChkStd - 1	ChkStd	26939	6.9390	-	7470A	PPB	24 Oct 2019 08:27:39	U	SMPL
PB124254BL - 1	PBW	6	-0.0438	-	7470A	PPB	24 Oct 2019 08:29:46	U	SMPL
PB124254BS - 1	LCSW	16355	4.1949	-	7470A	PPB	24 Oct 2019 08:31:53	U	SMPL
K5139-03 - 1	BU-03-102119	41	-0.0348	-	7470A	PPB	24 Oct 2019 08:34:00	U	SMPL
K5139-03DUP - 1	BU-03-102119DUP	57	-0.0306	-	7470A	PPB	24 Oct 2019 08:36:07	U	SMPL
K5139-03MS - 1	BU-03-102119MS	16346	4.1926	-	7470A	PPB	24 Oct 2019 08:42:47	U	SMPL
K5139-03MSD - 1	BU-03-102119MSD	16436	4.2159	-	7470A	PPB	24 Oct 2019 08:44:55	U	SMPL
K5152-13 - 1	SU-03-102219	67	-0.0280	-	7470A	PPB	24 Oct 2019 08:47:02	U	SMPL
CCV02 - 1	CCV02	20052	5.1535	-	7470A	PPB	24 Oct 2019 08:49:08	U	SMPL
CCB02 - 1	CCB02	102	-0.0189	-	7470A	PPB	24 Oct 2019 08:51:15	U	SMPL
K5485-04 - 1	TP-24B-S	64	-0.0288	-	7470A	PPB	24 Oct 2019 08:53:22	U	SMPL
K5485-08 - 1	TP-24B-D	53	-0.0317	-	7470A	PPB	24 Oct 2019 08:55:28	U	SMPL
K5485-12 - 1	TP-23-S	54	-0.0314	-	7470A	PPB	24 Oct 2019 08:59:39	U	SMPL
K5489-03 - 1	MAIN-STOCKPILE-SLP	47	-0.0332	-	7470A	PPB	24 Oct 2019 09:01:46	U	SMPL
K5489-06 - 1	SECONDARY-PILE-SLP	42	-0.0345	-	7470A	PPB	24 Oct 2019 09:03:53	U	SMPL
K5495-01 - 1	WASTE	29	-0.0379	-	7470A	PPB	24 Oct 2019 09:06:01	U	SMPL
PB124254TB - 1	PB124185TB	27	-0.0384	-	7470A	PPB	24 Oct 2019 09:08:09	U	SMPL
K5139-03LX5 - 1		13	-0.0420	-	7470A	PPB	24 Oct 2019 09:10:17	U	SMPL
K5139-03A - 1		16748	4.2968	-	7470A	PPB	24 Oct 2019 09:12:26	U	SMPL
CCV03 - 1	CCV03	20230	5.1996	-	7470A	PPB	24 Oct 2019 09:14:33	U	SMPL
CCB03 - 1	CCB03	38	-0.0355	-	7470A	PPB	24 Oct 2019 09:16:40	U	SMPL