


LB105901

ISM02.4

INSTRUMENT ID : CU1

Linear



A= 0.0000e+000

B= 2.0687e-004

C= -3.4029e-002

Rho= 0.9999408

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.023	-0.023	53	0.000	53				
0.05	0.050									
0.20	0.200	0.190	-0.010	1081	0.0 %	1081				
2.50	2.500	2.577	0.077	12624	0.0 %	12624				
5.00	5.000	4.946	-0.054	24075	0.0 %	24075				
7.50	7.500	7.512	0.012	36476	0.0 %	36476				
10.0	10.000	9.998	-0.002	48494	0.0 %	48494				

%D

—

-5

3

-1

1

0

RS 10-24-19

LB105901
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	53	-		0 ISM02.4	PPB	10/24/2019 9:37	Std
0.2 S0.2	1081	-		0.2 ISM02.4	PPB	10/24/2019 9:39	Std
2.5 S2.5	12624	-		2.5 ISM02.4	PPB	10/24/2019 9:41	Std
5 S5.0	24075	-		5 ISM02.4	PPB	10/24/2019 9:43	Std
7.5 S7.5	36476	-		7.5 ISM02.4	PPB	10/24/2019 9:46	Std
10 S10.0	48494	-		10 ISM02.4	PPB	10/24/2019 9:48	Std
ICV	ICV	18788	3.8526 -	ISM02.4	PPB	10/24/2019 9:51	SMPL
ICB	ICB	11	-0.0318 -	ISM02.4	PPB	10/24/2019 9:54	SMPL
CCV004	CCV004	24071	4.9455 -	ISM02.4	PPB	10/24/2019 9:56	SMPL
CCB004	CCB004	6	-0.0328 -	ISM02.4	PPB	10/24/2019 9:58	SMPL
PB124263E	PBS263	3	-0.0334 -	ISM02.4	PPB	10/24/2019 10:01	SMPL
K5385-01	MH0GS9	943	0.161 -	ISM02.4	PPB	10/24/2019 10:03	SMPL
K5385-02	MH0GT0	3686	0.7285 -	ISM02.4	PPB	10/24/2019 10:05	SMPL
K5385-03	MH0GT1	3864	0.7653 -	ISM02.4	PPB	10/24/2019 10:07	SMPL
K5385-04	MH0GT2	945	0.1615 -	ISM02.4	PPB	10/24/2019 10:10	SMPL
K5385-05	MH0GT3	-24	-0.039 -	ISM02.4	PPB	10/24/2019 10:12	SMPL
K5385-06	MH0GT4	382	0.045 -	ISM02.4	PPB	10/24/2019 10:14	SMPL
K5385-07	MH0GT5	111	-0.0111 -	ISM02.4	PPB	10/24/2019 10:16	SMPL
K5385-08	MH0GT6	456	0.0603 -	ISM02.4	PPB	10/24/2019 10:19	SMPL
K5385-09	MH0GT7	1217	0.2177 -	ISM02.4	PPB	10/24/2019 10:21	SMPL
K5385-10	MH0GT8	1495	0.2752 -	ISM02.4	PPB	10/24/2019 10:23	SMPL
K5385-11	MH0GT9	4066	0.8071 -	ISM02.4	PPB	10/24/2019 10:26	SMPL
K5385-12	MH0GU0	1236	0.2217 -	ISM02.4	PPB	10/24/2019 10:28	SMPL
K5385-13	MH0GU1	2166	0.4141 -	ISM02.4	PPB	10/24/2019 10:30	SMPL
K5385-14	MH0GU3	2460	0.4749 -	ISM02.4	PPB	10/24/2019 10:32	SMPL
K5385-15	MH0GU4	6640	1.3396 -	ISM02.4	PPB	10/24/2019 10:35	SMPL
K5385-16	MH0GU5	1810	0.3404 -	ISM02.4	PPB	10/24/2019 10:37	SMPL
K5385-17	MH0GU6	963	0.1652 -	ISM02.4	PPB	10/24/2019 10:39	SMPL
K5385-18	MH0GU7	1352	0.2457 -	ISM02.4	PPB	10/24/2019 10:42	SMPL
K5385-19	MH0GU8	164	-0.0001 -	ISM02.4	PPB	10/24/2019 10:44	SMPL
CCV005	CCV005	24442	5.0223 -	ISM02.4	PPB	10/24/2019 10:46	SMPL
CCB005	CCB005	8	-0.0324 -	ISM02.4	PPB	10/24/2019 10:48	SMPL
K5385-20	MH0GU9	5200	1.0417 -	ISM02.4	PPB	10/24/2019 10:51	SMPL
K5385-21	MH0GU9D	5079	1.0167 -	ISM02.4	PPB	10/24/2019 10:53	SMPL
K5385-22	MH0GU9S	17475	3.581 -	ISM02.4	PPB	10/24/2019 10:55	SMPL
PB124264E	PBS264	-16	-0.0373 -	ISM02.4	PPB	10/24/2019 10:57	SMPL
K5386-01	MH0GV0	232	0.014 -	ISM02.4	PPB	10/24/2019 11:00	SMPL
K5386-02	MH0GV1	4838	0.9668 -	ISM02.4	PPB	10/24/2019 11:02	SMPL
K5386-03	MH0GV2	112	-0.0109 -	ISM02.4	PPB	10/24/2019 11:06	SMPL
K5386-04	MH0GV3	107	-0.0119 -	ISM02.4	PPB	10/24/2019 11:08	SMPL
K5386-05	MH0GV4	3830	0.7583 -	ISM02.4	PPB	10/24/2019 11:10	SMPL
K5386-06	MH0GV5	32768	6.7447 -	ISM02.4	PPB	10/24/2019 11:13	SMPL
K5386-07	MH0GV6	9232	1.8758 -	ISM02.4	PPB	10/24/2019 11:15	SMPL
K5386-08	MH0GV7	514	0.0723 -	ISM02.4	PPB	10/24/2019 11:17	SMPL
K5386-09	MH0GV8	774	0.1261 -	ISM02.4	PPB	10/24/2019 11:20	SMPL
K5386-10	MH0GV9	832	0.1381 -	ISM02.4	PPB	10/24/2019 11:22	SMPL

LB105901
INSTRUMENT ID : CV1

K5386-11	MH0GW0	10424	2.1224 -	ISM02.4	PPB	10/24/2019 11:24 SMPL
K5386-12	MH0GW1	435	0.056 -	ISM02.4	PPB	10/24/2019 11:26 SMPL
K5386-13	MH0GW2	336	0.0355 -	ISM02.4	PPB	10/24/2019 11:29 SMPL
K5386-14	MH0GW3	10511	2.1404 -	ISM02.4	PPB	10/24/2019 11:31 SMPL
K5386-15	MH0GW4	6165	1.2413 -	ISM02.4	PPB	10/24/2019 11:33 SMPL
K5386-16	MH0GW5	3829	0.7581 -	ISM02.4	PPB	10/24/2019 11:36 SMPL
CCV006	CCV006	24340	5.0012 -	ISM02.4	PPB	10/24/2019 11:38 SMPL
CCB006	CCB006	22	-0.0295 -	ISM02.4	PPB	10/24/2019 11:40 SMPL
K5386-17	MH0GW6	215	0.0104 -	ISM02.4	PPB	10/24/2019 11:42 SMPL
K5386-18	MH0GW7	3215	0.6311 -	ISM02.4	PPB	10/24/2019 11:45 SMPL
K5386-19	MH0GW8	12131	2.4755 -	ISM02.4	PPB	10/24/2019 11:47 SMPL
K5386-20	MH0GW9	632	0.0967 -	ISM02.4	PPB	10/24/2019 11:49 SMPL
K5386-21	MH0GW9C	630	0.0963 -	ISM02.4	PPB	10/24/2019 11:51 SMPL
K5386-22	MH0GW9S	14160	2.8952 -	ISM02.4	PPB	10/24/2019 11:54 SMPL
PB124265E	PBS265	22	-0.0295 -	ISM02.4	PPB	10/24/2019 11:56 SMPL
K5387-01	MH0GX0	11846	2.4165 -	ISM02.4	PPB	10/24/2019 11:58 SMPL
K5387-02	MH0GX1	73	-0.0189 -	ISM02.4	PPB	10/24/2019 12:01 SMPL
K5387-03	MH0GX2	15750	3.2242 -	ISM02.4	PPB	10/24/2019 12:04 SMPL
K5387-04	MH0GX3	7838	1.5874 -	ISM02.4	PPB	10/24/2019 12:07 SMPL
K5387-05	MH0GX4	848	0.1414 -	ISM02.4	PPB	10/24/2019 12:09 SMPL
K5387-06	MH0GX5	1148	0.2035 -	ISM02.4	PPB	10/24/2019 12:11 SMPL
K5387-07	MH0GX6	1557	0.2881 -	ISM02.4	PPB	10/24/2019 12:14 SMPL
K5387-08	MH0GX7	71	-0.0193 -	ISM02.4	PPB	10/24/2019 12:16 SMPL
K5387-09	MH0GX8	1686	0.3148 -	ISM02.4	PPB	10/24/2019 12:18 SMPL
K5387-10	MH0GX9	852	0.1422 -	ISM02.4	PPB	10/24/2019 12:20 SMPL
K5387-11	MH0GY0	85	-0.0164 -	ISM02.4	PPB	10/24/2019 12:23 SMPL
K5387-12	MH0GY1	16051	3.2864 -	ISM02.4	PPB	10/24/2019 12:25 SMPL
K5387-13	MH0GY2	277	0.0233 -	ISM02.4	PPB	10/24/2019 12:27 SMPL
CCV007	CCV007	23247	4.7751 -	ISM02.4	PPB	10/24/2019 12:29 SMPL
CCB007	CCB007	297	0.0274 -	ISM02.4	PPB	10/24/2019 12:32 SMPL
K5387-14	MH0GY3	1937	0.3667 -	ISM02.4	PPB	10/24/2019 12:34 SMPL
K5387-15	MH0GY4	278	0.0235 -	ISM02.4	PPB	10/24/2019 12:36 SMPL
K5387-16	MH0GY5	299	0.0278 -	ISM02.4	PPB	10/24/2019 12:39 SMPL
K5387-17	MH0GY6	2721	0.5289 -	ISM02.4	PPB	10/24/2019 12:41 SMPL
K5387-18	MH0GY7	1824	0.3433 -	ISM02.4	PPB	10/24/2019 12:43 SMPL
K5387-19	MH0GY8	68596	14.1564 -	ISM02.4	PPB	10/24/2019 12:45 SMPL
K5387-20	MH0GY9	11702	2.3868 -	ISM02.4	PPB	10/24/2019 12:48 SMPL
K5387-21	MH0GY9D	11619	2.3696 -	ISM02.4	PPB	10/24/2019 12:50 SMPL
K5387-22	MH0GY9S	25670	5.2763 -	ISM02.4	PPB	10/24/2019 12:52 SMPL
PB124266E	PBS266	304	0.0289 -	ISM02.4	PPB	10/24/2019 12:55 SMPL
K5388-01	MH0GZ0	479	0.0651 -	ISM02.4	PPB	10/24/2019 12:57 SMPL
K5388-02	MH0GZ1	1464	0.2688 -	ISM02.4	PPB	10/24/2019 12:59 SMPL
K5388-03	MH0GZ2	1224	0.2192 -	ISM02.4	PPB	10/24/2019 13:01 SMPL
K5388-04	MH0GZ3	2225	0.4263 -	ISM02.4	PPB	10/24/2019 13:04 SMPL
K5388-05	MH0GZ4	4604	0.9184 -	ISM02.4	PPB	10/24/2019 13:06 SMPL
K5388-06	MH0GZ5	2256	0.4327 -	ISM02.4	PPB	10/24/2019 13:08 SMPL
K5388-07	MH0GZ6	12878	2.63 -	ISM02.4	PPB	10/24/2019 13:11 SMPL

LB105901
INSTRUMENT ID : CV1

K5388-08	MHOGZ7	1128	0.1993 -	ISM02.4	PPB	10/24/2019 13:13 SMPL
K5388-09	MHOGZ8	4621	0.9219 -	ISM02.4	PPB	10/24/2019 13:15 SMPL
K5388-10	MHOGZ9	673	0.1052 -	ISM02.4	PPB	10/24/2019 13:17 SMPL
CCV008	CCV008	24064	4.9441 -	ISM02.4	PPB	10/24/2019 13:20 SMPL
CCB008	CCB008	14	-0.0311 -	ISM02.4	PPB	10/24/2019 13:22 SMPL
K5388-11	MH0HA0	525	0.0746 -	ISM02.4	PPB	10/24/2019 13:24 SMPL
K5388-12	MH0HA1	854	0.1426 -	ISM02.4	PPB	10/24/2019 13:27 SMPL
K5388-13	MH0HA2	4259	0.847 -	ISM02.4	PPB	10/24/2019 13:29 SMPL
K5388-14	MH0HA3	436	0.0562 -	ISM02.4	PPB	10/24/2019 13:31 SMPL
K5388-15	MH0HA4	1579	0.2926 -	ISM02.4	PPB	10/24/2019 13:33 SMPL
K5388-16	MH0HA5	4361	0.8681 -	ISM02.4	PPB	10/24/2019 13:36 SMPL
K5388-17	MH0HA5D	4372	0.8704 -	ISM02.4	PPB	10/24/2019 13:38 SMPL
K5388-18	MH0HA5S	17498	3.5858 -	ISM02.4	PPB	10/24/2019 13:40 SMPL
K5387-19DLX3		21960	4.5088 -	ISM02.4	PPB	10/24/2019 13:43 SMPL
CCV009	CCV009	23405	4.8077 -	ISM02.4	PPB	10/24/2019 13:45 SMPL
CCB009	CCB009	177	0.0026 -	ISM02.4	PPB	10/24/2019 13:47 SMPL