



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: AMANDEEP

Run Number: LB105905

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/24/2019	10:55
2	K5505-01		1	9	100.00	>212.0	>212.0	10/24/2019	11:20
3	K5505-01DUP		1	9	100.00	>212.0	>212.0	10/24/2019	11:45
4	K5514-01		1	7	100.00	>212.0	>212.0	10/24/2019	12:10

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp102419

WorkList ID : 132545

Department : Wet-Chemistry

Date : 10-24-2019 09:05:55

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
11/05/2019	Water	K5505-01	Flash Point	Cool 4 deg C	DIST01	E11	55-ST-22	10/23/2019	1010A
10/30/2019	Water	K5514-01	Flash Point	Cool 4 deg C	PSEG01	N51	ROLL-OFF-WATER	10/23/2019	1010A

Date/Time 10/24/19 10:30
 Received by: KS
 Relinquished by: JP

Date/Time 10/24/19 12:15
 Received by: JP
 Relinquished by: KS