

TOTAL SOLIDS - SM2540B

SUPERVISOR: apatel

ANALYST: JIGNESH

Date: 10/24/2019

OVEN IN TEMP1: 98 °C

OVEN OUT TEMP1: 104 °C

DATE & TIME IN1: 10/23/2019 16:00

DATE & TIME OUT1: 10/24/2019 07:30

OVEN IN TEMP2: 104 °C

OVEN OUT TEMP2: 105 °C

DATE & TIME IN2: 10/24/2019 07:35

DATE & TIME OUT2: 10/24/2019 12:05

Run Number: LB105906

BalanceID: WC SC-3

OvenID: WC OVEN-1

Dish #	Lab ID	Client ID	Empty Dish Weight (g)	Final Empty Dish Weight (g)	Sample Vol (ml)	1st Dish+Sample weight after Drying @103-@105°C (g)	2nd Dish+Sample weight after Drying @103-@105°C (g)	Final Dish+Sample weight after Drying @103-@105°C (g)	Weight (g)	Result (mg/L)
1	LB105906BL	LB105906BL	86.5791	86.5791	100	86.5791	86.5791	86.5791	0	0
2	K5505-01	55-ST-22	79.5122	79.5122	100	79.6359	79.6359	79.6359	0.1237	1237
3	K5505-01DUP	55-ST-22DUP	82.4679	82.4679	100	82.5883	82.5883	82.5883	0.1204	1204

A = Sample Volume (ml)

B = Final Empty Dish Weight (g)

C = Final Dish+Sample weight after Drying @103-@105°C (g)

$$\text{Result mg/L} = ((C - B) / A) * 1000 * 1000$$

SP ~~105882~~ 105906

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TS K5505 WorkList ID : 132475 Department : Wet-Chemistry Date : 10-23-2019 12:32:03

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/30/2019	Water	K5505-01 -D	TS	Cool 4 deg C	DIST01	E11	55-ST-22	10/23/2019	SM2540 B

Date/Time 10/23/19 14:40
Received by: Jn
Relinquished by: ca

Date/Time 10/23/19 16:00
Received by: ca
Relinquished by: Jn