

LB105908


ISM02.4

INSTRUMENT ID : CV1

Linear

 μ Abs.:

40644



A= 0.0000e+000

B= 2.4676e-004

C= -9.1866e-002

Rho= 0.9998544

Accept=Accepted

slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.048	-0.048	176	0.000	176				
0.05	0.050	0.059	0.009	612	0.0 %	612				
0.20	0.200	0.241	0.041	1347	0.0 %	1347				
2.50	2.500	2.418	-0.082	10173	0.0 %	10173				
5.00	5.000	5.112	0.112	21087	0.0 %	21087				
7.50	7.500	7.531	0.031	30894	0.0 %	30894				
10.0	10.000	9.937	-0.063	40644	0.0 %	40644				

10
18
21
-3
2
0
-1

RS 10.24.19

LB105908
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	
0 S0.0	176	-	0	ISM02.4	PPB	10/24/2019 14:11	Std	
0.05 S0.05	612	-	0.05	ISM02.4	PPB	10/24/2019 14:13	Std	
0.2 S0.2	1347	-	0.2	ISM02.4	PPB	10/24/2019 14:16	Std	
2.5 S2.5	10173	-	2.5	ISM02.4	PPB	10/24/2019 14:18	Std	
5 S5.0	21087	-	5	ISM02.4	PPB	10/24/2019 14:20	Std	
7.5 S7.5	30894	-	7.5	ISM02.4	PPB	10/24/2019 14:22	Std	
10 S10.0	40644	-	10	ISM02.4	PPB	10/24/2019 14:25	Std	
ICV	ICV	16268	3.9224	-	ISM02.4	PPB	10/24/2019 14:27	SMPL
ICB	ICB	136	-0.0583	-	ISM02.4	PPB	10/24/2019 14:30	SMPL
CCV010	CCV010	21356	5.1779	-	ISM02.4	PPB	10/24/2019 14:32	SMPL
CCB010	CCB010	113	-0.064	-	ISM02.4	PPB	10/24/2019 14:34	SMPL
PB124282E	PBW282	163	-0.0516	-	ISM02.4	PPB	10/24/2019 14:36	SMPL
K5295-01	MBF537	118	-0.0627	-	ISM02.4	PPB	10/24/2019 14:39	SMPL
K5295-02	MBF537D	102	-0.0667	-	ISM02.4	PPB	10/24/2019 14:41	SMPL
K5295-03	MBF537S	3518	0.7762	-	ISM02.4	PPB	10/24/2019 14:43	SMPL
K5295-04	MBF539	45	-0.0808	-	ISM02.4	PPB	10/24/2019 14:45	SMPL
K5295-05	MBF541	79	-0.0724	-	ISM02.4	PPB	10/24/2019 14:48	SMPL
K5295-06	MBF542	68	-0.0751	-	ISM02.4	PPB	10/24/2019 14:50	SMPL
K5295-07	MBF4K3	87	-0.0704	-	ISM02.4	PPB	10/24/2019 14:52	SMPL
K5295-08	MBF559	77	-0.0729	-	ISM02.4	PPB	10/24/2019 14:55	SMPL
K5295-09	MBF543	99	-0.0674	-	ISM02.4	PPB	10/24/2019 14:57	SMPL
K5295-10	MBF544	101	-0.0669	-	ISM02.4	PPB	10/24/2019 14:59	SMPL
K5295-11	MBF545	112	-0.0642	-	ISM02.4	PPB	10/24/2019 15:01	SMPL
K5295-12	MBF535	152	-0.0544	-	ISM02.4	PPB	10/24/2019 15:04	SMPL
K5295-13	MBF536	147	-0.0556	-	ISM02.4	PPB	10/24/2019 15:06	SMPL
K5295-14	MBF538	203	-0.0418	-	ISM02.4	PPB	10/24/2019 15:08	SMPL
K5295-15	MBF540	128	-0.0603	-	ISM02.4	PPB	10/24/2019 15:10	SMPL
PB124283E	PBW283	127	-0.0605	-	ISM02.4	PPB	10/24/2019 15:13	SMPL
K5320-01	ME0AA4	99	-0.0674	-	ISM02.4	PPB	10/24/2019 15:15	SMPL
K5320-02	ME0AA7	-26	-0.0983	-	ISM02.4	PPB	10/24/2019 15:17	SMPL
K5320-03	ME0AB3	142	-0.0568	-	ISM02.4	PPB	10/24/2019 15:20	SMPL
CCV011	CCV011	21071	5.1076	-	ISM02.4	PPB	10/24/2019 15:22	SMPL
CCB011	CCB011	129	-0.06	-	ISM02.4	PPB	10/24/2019 15:24	SMPL
K5320-04	ME0AB4	129	-0.06	-	ISM02.4	PPB	10/24/2019 15:26	SMPL
K5320-05	ME0AA2	119	-0.0625	-	ISM02.4	PPB	10/24/2019 15:29	SMPL
K5320-06	ME0AA3	109	-0.065	-	ISM02.4	PPB	10/24/2019 15:31	SMPL
K5320-07	ME0AA9	168	-0.0504	-	ISM02.4	PPB	10/24/2019 15:33	SMPL
K5320-08	ME0AB2	232	-0.0346	-	ISM02.4	PPB	10/24/2019 15:35	SMPL
K5320-09	ME0AB6	251	-0.0299	-	ISM02.4	PPB	10/24/2019 15:38	SMPL
K5320-10	ME0AB7	293	-0.0196	-	ISM02.4	PPB	10/24/2019 15:40	SMPL
K5320-11	ME0AA1	106	-0.0657	-	ISM02.4	PPB	10/24/2019 15:42	SMPL
K5320-12	ME0AA5	224	-0.0366	-	ISM02.4	PPB	10/24/2019 15:45	SMPL
K5320-13	ME0AA8	208	-0.0405	-	ISM02.4	PPB	10/24/2019 15:47	SMPL
K5320-14	ME0AB5	164	-0.0514	-	ISM02.4	PPB	10/24/2019 15:49	SMPL
K5320-15	ME0AA6	155	-0.0536	-	ISM02.4	PPB	10/24/2019 15:51	SMPL
K5320-16	ME0AA6D	177	-0.0482	-	ISM02.4	PPB	10/24/2019 15:54	SMPL

LB105908

INSTRUMENT ID : CV1

K5320-17	ME0AA6S	5113	1.1698 -	ISM02.4	PPB	10/24/2019 15:56 SMPL
K5320-18	ME0AB0	193	-0.0442 -	ISM02.4	PPB	10/24/2019 15:58 SMPL
K5320-19	ME0AB1	232	-0.0346 -	ISM02.4	PPB	10/24/2019 16:01 SMPL
CCV012	CCV012	21154	5.1281 -	ISM02.4	PPB	10/24/2019 16:03 SMPL
CCB012	CCB012	105	-0.066 -	ISM02.4	PPB	10/24/2019 16:05 SMPL