


LB105913

ISM02.4

INSTRUMENT ID: CV1



A= 0.0000e+000

B= 2.4676e-004

C= -9.1866e-002

Rho= 0.9998544

Accept = Accepted

Slope
y-intercept

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.048	-0.048	176	0.000	176					18
0.05	0.050	0.059	0.009	612	0.0 %	612					21
0.20	0.200	0.241	0.041	1347	0.0 %	1347					3
2.50	2.500	2.418	-0.082	10173	0.0 %	10173					2
5.00	5.000	5.112	0.112	21087	0.0 %	21087					0
7.50	7.500	7.531	0.031	30894	0.0 %	30894					0
10.0	10.000	9.937	-0.063	40644	0.0 %	40644					-1

12 1024.19

LB105908
INSTRUMENT ID : CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0.0	176	-		0 ISM02.4	PPB	10/24/2019 14:11	Std
0.05 S0.05	612	-		0.05 ISM02.4	PPB	10/24/2019 14:13	Std
0.2 S0.2	1347	-		0.2 ISM02.4	PPB	10/24/2019 14:16	Std
2.5 S2.5	10173	-		2.5 ISM02.4	PPB	10/24/2019 14:18	Std
5 S5.0	21087	-		5 ISM02.4	PPB	10/24/2019 14:20	Std
7.5 S7.5	30894	-		7.5 ISM02.4	PPB	10/24/2019 14:22	Std
10 S10.0	40644	-		10 ISM02.4	PPB	10/24/2019 14:25	Std
ICV	ICV	16268	3.9224 -	ISM02.4	PPB	10/24/2019 14:27	SMPL
ICB	ICB	136	-0.0583 -	ISM02.4	PPB	10/24/2019 14:30	SMPL
CCV010	CCV010	21356	5.1779 -	ISM02.4	PPB	10/24/2019 14:32	SMPL
CCB010	CCB010	113	-0.064 -	ISM02.4	PPB	10/24/2019 14:34	SMPL
PB124282E	PBW282	163	-0.0516 -	ISM02.4	PPB	10/24/2019 14:36	SMPL
K5295-01	MBF537	118	-0.0627 -	ISM02.4	PPB	10/24/2019 14:39	SMPL
K5295-02	MBF537D	102	-0.0667 -	ISM02.4	PPB	10/24/2019 14:41	SMPL
K5295-03	MBF537S	3518	0.7762 -	ISM02.4	PPB	10/24/2019 14:43	SMPL
K5295-04	MBF539	45	-0.0808 -	ISM02.4	PPB	10/24/2019 14:45	SMPL
K5295-05	MBF541	79	-0.0724 -	ISM02.4	PPB	10/24/2019 14:48	SMPL
K5295-06	MBF542	68	-0.0751 -	ISM02.4	PPB	10/24/2019 14:50	SMPL
K5295-07	MBF4K3	87	-0.0704 -	ISM02.4	PPB	10/24/2019 14:52	SMPL
K5295-08	MBF559	77	-0.0729 -	ISM02.4	PPB	10/24/2019 14:55	SMPL
K5295-09	MBF543	99	-0.0674 -	ISM02.4	PPB	10/24/2019 14:57	SMPL
K5295-10	MBF544	101	-0.0669 -	ISM02.4	PPB	10/24/2019 14:59	SMPL
K5295-11	MBF545	112	-0.0642 -	ISM02.4	PPB	10/24/2019 15:01	SMPL
K5295-12	MBF535	152	-0.0544 -	ISM02.4	PPB	10/24/2019 15:04	SMPL
K5295-13	MBF536	147	-0.0556 -	ISM02.4	PPB	10/24/2019 15:06	SMPL
K5295-14	MBF538	203	-0.0418 -	ISM02.4	PPB	10/24/2019 15:08	SMPL
K5295-15	MBF540	128	-0.0603 -	ISM02.4	PPB	10/24/2019 15:10	SMPL
PB124283E	PBW283	127	-0.0605 -	ISM02.4	PPB	10/24/2019 15:13	SMPL
K5320-01	ME0AA4	99	-0.0674 -	ISM02.4	PPB	10/24/2019 15:15	SMPL
K5320-02	ME0AA7	-26	-0.0983 -	ISM02.4	PPB	10/24/2019 15:17	SMPL
K5320-03	ME0AB3	142	-0.0568 -	ISM02.4	PPB	10/24/2019 15:20	SMPL
CCV011	CCV011	21071	5.1076 -	ISM02.4	PPB	10/24/2019 15:22	SMPL
CCB011	CCB011	129	-0.06 -	ISM02.4	PPB	10/24/2019 15:24	SMPL
K5320-04	ME0AB4	129	-0.06 -	ISM02.4	PPB	10/24/2019 15:26	SMPL
K5320-05	ME0AA2	119	-0.0625 -	ISM02.4	PPB	10/24/2019 15:29	SMPL
K5320-06	ME0AA3	109	-0.065 -	ISM02.4	PPB	10/24/2019 15:31	SMPL
K5320-07	ME0AA9	168	-0.0504 -	ISM02.4	PPB	10/24/2019 15:33	SMPL
K5320-08	ME0AB2	232	-0.0346 -	ISM02.4	PPB	10/24/2019 15:35	SMPL
K5320-09	ME0AB6	251	-0.0299 -	ISM02.4	PPB	10/24/2019 15:38	SMPL
K5320-10	ME0AB7	293	-0.0196 -	ISM02.4	PPB	10/24/2019 15:40	SMPL
K5320-11	ME0AA1	106	-0.0657 -	ISM02.4	PPB	10/24/2019 15:42	SMPL
K5320-12	ME0AA5	224	-0.0366 -	ISM02.4	PPB	10/24/2019 15:45	SMPL
K5320-13	ME0AA8	208	-0.0405 -	ISM02.4	PPB	10/24/2019 15:47	SMPL
K5320-14	ME0AB5	164	-0.0514 -	ISM02.4	PPB	10/24/2019 15:49	SMPL
K5320-15	ME0AA6	155	-0.0536 -	ISM02.4	PPB	10/24/2019 15:51	SMPL
K5320-16	ME0AA6D	177	-0.0482 -	ISM02.4	PPB	10/24/2019 15:54	SMPL

LB105908

INSTRUMENT ID : CV1

K5320-17	MEOAA6S	5113	1.1698 -	ISM02.4	PPB	10/24/2019 15:56 SMPL
K5320-18	MEOAB0	193	-0.0442 -	ISM02.4	PPB	10/24/2019 15:58 SMPL
K5320-19	MEOAB1	232	-0.0346 -	ISM02.4	PPB	10/24/2019 16:01 SMPL
CCV012	CCV012	21154	5.1281 -	ISM02.4	PPB	10/24/2019 16:03 SMPL
CCB012	CCB012	105	-0.066 -	ISM02.4	PPB	10/24/2019 16:05 SMPL