



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : AMANDEEP

Run Number: LB105930

Slope : 97.8

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/25/2019	10:45
2	CAL2	1	Water	NA	NA	20.1	6.99	10/25/2019	10:46
3	CAL3	1	Water	NA	NA	20.1	10.04	10/25/2019	10:47
4	ICV	1	Water	NA	NA	20.1	7.01	10/25/2019	10:49
5	CCV1	1	Water	NA	NA	20.1	2.01	10/25/2019	10:50
6	K5556-01	1	Water	NA	NA	15.1	4.70	10/25/2019	10:52
7	K5556-01DUP	1	Water	NA	NA	15.2	4.71	10/25/2019	10:53
8	K5556-02	1	Water	NA	NA	15.3	4.01	10/25/2019	10:55
9	K5556-03	1	Water	NA	NA	15.9	4.36	10/25/2019	11:00
10	CCV2	1	Water	NA	NA	20.1	12.02	10/25/2019	11:02

213105930

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5556 WorkList ID : 132608 Department : Wet-Chemistry Date : 10-25-2019 10:37:24

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/24/2019	Water	K5556-01 - 17	pH	Cool 4 deg C	STER09	E21	CELL-A-POOL-WATER-(DE-L	10/24/2019	SM 4500-H B
10/24/2019	Water	K5556-02 - 17	pH	Cool 4 deg C	STER09	E21	CELL-B-C-POOL-WATER-(DE	10/24/2019	SM 4500-H B
10/24/2019	Water	K5556-03 - 17	pH	Cool 4 deg C	STER09	E21	CELL-D-POOL-WATER-(DE-L	10/24/2019	SM 4500-H B

Date/Time 10-25-19 10:40
Received by: Jn
Relinquished by: ed

Date/Time 10-25-19 11:40
Received by: SC
Relinquished by: Jn