


LB105933

7470A

INSTRUMENT ID : CV1



Linear

A= 0.0000e+000

B= 2.5237e-004

C= -1.2281e-001

Rho= 0.9986929

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.062	-0.062	242	0.000	242				
0.2	0.200	0.226	0.026	1381	0.0 %	1381				
2.5	2.500	2.340	-0.160	9757	0.0 %	9757				
5.0	5.000	5.245	0.245	21270	0.0 %	21270				
7.5	7.500	7.723	0.223	31087	0.0 %	31087				
10.0	10.000	9.729	-0.271	39036	0.0 %	39036				

LB105933 INSTRUMENT ID : CV1

Method: 7470A		Operator: Admin		Date of Analysis: 25 Oct 2019 12:29:13					
Sample ID	Extended ID	µ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	242	-	0.0000	7470A	PPB	25 Oct 2019 12:35:55	S	Std
0.2 - 1	50.2	1381	-	0.2000	7470A	PPB	25 Oct 2019 12:38:04	S	Std
2.5 - 1	52.5	9721	-	2.5000	7470A	PPB	25 Oct 2019 12:40:13	S	Std
5.0 - 1	55.0	0	-	5.0000	7470A	PPB	25 Oct 2019 12:42:22	S	Std
2.5 - 1	57.5	9757	-	2.5000	7470A	PPB	25 Oct 2019 12:48:28	S	Std
5.0 - 1	510.0	21270	-	5.0000	7470A	PPB	25 Oct 2019 12:50:44	S	Std
7.5 - 1		31087	-	7.5000	7470A	PPB	25 Oct 2019 12:52:52	S	Std
10.0 - 1		39036	-	10.0000	7470A	PPB	25 Oct 2019 12:57:24	S	Std
ICV74 - 1	ICV74	16085	3.9366	-	7470A	PPB	25 Oct 2019 13:00:11	U	SMPL
ICB74 - 1	ICB74	178	-0.0779	-	7470A	PPB	25 Oct 2019 13:02:18	U	SMPL
CCV19 - 1	CCV19	20346	5.0119	-	7470A	PPB	25 Oct 2019 13:04:24	U	SMPL
CCB19 - 1	CCB19	184	-0.0764	-	7470A	PPB	25 Oct 2019 13:06:31	U	SMPL
CRA - 1	CRA	1341	0.2156	-	7470A	PPB	25 Oct 2019 13:08:37	U	SMPL
HighStd - 1	HighStd	39242	9.7807	-	7470A	PPB	25 Oct 2019 13:10:44	U	SMPL
ChkStd - 1	ChkStd	28317	7.0236	-	7470A	PPB	25 Oct 2019 13:12:51	U	SMPL
PB124319BL - 1	PBW	265	-0.0559	-	7470A	PPB	25 Oct 2019 13:14:59	U	SMPL
PB124319BS - 1	LCSW	16645	4.0779	-	7470A	PPB	25 Oct 2019 13:17:06	U	SMPL
K5145-06 - 1	HD-02-102319	70	-0.1051	-	7470A	PPB	25 Oct 2019 13:19:15	U	SMPL
K5145-06DUP - 1	HD-02-102319DUP	106	-0.0961	-	7470A	PPB	25 Oct 2019 13:21:23	U	SMPL
K5145-06MS - 1	HD-02-102319MS	18068	4.4370	-	7470A	PPB	25 Oct 2019 13:23:31	U	SMPL
K5145-06MSD - 1	HD-02-102319MSD	17200	4.2180	-	7470A	PPB	25 Oct 2019 13:25:40	U	SMPL
K5148-12 - 1	OR-02-102319	141	-0.0872	-	7470A	PPB	25 Oct 2019 13:27:47	U	SMPL
CCV20 - 1	CCV20	20491	5.0485	-	7470A	PPB	25 Oct 2019 13:29:55	U	SMPL
CCB20 - 1	CCB20	238	-0.0627	-	7470A	PPB	25 Oct 2019 13:32:02	U	SMPL
PB124320BL - 1	PBW	288	-0.0501	-	7470A	PPB	25 Oct 2019 13:34:09	U	SMPL
PB124320BS - 1	LCSW	16725	4.0981	-	7470A	PPB	25 Oct 2019 13:36:16	U	SMPL
K5514-01 - 1	ROLL-OFF-WATER	191	-0.0746	-	7470A	PPB	25 Oct 2019 13:38:23	U	SMPL
K5514-01DUP - 1	ROLL-OFF-WATERDUP	185	-0.0761	-	7470A	PPB	25 Oct 2019 13:40:30	U	SMPL
K5514-01MS - 1	ROLL-OFF-WATERMS	17386	4.2649	-	7470A	PPB	25 Oct 2019 13:45:35	U	SMPL
K5514-01MSD - 1	ROLL-OFF-WATERMSD	17886	4.3911	-	7470A	PPB	25 Oct 2019 13:47:43	U	SMPL
PB124321BL - 1	PBW	266	-0.0557	-	7470A	PPB	25 Oct 2019 13:49:50	U	SMPL
PB124321BS - 1	LCSW	16709	4.0940	-	7470A	PPB	25 Oct 2019 13:51:59	U	SMPL
K5153-07 - 1	TR-05-102419	130	-0.0900	-	7470A	PPB	25 Oct 2019 13:54:07	U	SMPL
K5153-07DUP - 1	TR-05-102419DUP	124	-0.0915	-	7470A	PPB	25 Oct 2019 13:56:14	U	SMPL
CCV21 - 1	CCV21	20546	5.0624	-	7470A	PPB	25 Oct 2019 13:58:21	U	SMPL
CCB21 - 1	CCB21	260	-0.0572	-	7470A	PPB	25 Oct 2019 14:00:27	U	SMPL
K5153-07MS - 1	TR-05-102419MS	17800	4.3694	-	7470A	PPB	25 Oct 2019 14:06:26	U	SMPL
K5153-07MSD - 1	TR-05-102419MSD	17674	4.3376	-	7470A	PPB	25 Oct 2019 14:08:33	U	SMPL
PB124322BL - 1	PBW	276	-0.0532	-	7470A	PPB	25 Oct 2019 14:10:40	U	SMPL
PB124322BS - 1	LCSW	16709	4.0940	-	7470A	PPB	25 Oct 2019 14:14:22	U	SMPL
K4660-07 - 1	TP-1-GROUNDWATER	66	-0.1062	-	7470A	PPB	25 Oct 2019 14:16:30	U	SMPL
K4660-07DUP - 1	TP-1-GROUNDWATERDUP	64	-0.1067	-	7470A	PPB	25 Oct 2019 14:18:37	U	SMPL
K4660-07MS - 1	TP-1-GROUNDWATERMS	16454	4.0297	-	7470A	PPB	25 Oct 2019 14:20:46	U	SMPL
K4660-07MSD - 1	TP-1-GROUNDWATERMSD	17538	4.3033	-	7470A	PPB	25 Oct 2019 14:22:55	U	SMPL
K5494-10 - 1	FB-102119	333	-0.0388	-	7470A	PPB	25 Oct 2019 14:25:04	U	SMPL
K5542-21 - 1	EQUIPMENT-BLANK-102419	233	-0.0640	-	7470A	PPB	25 Oct 2019 14:27:11	U	SMPL
CCV22 - 1	CCV22	20394	5.0240	-	7470A	PPB	25 Oct 2019 14:32:36	U	SMPL
CCB22 - 1	CCB22	165	-0.0812	-	7470A	PPB	25 Oct 2019 14:34:43	U	SMPL
PB124319TB - 1	PB124240TB	296	-0.0481	-	7470A	PPB	25 Oct 2019 14:36:50	U	SMPL
PB124320TB - 1	PB124242TB	302	-0.0466	-	7470A	PPB	25 Oct 2019 14:38:56	U	SMPL
PB124321TB - 1	PB124286TB	303	-0.0463	-	7470A	PPB	25 Oct 2019 14:41:03	U	SMPL
K5145-06LX5 - 1		37	-0.1135	-	7470A	PPB	25 Oct 2019 14:43:10	U	SMPL
K5145-06A - 1		17977	4.4141	-	7470A	PPB	25 Oct 2019 14:45:17	U	SMPL
K5514-01LX5 - 1		94	-0.0991	-	7470A	PPB	25 Oct 2019 14:47:25	U	SMPL
K5514-01A - 1		18002	4.4204	-	7470A	PPB	25 Oct 2019 14:49:34	U	SMPL
K5153-07LX5 - 1		35	-0.1140	-	7470A	PPB	25 Oct 2019 14:51:42	U	SMPL
K5153-07A - 1		16666	4.0832	-	7470A	PPB	25 Oct 2019 14:53:51	U	SMPL
K4660-07LX5 - 1		157	-0.0832	-	7470A	PPB	25 Oct 2019 14:55:58	U	SMPL
CCV23 - 1	CCV23	19940	4.9095	-	7470A	PPB	25 Oct 2019 14:58:05	U	SMPL
CCB23 - 1	CCB23	195	-0.0736	-	7470A	PPB	25 Oct 2019 15:00:12	U	SMPL
K4660-07A - 1		16836	4.1261	-	7470A	PPB	25 Oct 2019 15:02:19	U	SMPL
CCV24 - 1	CCV24	19624	4.8297	-	7470A	PPB	25 Oct 2019 15:04:27	U	SMPL
CCB24 - 1	CCB24	38	-0.1132	-	7470A	PPB	25 Oct 2019 15:06:36	U	SMPL