



Analytical Summary Report

Analysis Method: 1010A

Reviewed By: ketankumar

Parameter: Flash Point

Supervisor Review By: apatel

Run Number: LB105934

Ambient Barometric Pressure (mmHg): 760.00

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W2168

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	7	27.00	80.6	80.6	10/25/2019	11:05
2	K5534-01		1	9	100.00	>212.0	>212.0	10/25/2019	11:35
3	K5534-01DUP		1	9	100.00	>212.0	>212.0	10/25/2019	11:55
4	K5534-03		1	7	94.00	201.2	201.2	10/25/2019	12:25

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

LB105934

WorkList Name : fp102519 WorkList ID : 132607 Department : Wet-Chemistry Date : 10-25-2019 10:33:13

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/31/2019	Solid	K5534-01	Flash Point	Cool 4 deg C	PSEG01	N51	OIL-E	10/24/2019	1010A
10/31/2019	Solid	K5534-03	Flash Point	Cool 4 deg C	PSEG01	N51	OIL-A-C	10/24/2019	1010A

Date/Time 10/25/19 10:35
 Received by: KS
 Relinquished by: JP

Date/Time 10/25/19 12:30
 Received by: JP
 Relinquished by: KS