

Analytical Summary ReportAnalysis Method: 365.3ANALYST: AMANDEEPParameter: Phosphorus-TotalSUPERVISOR REVIEW BY: apatelRun Number: LB105937

Reagent/Standard	Lot/Log #
5N sulfuric acid	WP76600
calibration std. phosphate 1 ppm	WP79363
calibration std. phosphate 0.5 ppm	WP79362
calibration std. phosphate 0.3 ppm	WP79361
calibration std. 0 ppm	WP79358
phosphate ICV std.	WP79365
phosphate CCV std.	WP79364
calibration std. phosphate 0.1 ppm	WP79360
calibration std. phosphate 0.05 ppm	WP79359
Combined reagent	WP79367
11 N sulfuric acid	WP77312
Phenolphthalein indicator	WP77313
Sodium hydroxide, 1N	WP77314

Intercept: -0.0006Slope: 0.6487Regression: 0.999856

Seq	Lab ID	True Value (mg/L)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	AnalDate	AnalTime
1	CAL1	0.00	1	50	50	0.000	10/25/2019	12:29
2	CAL2	0.05	1	50	50	0.033	10/25/2019	12:30
3	CAL3	0.10	1	50	50	0.066	10/25/2019	12:30
4	CAL4	0.30	1	50	50	0.186	10/25/2019	12:31
5	CAL5	0.50	1	50	50	0.328	10/25/2019	12:32
6	CAL6	1.00	1	50	50	0.648	10/25/2019	12:32



Analytical Summary Report

Analysis Method: 365.3

ANALYST: AMANDEEP

Parameter: Phosphorus-Total

SUPERVISOR REVIEW BY: apatel

Run Number: LB105937

Seq	Lab ID	True Value (mg/l)	DF	Initial Volume (mL)	Final Volume (mL)	Absorbance Reading at 880nm	Result (mg/L)	AnalDate	AnalTime
1	ICV	0.50	1	50	50	0.321	0.496	10/25/2019	12:33
2	ICB		1	50	50	0.000	0.001	10/25/2019	12:33
3	CCV1	0.50	1	50	50	0.324	0.500	10/25/2019	12:34
4	CCB1		1	50	50	0.000	0.001	10/25/2019	12:34
5	PB124326BL		1	50	50	0.000	0.001	10/25/2019	12:35
6	PB124326BS	0.50	1	50	50	0.299	0.462	10/25/2019	12:36
7	K5490-13		1	50	50	0.147	0.228	10/25/2019	12:36
8	K5490-13DUP		1	50	50	0.150	0.232	10/25/2019	12:37
9	K5490-13MS	0.50	1	50	50	0.467	0.721	10/25/2019	12:37
10	K5490-13MSD	0.50	1	50	50	0.469	0.724	10/25/2019	12:38
11	CCV2	0.50	1	50	50	0.321	0.496	10/25/2019	12:38
12	CCB2		1	50	50	0.000	0.001	10/25/2019	12:39