



Analytical Summary Report

Analysis Method: SM 4500-H B

Analyst By : JIGNESH

Parameter: pH

Supervisor Review By : apatel

Run Number: LB105942

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.1	4.01	10/25/2019	15:45
2	CAL2	1	Water	NA	NA	20.1	7.00	10/25/2019	15:46
3	CAL3	1	Water	NA	NA	20.1	10.03	10/25/2019	15:47
4	ICV	1	Water	NA	NA	20.2	6.99	10/25/2019	15:48
5	CCV1	1	Water	NA	NA	20.1	2.02	10/25/2019	15:50
6	K5570-01	1	Water	NA	NA	19.3	5.22	10/25/2019	15:52
7	K5570-01DUP	1	Water	NA	NA	19.4	5.23	10/25/2019	15:53
8	CCV2	1	Water	NA	NA	20.1	12.02	10/25/2019	16:00

LB105942

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH K5570 WorkList ID : 132620 Department : Wet-Chemistry Date : 10-25-2019 15:39:12

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/22/2019	Water	K5570-01 ✓	pH	Cool 4 deg C	STER17	E21	JS9600	10/22/2019	SM 4500-H B

Date/Time 10-25-19 15:40
Received by: JS
Relinquished by: CL

Date/Time 10-25-19 16:30
Received by: SP
Relinquished by: JG