



Analytical Summary Report

Analysis Method: 9040C

Analyst By : JIGNESH

Parameter: Corrosivity

Supervisor Review By : apatel

Run Number: LB105950

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W2568
buffer solution pH 7 yellow	W2447
PH 10.01 BUFFER,COLOR CD 475ML	W2570
buffer solution pH 7 yellow	W2446
Buffer Solution, PH2 (500ml)	W2592
Buffer Solution, PH12 (500ml)	W2467

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/25/2019	16:25
2	CAL2	1	Water	NA	NA	20.2	7.00	10/25/2019	16:26
3	CAL3	1	Water	NA	NA	20.2	10.03	10/25/2019	16:27
4	ICV	1	Water	NA	NA	20.1	6.99	10/25/2019	16:30
5	CCV1	1	Water	NA	NA	20.1	2.02	10/25/2019	16:33
6	K5569-02	1	Water	NA	NA	18.4	9.17	10/25/2019	16:35
7	K5569-02DUP	1	Water	NA	NA	18.3	9.18	10/25/2019	16:36
8	CCV2	1	Water	NA	NA	20.1	12.02	10/25/2019	16:37

18105950

WORKLIST(Hardcopy Internal Chain)

WorkList Name : CORROSIVITY K5569

WorkList ID : 132623

Department : Wet-Chemistry

Date : 10-25-2019 15:50:51

Due Date	Matrix	Sample	Test	Preservative	Customer	Storage Location	Customer Sample	Collect Date	Method
10/25/2019	Water	K5569-02	☒ Corrosivity	Cool 4 deg C	JACO05		IDW-DECON-WATER-102519	10/25/2019	9040C

Date/Time 10-25-19 16:00
 Received by: Jo
 Relinquished by: ea

Date/Time 10-25-19 17:00
 Received by: ea
 Relinquished by: ea